

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63832

Customer	Alorif Hospital	Equipment	Washer unit.
		Model No.	Uniclean PL 11
		Serial No.	230506
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	CSSD
unloading door not opening		Contract / Order No.	
Services / Check Carried out ☐ Inspection / Cleaning ☐ Calibration / Performance ☐ Interface / Communication ☐ Safety Checks ☐ Electrical / Electronics ☐ Accessories Check ☐ Optics Checks ☐ Pneumatic Checks ☐ Hydraulic Checks ☐ Operational Error ☐ Mechanical Checks ☐		☐ Installation	☐ Warranty
		☒ Repair	☐ On Charge
		☐ Maintenance	☐ Contract
		☐ Upgrade	☐
		Time Spent	
Remarks Service call attended, we inspect the Unit and found that the unloading door not opening. we opened the door manually and reset the relay, now the door is working properly, we done one cycle and the door opens properly, we are keeping the unit under observation.			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer NIDON ATCO Technology Co. Signature Date 15/02/2025		Authorized Person Signature Date	
		Customer / User Signature Date	