

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63670

Customer	Alort Hospital	Equipment	ENT workstation
		Model No.	Futurent III
		Serial No.	111632045
Authorized Person		Manufacturer	Entered
Customer Complaint		Location	ENT 5th floor
No water / LEAKING		Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☒ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		
Remarks we replaced the water filter inside the workstation with new one. we repaired and cleaned the entire machine. we fixed the unit properly. checked and confirmed that the unit is working properly			
No.	Spare Parts Used	Part No.	Qty.
Service Engineer ATCO Technology Co. Signature Date 20/05/2025		Authorized Person Signature Date	
		Customer / User Signature Date	