

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO : 336218



شركة أتكوتكنولوجيه

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63671

Customer	ALORF Hospital	Equipment	Washer Disinfectant
	Jahra	Model No.	Uniclean PL10
		Serial No.	230504
Authorized Person		Manufacturer	MMM
Customer Complaint		Location	ess D
	clean side door not opening	Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☐ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Maintenance	☐ Contract
☐ Calibration / Performance	☐ Upgrade	☐
☐ Interface / Communication		
☐ Safety Checks		
☐ Electrical / Electronics		
☐ Accessories Check		
☐ Optics Checks		
☐ Pneumatic Checks		
☐ Hydraulic Checks		
☐ Operational Error		
☐ Mechanical Checks		
☐		
	Time Spent	

Remarks
Checked found clean side door not open
Reset clean side door relay
All connectors are tested ok
One cycle checked completely
Washer working good

No.	Spare Parts Used	Part No.	Qty.

Service Engineer	Authorized Person	Customer / User
J.P. & Since		
ATCO Technology Co.		
Signature	Signature	Signature
Date 22/05/2015	Date 20/5/2015	Date