

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهر، ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63828

Customer	Aloft Hospital	Equipment	ENT Chair
		Model No.	Futurent III / 760110ES
		Serial No.	2010K0023
Authorized Person		Manufacturer	Entemed
Customer Complaint		Location	5th floor Dr. Evan
		Contract / Order No.	

Services / Check Carried out

- | | |
|---|---|
| ☐ Inspection / Cleaning | ☐ Optics Checks |
| ☐ Calibration / Performance | ☐ Pneumatic Checks |
| ☐ Interface / Communication | ☐ Hydraulic Checks |
| ☐ Safety Checks | ☐ Operational Error |
| ☐ Electrical / Electronics | ☐ Mechanical Checks |
| ☐ Accessories Check | ☐ |

- | | |
|---|-------------------------------------|
| ☐ Installation | ☐ Warranty |
| ☒ Repair | ☐ On Charge |
| ☐ Maintenance | ☐ Contract |
| ☐ Upgrade | ☒ Em |

Time Spent

Remarks: ENT chair lock not working properly. We fixed the lock and tested the unit, checked and confirmed that the unit is working properly.

No.	Spare Parts Used	Part No.	Qty.

Service Engineer

Mixom

ATCO Technology Co.

Signature

Date: 26/01/2025

Authorized Person

Signature

Date

Customer / User

كوبيتا شاتاف
رئيس مجلس أمناء وأمين خزانة
مستشفى الحرف
22023120704

Signature

Date: 28/1/25