

Service Report



مجموعة وناس الهندسية
Wanas Engineering Group

Customer Name & Address

AL-Orf Hospital
Dermatology clinic

1108

Instrument Pulsed dye laser

Model No. Vbeam Perfecta

Serial No. 9914-0300-3443

Manufacturer Candela

Contract No.

LOCATION

Customer ☒

Installation ☐

Repair ☒

Preventive Maintenance ☐

Workshop ☐

Warranty ☐

Charge ☒

Contract ☐

ENGINEER'S REPORT & FINAL STATUS OF INSTRUMENT

PARTS USED FOR REPAIR

- As per the customer's order one new Dye Kit was supplied and installed into the laser unit
- Doing system energy and circuit calibrations after installing The new Dye Kit
- system performance checked
- later is working OK

Description

Parts No.

Qty.

Dye Kit

1

Notes

Customers comments

Work Completed

YES ☒

NO. ☐

Name of Engineer

M. Wanas

Date

23/7/2024

Signature

Name of Engineer

Date

Signature

Name of Customer & Title

Date

Signature

د. ايضيت توبيا نوحا
استشاري امراض جلدية وتناسية
مستشفى الاعرف
140925