

MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 05/05/25

Requested by Anumol Antony

Department PT Location Female

SR. No.	Machine	Model	Complaint
<u>NA</u>	<u>Patient Bed.</u>		<u>Pedal not working</u>

Check within Week

☒ ASAP

Engineer Comments

End Comments

Anumol
Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO. 336218



شركة أتكو التكنولوجية

تلفون: ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس: ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب: ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61172

Customer	ALORF HOSPITAL	Equipment	Examination bed
	Jahra	Model No.	3 sections Electric Plinth
		Serial No.	
Authorized Person		Manufacturer	Medisa
Customer Complaint		Location	PT - Ground Floor
	Foot switch not working	Contract / Order No.	
Services / Check Carried out		☐ Installation	☐ Warranty
☐ Inspection / Cleaning	☐ Optics Checks	☐ Repair	☐ On Charge
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Maintenance	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade	☐ _____
☐ Safety Checks	☐ Operational Error	Time Spent	
☐ Electrical / Electronics	☐ Mechanical Checks		
☐ Accessories Check	☐ _____		

Remarks
Checked found foot switch defective
Repaired foot switch. All connections are tested on
Bed working good

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Jinice Cherran ATCO Technology Co. Signature Date 15/05/25	Authorized Person د. نادية أحمد محمد مستشفى الطيبعي والتأهيل 133161983 Signature Date 18/5/2025	Customer / User Signature Date 15/05/25
---	--	---