ATCO Technology Co.
Tel.: (+965) 24743017 / 24567725
Fax: (+965) 24743019 / 24567792
P.O. Box: 2355 Jahra, 01025 - Kuwait
Email: sales@atco-kuwait.com
www.atco-kuwait.com
CR NO . 336218

ATCOTECH

loyno-2132

شركة أتكو التكنولوجية
تلفون : ٢٤٥٦٧٧٢٥ / ٢٤٧٤٣٠١٧ (٩٦٥+)
فاكس : ٢٤٥٦٧٧٩٢ / ٢٤٧٤٣٠١٩ (٩٦٥+)
ص.ب : ٢٣٥٥ الجهراء, ٠١٠٢٥ - الكويت
Email: sales@atco-kuwait.com
www.atco-kuwait.com
س.ت. ٣٣٦٢١٨

SERVICE / ~~INSTALLATION~~ REPORT

No.: 60981

Customer Alorf Hospital	**Equipment** Surgical microscope
	Model No. IVU 55
	Serial No. 1508-01400
Authorized Person	**Manufacturer** Labomed - Prima
Customer Complaint	**Location** OR
Light Source Not working	**Contract / Order No.**

Services / Check Carried out

- O Inspection / Cleaning
- O Calibration / Performance
- O Interface / Communication
- O Safety Checks
- O Electrical / Electronics
- O Accessories Check
- O Optics Checks
- O Pneumatic Checks
- O Hydraulic Checks
- O Operational Error
- O Mechanical Checks
- O ———

- O Installation
- O ~~Repair~~
- O Maintenance
- O Upgrade
- O Warranty
- O On Charge
- O Contract
- O ———

Time Spent

Remarks Service call attended.
We inspect the unit and found that the light source cable is removed from the main source. We fixed the light source cable and now the unit is working Properly.
checked and confirmed that the unit is working Properly

No.	Spare Parts Used	Part No.	Qty.
	working now		

Service Engineer	Authorized Person	Customer / User
Nixon ATCO Technology Co.	16/7/24	
Signature	Signature	Signature
Date 13/07/2024	Date	Date