

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرام, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 63009

Customer	Aloof Hospital	Equipment	Blanket warmer
		Model No.	FWL/SP
		Serial No.	11 J 1790
Authorized Person		Manufacturer	Hasami
Customer Complaint		Location	OR
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☒ Repair	☐ On Charge
☐ Inspection / Cleaning	☐ Optics Checks	☐ Maintenance
☐ Calibration / Performance	☐ Pneumatic Checks	☐ Contract
☐ Interface / Communication	☐ Hydraulic Checks	☐ Upgrade
☐ Safety Checks	☐ Operational Error	☐ _____
☐ Electrical / Electronics	☐ Mechanical Checks	
☐ Accessories Check	☐ _____	
	Time Spent	

Remarks	Service call attended. we found the temperature display not working. we replaced the display battery with new one, checked and confirmed that the unit is working in good condition.
---------	--

No.	Spare Parts Used	Part No.	Qty.
	Battery ATF6		1

Service Engineer Nisam ATCO Technology Co. Signature Date 03/05/2025	Authorized Person Signature Date	Customer / User Hussein / 3/5/25 Signature Date
--	--	--