



MEDICAL EQUIPMENT REPAIR REQUEST

No. _____

Date: 13/10/24

Requested by Geetha mal Sumon

Department OBG1[OPD] Location 1st Floor old Building

SR. No.	Machine	Model	Complaint
<u>19101628</u>	<u>Laser</u>	<u>Fotomg M021-7AF/3</u>	<u>Foot Pedal Not delivering shots</u>

Check within

☐ Week

☐ ASAP

Engineer Comments

End Comments

Requested by

Checked By

Approved by

ATCO Technology Co.

Tel.: (+965) 24743017 / 24567725

Fax: (+965) 24743019 / 24567792

P.O. Box: 2355 Jahra, 01025 - Kuwait

Email: sales@atco-kuwait.com

www.atco-kuwait.com

CR NO . 336218



1807
شركة أتكو التكنولوجية

تلفون : ٢٤٧٤٣٠١٧ / ٢٤٥٦٧٧٢٥ (+٩٦٥)

فاكس : ٢٤٧٤٣٠١٩ / ٢٤٥٦٧٧٩٢ (+٩٦٥)

ص.ب : ٢٣٥٥ الجهرء, ٠١٠٢٥ - الكويت

Email: sales@atco-kuwait.com

www.atco-kuwait.com

س.ت. ٣٣٦٢١٨

SERVICE / INSTALLATION REPORT

No.: 61012

Customer	Aloof Hospital Jahra	Equipment	Cynae Laser
		Model No.	SP Dynamus
		Serial No.	19101628
Authorized Person		Manufacturer	Fotona
Customer Complaint	Not. hinoz	Location	1st FLR. OB
		Contract / Order No.	

Services / Check Carried out	☐ Installation	☐ Warranty
	☐ Repair	☐ On Charge
	☐ Maintenance	☐ Contract
	☐ Upgrade	☐ _____
	Time Spent	

☐ Inspection / Cleaning	☐ Optics Checks
☐ Calibration / Performance	☐ Pneumatic Checks
☐ Interface / Communication	☐ Hydraulic Checks
☐ Safety Checks	☐ Operational Error
☐ Electrical / Electronics	☐ Mechanical Checks
☐ Accessories Check	☐ _____

Remarks	Service call attended. Checked the unit and suspected problem with A-runner interface. Taking interface module to company for checking.
---------	---

No.	Spare Parts Used	Part No.	Qty.

Service Engineer Signature: Date: 14/10/2024	Authorized Person Signature: Date: 19/10/2024	Customer / User Signature: _____ Date: _____
--	---	--