

PGA Rationale Memo - JB-LED-01

Generated by Azali - 2 pages

Determination	FDA DEV flag required - confidence 0.74
Basis	21 CFR Sec. 878.4810 - light-based devices for medical purposes
Exemption	General-wellness exemption likely unavailable on current claims
Forward fix	Artwork/claims cleanup could restore the exemption on future POs