分析證書
ANALYSIS CERTIFICATE

Product	Storage Solution	Quantity	1000 ml x 20
Ref.standard	Q/ZRS 030S-2010	Batch No.	20260310
Date	10.03.2026	Exp Date	10.03.2029
Items	Specifications	Results	
Conclusion	Conformed with the requirements of Company standard		

Electrodes Fill Solution (3M KCl)

Each 1 liter = 223.7 grams of potassium chloride,

Do NOT store the electrode in deionizer water.

Test by:Fang He   Reviewby:Yan Yan Ren   Surveyor:Yanjun Xiao

測試者：何芳Reviewby ：顏燕人測量師：王彥軍肖

地址：广州市东风西路132号（流花广场）园景中心911室
TEL: 0086-20-81368125   FAX: 0086-20-81368157
E-mail: oscltd.2007@yahoo.com.cn

This certificate is an electronic copy of the certificate available in our laboratory, and does not require signature

Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az chem Az