

SYNTHETIC TEST DOCUMENT — LIFE SCIENCES QUALITY ASSURANCE TEST DATASET — STRICTLY FOR DEMONSTRATION & VALIDATION

Analytical Balance Calibration Out-of-Tolerance during Potency Assay

1. GENERAL INCIDENT INFORMATION

Date of Occurrence:	2026-09-21	Discovery Source:	Routine Daily QC Balance Check (SOP-QC-011)
Manufacturing Site:	Analytical QC Testing Lab 3, Cambridge Center	Department / Area:	Quality Control Laboratory
Related Product / Material:	Atorvastatin Calcium Standard Solution	Batch / Lot Number:	QC-STD-2026-09

2. EVENT SUMMARY & DEVIATION DESCRIPTION

Analytical balance BAL-004 failed routine 100 mg calibration check, displaying 104.8 mg (tolerance ± 0.5 mg). Balance had been used earlier that morning to weigh active standards for release assay of commercial batch ATOR-3301.

3. IMMEDIATE ACTIONS & CONTAINMENT MEASURES

Balance BAL-004 was immediately tagged Out of Service. All HPLC potency testing performed on September 21 utilizing standards weighed on BAL-004 was invalidated.

4. INITIAL LOGGING VERIFICATION

Logged By: Operations Shift Lead Date: 2026-09-21	QA Reviewer: Quality Assurance On-Duty Date: 2026-09-21	Workflow Stage: Awaiting AI Extraction & Human Review
--	--	---