LU2006.01_bsl	
Director:	Panel: 6 Color TBNK + TruC Acquired: 8/31/2020 2:34:13 PM Analyzed: 8/31/2020 2:34:13 PM TruC Lot ID: 19064 Bead/Pellet: 49100 Status: OK Operator: BD Reviewer: Results: 31082020.csv
Column #1:	Column #2: Column #3:

BD FACSCanto II V96300506 BD FACSCanto v.3.1.5878.21241

CD3/CD16+56/CD45/CD4/CD19/CD8 TruC Total Events: 10115



LU2006.01_bsl001.001.fcs Reagent Lot ID: 00000

Parameter	Percent	Value/AbsCnt
Lymph Events		2505
Bead Events		1231
CD3+	44.71	893.45
CD3+CD8+	19.56	390.89
CD3+CD4+	24.71	493.79
CD3+CD4+CD8+	0.92	18.35
CD16+CD56+	51.50	1029.07
CD19+	3.51	70.20
CD45+		1998.30
4/8 Ratio		1.26

QC Messages

% T-Sum is: 0.44
Lymphosum is: 99.72
4/8 ratio is: 1.26
Cytometer settings were generated from a failed setup result.

Comments