

Serial Protocol Decode — user manual

Ian Ameline · version 0.9 — **beta** · MIT licence

This app turns a Keithley DMM6500 into a serial decoder. Clip onto a UART line, press **Capture**, and read the bytes on the front panel. You do not have to tell it the baud rate, the frame format or which way up the signal is — it works all of that out from the signal itself.

Three things make it worth reaching for over a scope's decode: the **window is large** (a screenful is ~240 bytes, and a recording holds 32 768), every capture is **appended to the USB key automatically** while a key is in the slot, so you have a log without asking for one, and the **serial parameters are detected rather than configured**. For most debugging that is the whole job — connect, press, read.

This is a beta. Every number in this manual was measured on a real DMM6500 against a real signal generator unless the text says otherwise — and where something is calculated rather than measured, or untested, it says so in that spot. What that cannot cover is the variety of real devices, so if the app tells you something you can prove is wrong, that is the bug worth reporting. It is built so that **refusing with a reason is acceptable and confident nonsense is not**, and anything it is unsure about it says on the note row at the bottom of the screen.

Hooking it up — read this part

The line must stay between -10 V and +10 V. The app fixes the meter on its 10 V range. 3.3 V CMOS and 5 V TTL are both fine as they are. For a 12 V line, put a 2:1 resistor divider in front of it — and if it is an open-drain bus, pick resistors big enough not to load it down.

Connect INPUT LO to the ground of the circuit you are probing. Everything is measured at INPUT HI relative to INPUT LO, so without a shared ground the readings mean nothing.

Do not put a differential bus straight across HI and LO. RS-485, CAN and LIN need to go through a transceiver first, or you need to tap one side of the pair against ground.

A DC offset does not matter. The app measures the two voltage levels on your line and puts its decision threshold between them, so it does not care where they sit — only that both fit inside ± 10 V. A 1.6 V swing sitting up at 6 V decodes fine. **1.6 V is the smallest swing tested.** A 60 mV swing is refused outright.

Quick start

1. Connect the line to **INPUT HI** and its ground to **INPUT LO**.
2. Press **Capture** while the device is talking.
3. Read the bytes. Press **View** to switch between hex and plain text.

That is the whole normal workflow. The first capture also locks the baud rate when it is confident, so later captures are faster.

Two more things worth knowing before you need them: **Mode** turns Capture into a recording of up to 32 768 bytes straight to the USB key, still one press — and the physical **TRIGGER** key stops a long job, keeping everything decoded so far.

The screen

The screenshot shows the main interface of the 'SERIAL PROTOCOL DECODE' application. At the top, a status bar displays various parameters: BAUD (9600), FORMAT (8N1), IDLE (HIGH), LOGIC (3V3 CMOS), THRESH (1.63 V), SA/BIT (8.3), RATE (80 kS/s), BYTES (152), ERR (0), and FIT (0.98). Below this, a table displays the captured data in hex view, showing frame numbers, hex values, and the corresponding ASCII text. The data appears to be a sequence of 'Hello, World!' messages. At the bottom, there are several buttons: Capture, View, Mode, NewLog, Save, and Options. A status bar at the very bottom shows 'FRAME HEX pg 1/1 bytes 1-152/152 win 240 [done]' and 'log: bytes019.txt 303 B'.

BAUD	FORMAT	IDLE	LOGIC	THRESH	SA/BIT	RATE	BYTES	ERR	FIT
9600	8N1	HIGH	3V3 CMOS	1.63 V	8.3	80 kS/s	152	0	0.98

FRAME	HEX	ASCII
0000	C7 47 17 FE 48 65 6C 6C	???Hello, World
0016	21 48 65 6C 6C 6F 2C 20	!Hello, World!He
0032	6C 6C 6F 2C 20 57 6F 72	llo, World!Hello
0048	6C 64 21 48 65 6C 6C 6F	, World!Hello, W
0064	2C 20 57 6F 72 6C 64 21	orld!Hello, Worl
0080	6F 72 6C 64 21 48 65 6C	d!Hello, World!H
0096	64 21 48 65 6C 6C 6F 2C	ello, World!Hell
0112	65 6C 6C 6F 2C 20 57 6F	o, World!Hello,
0128	6F 2C 20 57 6F 72 6C 64	World!Hello, Wor
0144	57 6F 72 6C 64 21 48 65	ld!Hello

FRAME HEX pg 1/1 bytes 1-152/152 win 240 [done] log: bytes019.txt 303 B EXT TRIG

Buttons: Capture, View, Mode, NewLog, Save, Options

Figure 1: Main screen, hex view

Along the top are the things the app worked out about your line:

Field	Meaning
BAUD	The rate it found. A padlock means it is locked and will not be re-detected.
FORMAT	Data bits, parity, stop bits — 8N1, 7E1 and so on.
IDLE	What the line rests at: HIGH for CMOS/TTL, LOW for an RS-232-style line.
LOGIC	The family it looks like, e.g. 3V3 CMOS, 5V TTL.
THRESH	The voltage it is using to tell 0 from 1.
SA/BIT	Samples per bit. More is better; 8 is comfortable, 4 still works.
RATE	How fast the meter sampled.
BYTES / ERR	How many bytes came out, and how many were broken.
FIT	How well one bit-length explains what it measured, from 0 to 1. Below about 0.9 means the timing was messy.

A dash means “not measured yet”.

Underneath: the **mode cell**, the bytes, then two rows of text. The **status row** tells you where you are (FRAME HEX pg 1/1 bytes 1-155/155 win 240 [done]) and names the log file.

The mode cell is colour-coded, so a glance says which mode you are in without reading: grey for FRAME, blue for the 8 kB recording window, amber for 32 kB. During a recording the status row carries a live counter — recording... 40 % of the buffer, then decoding... 4096/8192 bytes — and the cell to its right reads no stop until it ends, because while the app is busy nothing on the instrument can interrupt it.

The note row is the important one. Warnings appear there — an ambiguous baud rate, a line that disagrees with the rate you locked, a recording that stopped early. If more than one applies it ends with (+2 more), and pressing **Save** writes all of them to a file.

An empty note row is not a promise that everything is perfect. It means the app did not detect anything wrong. A noise spike landing inside a data bit changes that byte with nothing to give it away.

The buttons

Button	Does
Capture	Take a capture and decode it. In a recording mode, does the whole recording.
View	Switch the display: hex + ASCII → plain text → back.
Mode	Choose what Capture does: a screenful, an 8 kB recording, or a 32 kB one.
NewLog	Start a new numbered log file on the USB key.
Save	Write a full report of what is on screen to the USB key.
Options	The settings screen.
Page ▲ / ▼	Right-hand edge. Only appear when there is more than one page.
Lock Rate	Right-hand edge. Only appears when the rate is not locked.

Every button responds immediately except the ones that take a capture. **Capture takes about 1.7 seconds when the rate is locked, and 4 to 9 seconds when it is not** — because to find an unknown rate it has to capture two or three times. Slower lines take longer, not less. The exact time varies from press to press because it depends on where the traffic happens to be when you press.

A press during a capture is not lost — it is queued and acted on when the capture finishes. So if the panel seems to ignore you, wait rather than pressing again.

The TRIGGER key — the one hardware button that matters

The **TRIGGER** key is the physical key on the front panel, to the left of the screen. It is not part of the app's own row of buttons. It was intended to do two things nothing on the glass can do; **one of them works** — timing a capture — and the other, stopping a long job, does not. Both are described below, because knowing which is which is the difference between waiting confidently and pressing a key that does nothing.

1. It does not stop a long job, and nothing else does either

A recording runs to its own end. Once you press **Capture** in 8 kB or 32 kB, nothing on the instrument will cut it short — not a button on the glass, and not the TRIGGER key. Decide before you press. The note row tells you the cost up front:

bounded at 32768 bytes -- ~34 s of this line, and nothing stops it early

and the cell on the status row reads `no stop until it ends` for as long as it is working. Allow more than that figure: it is the time to fill the buffer, and the decode afterwards takes longer than the capture did.

Earlier builds advertised TRIGGER as the stop. That was wrong, and it was measured wrong on the instrument on 2026-08-18: pressed 20 % of the way through a 32 kB decode — with tens of seconds of checking still to come — the run finished normally and the trigger latch was found empty afterwards. The key's press never reaches the app while a run started from the panel is executing.

The app's own side of this is sound, which is why it is worth fixing rather than accepting: when the identical cancel is delivered by a firmware timer instead of a finger, the run stops inside a second, keeps every byte decoded up to that point, and files them. What is missing is a route from a key press to that mechanism.

What does bound a run, so it can never hold the panel indefinitely:

- the byte ceiling for the mode — 8192 or 32768;
- a quiet line: one second with no activity ends it;
- under flow control, 32 windows or 20 minutes of wall clock, whichever comes first. The 20 minutes is what matters on slow lines, where a full 32 kB window would take 18 minutes at 300 baud.

A tap on Capture during a run is not a stop. It is queued by the instrument and delivered when the run finishes. The app recognises presses made during the run it has just finished and discards them, saying so on the note row, rather than starting another recording you did not ask for.

2. It can time a capture

Set Options ▶ Trigger = Trigger key, press **Capture**, and the capture waits for you: it fires when you press the physical key, so you decide the moment the line is sampled. Verified on this hardware — armed on a deliberately idle line, the capture fired on the key press rather than timing out.

If the trigger you chose never arrives, the capture goes ahead free-running after a few seconds rather than failing, and the note row says which one was missing.

These two jobs never collide: the key times a capture only when you have asked for that in Options, and stops a run only while one is going.

How much you can capture at once

Four answers, and almost everyone only ever needs the first. **All of them are one press.**

You need	Use	Get
a look at what a device is saying	FRAME — just press Capture	~ 240 bytes , on screen
a longer burst, quickly	Mode → 8 kB, then Capture	up to 8 192 bytes , to a file
the longest single recording	Mode → 32 kB, then Capture	up to 32 768 bytes , to a file
more than that, losslessly	flow control, and a device that obeys it	as much as you like, ~20 minutes per press

FRAME is the normal case and covers almost all debugging. One press gives you a screenful, already decoded, with the baud rate and frame format worked out for you and a copy appended to the USB key automatically. You do not have to set anything up or lock anything down.

Recording to a file

Press **Mode** until the cell reads 8 kB or 32 kB. **You must lock the baud rate first** — press **Lock Rate**, or type the rate into **Options**. Then **press Capture once**, and that press does the whole job: it records, decodes every byte, writes them to the USB key, and comes back with the tail on screen. There is nothing to press in the middle.

While it works, the counter on the status row moves — first recording... N % of the buffer, then decoding... N/8192 bytes — and the cell beside it reads no stop until it ends. **There is no way to finish it early** — see the section on the TRIGGER key above for what bounds it instead.

It stops by itself when the window is full or when the line has been quiet for a second. There is also a 20-minute backstop, so nothing can wait for ever — it only matters on very slow lines, where a full 32 kB window takes 18 minutes at 300 baud.

Nothing is dropped inside a recording. What you get is a continuous, in-order run of the line — verified with a 1024-byte non-repeating test pattern, where the decoded bytes were one unbroken slice of it.

Which window size to pick

The window is the unit of work: one press records a window, decodes it, and files it. So the size is a straight trade, and neither choice affects what gets decoded — both are lossless.

	8 kB	32 kB
Bytes per press	8 192	32 768
Recording, at 9600 baud	~8.5 s	~34 s
Decoding afterwards	~30 s	~2 minutes
Best for	seeing something soon, and stopping soon	the most data for the fewest presses

Decoding is the slow part, and it is slower than the line: the meter decodes roughly 300 bytes a second whatever the baud rate. That is why 8 kB feels much more responsive even though both are one press.

The app starts in FRAME, and one press of **Mode** reaches 8 kB, a second 32 kB. Stop at 8 kB when you would rather see bytes soon than see all of them, or while you are still working out whether you are probing the right pin; go on to 32 kB when you want the most data for the fewest presses.

More than one window: flow control

A window is not a total. If your device is built to wait, you can capture as much as you like without losing a byte, **and without pressing anything per window** — up to 32 windows or 20 minutes per press, whichever comes first. To carry on past that, press **Mode** to pick the window size again and then **Capture**; it resumes where it stopped, in a new file.

Set Options ▶ Rear BNC = FC Out. The rear EXT TRIG OUT then emits **one ~5 V, ~10 µs pulse each time a capture is armed** (measured on a scope: 4.92 V, 9.4–9.8 µs — the width is the meter’s own default and this firmware does not let the app change it) — a credit meaning “*send now, and send no more than one window’s worth*”. Then one press of **Capture** runs the whole conversation: credit, record, decode, file, credit again — round and round, with the window number on the status row, until the device stops sending. While the app is decoding, nothing is armed, no credit is issued, and a device that waits stays quiet, so the gap where bytes could be lost is closed by design. Everything lands in one file, in order.

It ends when your device goes quiet after a credit, when you press **TRIGGER**, or when it reaches one of two backstops: **32 windows** or **20 minutes**. Those exist so a device that never stops talking cannot leave the panel busy for ever, and the 20 minutes is usually the one you meet — decoding a window takes longer than recording it, so 20 minutes is about eight windows at 9600 baud rather than 32.

If it stops for either reason it says which, and it tells you how to carry on: **press Mode to choose the window size again, then Capture**. Two presses, not one — a finished recording always leaves you back in FRAME — and the bytes that follow go into a **new** numbered file rather than the one it was just writing.

No bytes are lost in the gap. Your device only transmits when it is credited, and no credit goes out until the next recording arms, so it is still waiting exactly where the run stopped. What you get is one conversation split across two files, in order, with nothing missing between them.

Three things your device has to get right:

- **Catch a 10 µs pulse.** That is short: use an interrupt, an edge-triggered input or a hardware latch. A firmware polling a GPIO in a loop can miss it, and at 250000 baud 10 µs is only 2.5 bit times. It is a pulse per window, not a level you can poll once.
- **Send no more than one window per pulse** — 8 192 or 32 768 bytes depending on the mode. The pulse carries no size information, so the limit has to be built into your firmware.

- **Go quiet when it has nothing to send.** That silence is what tells the app the transmission is over.

Honest limitation: the pulse itself is verified — 4.92 V, ~10 μ s, one per arm, measured on a scope — and the loop that issues them is tested. The full round trip, with a device that actually waits for the pulse, has **not** been tested here, because the signal generator on this bench cannot be gated by the meter's trigger output. The mechanism is sound by construction; you would be the first to close the loop.

How long a recording gets you

There are two different times here and it is worth keeping them apart:

- **Signal** — how much of the line ends up in the file.
- **Waiting** — how long you stand there before the buffer is full.

Below about 19200 baud they are the same. Above it they are not, because the meter files about **100 000 samples a second** into its memory while a fast line needs it to sample faster than that. The samples it takes are properly spaced and the record is complete — you just wait longer than the amount of signal you get.

Locked rate	Signal you get	Waiting for the full 32 kB
300 Bd	1092 s	the same
1200 Bd	273 s	the same
2400 Bd	136 s	the same
4800 Bd	68 s	the same
9600 Bd	34 s	the same
19200 Bd	17 s	the same
38400 Bd	8.5 s	about 17 s
57600 Bd	5.7 s	about 17 s
115200 Bd	2.8 s	about 18 s
160000 Bd	2.0 s	about 20 s

Every row is the same 32 768 bytes — what changes is how long that takes to go past. The 8 kB window is a quarter of each figure. Then add the decoding, which does not depend on the baud rate: about two minutes for 32 kB, half a minute for 8 kB.

You do not have to fill it: press **TRIGGER** whenever you like and you keep everything up to that point.

Above about 160000 baud there is no recording mode. Pressing **Mode** and then **Capture** says so, with the number, rather than recording at a sample rate too low to be trusted. Use **FRAME**, which works to the full 250000 baud.

Reading the bytes

The hex view shows two groups of eight bytes and then the same bytes as text. The offset down the left is dimmed because it is there to navigate by, not to read. A byte that could not be recovered shows as - -.

ERR 0 does not mean the bytes are definitely right. It counts broken stop bits and parity failures. Noise landing inside a data bit flips it with nothing to detect — which is why the note row matters more than the error count.

If you see **capture began mid-byte – the first N bytes are misaligned**: on a line that never goes quiet, the capture starts wherever it starts, and if that is halfway through a byte the first few bytes are chopped up wrongly. They are shown rather than hidden, but **do not trust their values**. Everything after the first gap in the traffic is fine. On traffic with normal gaps between messages it costs **5 to 12 bytes**; on a line that never pauses it can be far more — **up to 165 bytes measured** on back-to-back text at 250000 baud, because there is no gap to resynchronise on. A device that pauses between messages never shows it at all.

Locking the baud rate, and the amber padlock

The app locks the rate for you after a good capture, which is why later captures are quicker. It only does this when it is confident: the rate landed on a standard one, enough clean frames came through, FIT was 0.9 or better, and the voltage levels were steady.

When it is not confident it leaves the padlock **amber** and offers you the **Lock Rate** button. The bytes are still shown — they just are not trusted enough to carry forward. Press **Lock Rate** to pin it anyway, or capture again.

Auto-lock never locks polarity. It is worked out fresh on every capture, on purpose: guessing it wrong once and then sticking to it is how you get a confident, completely wrong decode. (**Lock Detected** on the Options screen *does* copy the detected polarity into the fields along with everything else — but there you can see what it picked before you commit it.)

Options

SERIAL DECODE OPTIONS End App

Baud Rate 0 = auto-detect	9.60000kBd
Data Bits auto, or fix it	8
Parity auto or fixed	None
Polarity idle level	Auto
Trigger capture start	Start-bit edge
Rear BNC trig in / flow-control out	Off

Auto Detect **Lock Detected** **Apply** **Cancel**

Figure 2: Options screen

Setting	What it does
Baud Rate	0 means auto-detect. Type a number to lock that rate.
Data Bits	Leave on Auto (7/8). Only use Auto (any) if you know the device sends 9-bit words — it can make a damaged capture look tidier while being wrong.
Parity	Auto, or pin None / Even / Odd.
Polarity	Auto, Idle high (normal CMOS/TTL) or Idle low (inverted). Auto is reliable unless the line is almost never idle.

Each setting is independent — pinning one leaves the rest on auto.

Apply commits your changes and takes a capture. **Cancel** puts everything back to how it was when you opened the screen. **Auto Detect** forgets everything locked, takes a fresh capture, and locks whatever that finds — it is the quick way back when you have locked onto the wrong thing. **Lock Detected** copies the last capture's settings into the fields so you can look at them before committing.

There is no stop-bit setting. A second stop bit looks exactly like a bit of idle line, which the decoder already copes with, so a 2-stop device decodes correctly and is reported as 1 stop.

Advanced

Two settings you can ignore unless you need them:

- **Trigger** — what starts the acquisition. `Start bit` (the default) waits for the line to move. `Free run` grabs immediately. `Trigger key` waits for the front-panel TRIGGER key, as described under **The TRIGGER key** above. **If the thing you asked for never happens, the capture goes ahead anyway** after a few seconds and the note row names the one that did not arrive — `edge trigger unavailable`; `captured free-running`, or `front` for the TRIGGER key. Those bytes are real but they are not lined up with the event you wanted.
- **Rear BNC** — lets an external trigger in on the rear connector, and `FC Out` sends the flow-control credit pulse out of EXT TRIG OUT. `FC Out` is what turns a recording into an unlimited one; see **More than one window: flow control** above.

This setting affects FRAME captures. **A recording ignores it** and starts as soon as you press Capture — for a recording, the thing that decides when the device talks is the flow-control credit, not a trigger.

Files on the USB key

Two different files, and they are not the same thing:

File	Written	Holds
<code>/usb1/bytesNNN.txt</code>	automatically, on every capture	the running log — every capture's bytes as text and hex. NewLog starts a new number.
<code>/usb1/serial_NNN.txt</code>	only when you press Save	a full report of what is on screen: settings, all the notes, the whole dump.

A recording decodes into the log file, so the file holds the whole run while the panel shows only the tail of it. The note row says so when they differ.

Recording needs a USB key — the file is the point of it. Normal FRAME captures work without one; they just cannot log or Save.

When something looks wrong

What you see	What it means
line is idle (no transitions)	Nothing is crossing the threshold. Check your connection, and check the device is actually talking while you capture.
no clear logic levels	Too much noise, or the swing is too small. 1.6 V works; 60 mV is refused outright.
Wrong rate reported	The traffic may be genuinely ambiguous (below). Lock the rate if you know it.
Bytes wrong but ERR 0, with the rate locked	The device is not running at the rate you locked. The note row warns when the line disagrees by more than 4 %.
A baud rate that is not a round number	It is reporting what it measured, because the measurement was not within 2 % of a standard rate.

Some traffic is genuinely ambiguous, and no decoder can fix it. Eight 0x00 bytes at 9600 baud 7N1 make exactly the same waveform as four 0x08 bytes at 4800 baud 8N1. The app tells you a rival rate fits rather than pretending to know. Irregular gaps between bytes remove the ambiguity completely.

What it can cope with

Measured at 8 samples per bit. Past every one of these limits the app **reports** the problem rather than quietly getting it wrong, except where noted.

Condition	Works up to
Wrong clock, rate locked	±4 % . Beyond that a fast device corrupts bytes with nothing to show it
Wrong clock, auto-detect	±12 % and past — it measures instead of assuming
Jitter	±15 % of a bit; past ±20 % it can fail quietly
Noise	about 40 % of the logic swing — roughly 1.3 V on a 3.3 V line
Slow edges, long cables, weak pull-ups	filtering up to 0.7 bit times
Logic swing	1.6 V up to 5 V TTL
DC offset	anything that keeps the line inside ±10 V

Rates from 300 to 250000 baud decode byte-exact on the bench, and so do odd ones — 900, 1500, 3600, 7200 and arbitrary values like 1379, 8123 and 104857. Two exceptions are known and both are **flagged rather than silent**: long gapless payloads at 2400 baud produce one to three damaged frames in a few hundred, and one 250000-baud run matched 497 of 499 bytes with the odd byte marked. A rate within 2 % of a standard one is reported as the standard one, because real devices use standard rates.

Tolerance follows samples per bit, not baud rate. A slow line is not automatically the safe one — read SA/BIT.

If you are writing the firmware at the other end, two things help: keep your edges within $\pm 15\%$ of a bit time (that is $\pm 15.6\ \mu\text{s}$ at 9600 baud but only $\pm 1.3\ \mu\text{s}$ at 115200, so if an interrupt can delay your transmit loop, send slower), and avoid padding with 0xFF, which gives the rate detector only one narrow pulse per frame to measure. 0x00 padding is fine.

Measured figures, the full tolerance envelope, press-by-press timings and the reasoning behind the design are in **REFERENCE.md**.