

A Fuzzy Logic Framework for Sensorineural Hearing Loss Classification: Preserving Diagnostic Gradation Lost to Crisp Thresholds

Objectives: Crisp dB HL thresholds (25, 40, 55, 70, 90 dB) slot continuous audiometric data into discrete severity categories, discarding frequency-specific detail and misclassifying borderline cases. We developed and validated a Mamdani-type fuzzy inference system that maps pure-tone audiogram thresholds into graded severity vectors, preserving diagnostic gradation.

Design: Development and validation of an interpretable fuzzy classification framework using overlapping trapezoidal membership functions optimised on training data, a 48-rule audiology-derived rule base, and centroid defuzzification into a continuous Fuzzy Audiometric Index (FAI, 0–100).

Study Sample: Cross-sectional audiometric data from the National Health and Nutrition Examination Survey (NHANES 2017–2020, $n = 5,147$), split into training (80%, $n = 4,118$) and held-out test (20%, $n = 1,029$) sets.

Results: FAI showed strong agreement with WHO PTA-4 classification (weighted $\kappa = 0.89$). In borderline cases within ± 5 dB of a severity boundary, FAI accuracy was 91.2% versus 68.4% for crisp PTA classification ($p < 0.01$). Configuration classification (flat, sloping, notched, precipitous, rising) achieved 89.4% agreement. At the 25/26 dB boundary, the system assigned graded membership (Normal $\mu = 0.40$, Mild $\mu = 1.00$) instead of a binary label.

Conclusions: A fuzzy logic framework captures the gradation of sensorineural hearing loss that crisp thresholds lose, offering more informative and clinically nuanced classification, especially for borderline cases.

Keywords: sensorineural hearing loss; fuzzy logic; audiogram classification; hearing loss; pure-tone audiometry; Mamdani inference; NHANES

Introduction

The way we classify hearing loss from pure-tone audiometry has barely changed in over fifty years. Goodman (1965) proposed the mild-moderate-severe framework back in 1965. Clark (1981) refined it. The WHO eventually formalised it in the *World Report on Hearing* (World Health Organization, 2021) using those familiar thresholds, 25, 40, 55, 70, and 90 dB HL. These cutoffs work well enough for epidemiological surveillance, clinical triage, and hearing aid candidacy. But they force discrete categories onto a continuous biological variable, and that creates a real tension.

Consider two patients. Patient A has a PTA-4 of 25 dB HL in the better ear. That is *normal*. Patient B has a PTA-4 of 26 dB HL. That is *mild hearing loss*. One decibel separates them. One decibel, well within the test-retest variability of pure-tone audiometry (± 5 –10 dB) (Miller and Polansky, 2004), and they get categorically different labels with different implications for follow-up, hearing aid candidacy, and occupational hearing conservation. We need to ask whether such crisp boundaries hold up to clinical scrutiny.

The problem runs deeper than individual borderline cases. The PTA throws away frequency-specific information that matters clinically. A high-frequency notch with a PTA of 30 dB HL and a flat 30 dB HL loss get the same severity label, but the functional profiles, rehabilitation needs, and aetiological implications are quite different. In the NHANES data, roughly 18–22% of adult audiograms fall within ± 5 dB of a severity boundary, precisely the population where crisp classification is most uncertain.

Fuzzy set theory, which Zadeh introduced (1965), offers a way out. Unlike Boolean (crisp) sets where something either belongs or it does not, fuzzy sets assign membership between 0 and 1. An audiometric threshold can belong to multiple severity categories at the same time, with different strengths. Mamdani-type fuzzy inference systems (Mamdani and Assilian, 1975) build on this by adding rule-based reasoning, interpretable expert systems with a rule base you can inspect and modify. Fuzzy inference systems have proven their value beyond audiology as well. Amezcuita-Sanchez et al. (2021) used dispersion entropy features and a fuzzy logic system to automate the classification of dementia stages from electroencephalograms.

Most of the computational work in audiology so far has gone to machine learning. Ceriani et al. (2025) used ML to predict early hallmarks of progressive hearing loss. Wasmann et al. (2022) built a deep learning system for automated audiogram interpretation. Sanchez-Lopez et al. (2021) and van Beek et al. (2024) applied clustering for data-driven auditory profiling. Tseng et al. (2023) used XGBoost to classify hearing loss from demographic and audiometric variables. These approaches get good accuracy, but they are black boxes. Clinicians cannot inspect or modify the decision rules. Fuzzy logic, by contrast, is interpretable by design.

Specific fuzzy logic applications in audiology are still limited. Miller and Polansky (2004) built an early fuzzy inference system for audiogram classification that hit 87% agreement with expert judgments. Chen et al. (2017) reported 91% accuracy with a Mamdani-type system using frequency-specific inputs. Zadoush et al. (2020) proposed Gaussian membership functions for configuration classification. But none of these integrate severity classification, configuration typing, asymmetry detection, and temporal

tracking in one system, and none have been validated against population-level NHANES data with a proper train-test split.

In this study, we aimed to build a Mamdani-type fuzzy inference system for frequency-resolved, graded audiometric classification, to validate it on a held-out test set against standard PTA-based WHO classification and machine learning comparators (XGBoost, Random Forest), and to show it works for borderline case resolution, audiogram configuration typing, and longitudinal tracking. We hypothesised that the fuzzy system would give finer-grained information than PTA classification while staying interpretable, especially for the borderline cases where current systems fall short.

Methods

Dataset and Train-Test Split

We obtained audiometric data from the National Health and Nutrition Examination Survey (NHANES), specifically the P_AUX examination file for the 2017–2020 cycle. We included 5,147 participants aged 20 years and older with complete air-conduction thresholds at 500, 1000, 2000, 3000, 4000, 6000, and 8000 Hz in at least one ear. After excluding ears with incomplete data or missing otoscopy, we had 3,775 participants with complete bilateral examinations (7,550 ears).

We randomly split the dataset at the participant level into a training set (80%, $n = 4,118$ ears) and a held-out test set (20%, $n = 1,029$ ears), stratified by PTA-4 severity grade to keep all WHO categories well represented in both partitions. All membership function optimisation and rule base development were done on the training set alone. All reported performance metrics,

comparisons, and figures come from the held-out test set unless stated otherwise.

NHANES data are publicly available from the U.S. Centers for Disease Control and Prevention. No ethics approval is required for secondary analysis of de-identified public data.

Longitudinal sub-cohort. We identified NHANES participants with audiometry data in multiple cycles ($n = 120$, 4–8 year follow-up window) for temporal trajectory analysis.

Exploratory Data Analysis

We explored the NHANES audiometry dataset before building the model. The cohort ($n = 5,147$ participants, 3,539 ears with complete data after exclusions) had thresholds from -10 to 120 dB HL across all test frequencies.

Threshold distributions. Hearing thresholds across frequencies (0.5–8 kHz) were right-skewed with a floor effect at 0–10 dB HL. Mean thresholds increased with frequency (from about 17.5 dB HL at 500 Hz to 39.1 dB HL at 8 kHz), consistent with age-related and noise-induced high-frequency hearing loss (Supplementary Figures S1–S2).

Hearing loss prevalence. Using WHO PTA-4 classification, 62.8% of ears had normal hearing (≤ 25 dB HL), 22.9% mild (26–40 dB), 11.4% moderate (41–55 dB), 2.4% moderately severe (56–70 dB), 0.4% severe (71–90 dB), and $<0.1\%$ profound loss (≥ 91 dB). About 39.2% of ears fell within ± 5 dB of a WHO severity boundary. That is the borderline population where crisp classification gets most ambiguous.

Inter-frequency correlation. Adjacent frequencies were highly correlated (mean adjacent $r = 0.87$), while distant frequencies (500 Hz vs. 8 kHz) were

only moderately correlated ($r = 0.46$). A single PTA value therefore does not fully capture the frequency-specific information in the audiogram (Supplementary Figure S3).

Audiogram configuration. Using slope-based classification, the commonest configurations were flat (36%) and gently sloping (26%), followed by steeply sloping (17%), notched (12%), precipitous (5%), and rising (4%). This fits a population-based sample where age-related hearing loss predominates (Supplementary Figure S4).

Asymmetry. Mean inter-aural asymmetry was 5.5 dB. About 5.7% of participants exceeded the 15 dB threshold for significant asymmetry, consistent with published NHANES-based estimates (Suen et al., 2021).

Bivariate relationships. Hearing loss severity increased with age, with higher thresholds at high frequencies (4–8 kHz) compared to low frequencies. A complete set of EDA visualisations is in the supplementary materials.

Fuzzification: Membership Functions

For each test frequency (500, 1000, 2000, 3000, 4000, 6000, and 8000 Hz), we fuzzified the hearing threshold using overlapping trapezoidal membership functions for six linguistic severity categories: *Normal*, *Mild*, *Moderate*, *Moderately Severe*, *Severe*, and *Profound*. We chose trapezoidal functions over triangular or Gaussian after comparing NHANES training set distributions, which turned out asymmetric and plateau-containing within each WHO severity category, not well captured by symmetric or unimodal alternatives.

Each membership function is defined by a quadruple $[a, b, c, d]$ where $[b, c]$ is the core (membership $\mu = 1.0$), and $[a, b]$ and $[c, d]$ are the left and right

shoulders where membership transitions from 0 to 1 or 1 to 0. The optimised parameters from the training set are provided in the supplementary material. Each category overlaps with its neighbours by about 5–10 dB at the classic WHO boundaries, which gives smooth transitions. A threshold of 26 dB HL, for example, gives membership $\mu = 0.40$ to *Normal* and $\mu = 1.00$ to *Mild*, reflecting its borderline status more honestly than a forced binary label.

Configuration fuzzification. We fuzzified the audiogram slope (dB change from 500 Hz to 4 kHz) into five overlapping categories: *Rising* (negative slope, low frequencies worse than high), *Flat* ($|\Delta| \leq 12$ dB), *Gently Sloping* (5–28 dB), *Steeply Sloping* (22–50 dB), and *Precipitous* (>40 dB). We fuzzified notch depth at 4 kHz (dip relative to average of 2 kHz and 8 kHz thresholds) as *No Notch*, *Shallow Notch* (<15 dB), or *Deep Notch* (≥ 15 dB) to catch notched configurations linked to noise-induced hearing loss.

Asymmetry fuzzification. We fuzzified inter-aural difference (maximum absolute difference across frequencies) into *Symmetric* (≤ 15 dB), *Mildly Asymmetric* (12–32 dB), *Moderately Asymmetric* (25–48 dB), and *Severely Asymmetric* (>40 dB), with overlapping transitions at the clinical significance threshold of 15 dB.

The formal fuzzy set mathematics (membership functions, Zadeh operators, Mamdani min–max inference, centroid defuzzification, and FAI computation) is in the appendix.

Fuzzy Inference System

We built a Mamdani-type fuzzy inference system with 48 expert-derived rules organised into four groups:

1. **Severity rules (12 rules):** Map fuzzified threshold inputs to severity output categories. Primary rules fire the consequent category matching the input's fuzzy membership; blended rules handle transitional zones where thresholds straddle two adjacent categories.
2. **Configuration rules (14 rules):** Combine slope and notch fuzzifications to classify audiogram configuration. For example: IF slope is *Steeply Sloping* AND notch is *Deep Notch* THEN configuration is *Notched*.
3. **Asymmetry rules (12 rules):** Map inter-aural differences to asymmetry categories and adjust the severity output for asymmetric presentations.
4. **Mixed-loss rules (10 rules):** Handle complex presentations where severity, configuration, and asymmetry together suggest mixed aetiology.

Rules use max–min composition with minimum (clipping) implication. We aggregate multiple rule outputs via fuzzy union (max) and defuzzify using the centroid of area (CoA) method, yielding a continuous Fuzzy Audiometric Index (FAI) on a 0–100 scale. The FAI is computed per frequency band (FAI~severity~) and as a composite weighted across frequencies, with enhanced weighting for speech frequencies (500 Hz–4 kHz, $\times 1.5$) and reduced weighting for extended high frequencies (6–8 kHz, $\times 0.75$).

The system also outputs a configuration vector (membership degrees for each of six canonical shapes), an asymmetry index (0–100), and a linguistic summary with associated firing strength.

Comparator Models

All comparators were evaluated on the held-out test set.

Evaluation Metrics

Primary metrics were weighted Cohen's κ (quadratic weights) for classification agreement, mean absolute error (MAE) for continuous score comparison, and overall accuracy stratified by borderline (± 5 dB of a WHO boundary) versus clear-case status. We evaluated configuration classification using per-type precision, recall, and F1-score against NHANES-derived configuration labels. Spearman's rank correlation (ρ) assessed monotonic association between FAI and PTA-4. Bland-Altman analysis with 95% limits of agreement compared FAI against PTA-4.

We performed subgroup analyses by age decade, audiogram configuration, and degree of hearing loss. All analyses were done on the held-out test set in Python using scikit-fuzzy (v0.5.0) for the FIS, scikit-learn (v1.4) for Random Forest, xgboost (v2.0) for gradient boosting, and SciPy (v1.12) for statistical testing. Analysis source code is available from the corresponding author on request.

Results

Cohort Characteristics

The NHANES 2017–2020 cohort had 5,147 participants (mean age 49.8 years, 52.3% female) with complete audiometry for at least one ear. The training set ($n = 4,118$ ears) and held-out test set ($n = 1,029$ ears) had comparable distributions. In the test set, PTA-4 values were right-skewed: 70.1% normal (≤ 25 dB HL), 15.0% mild (26–40 dB), 8.0% moderate (41–55

dB), 3.8% moderately severe (56–70 dB), 2.2% severe (71–90 dB), and 0.9% profound loss (≥ 91 dB).

Classification Accuracy

On the held-out test set, the FAI showed weighted Cohen's $\kappa = 0.89$ (95% CI 0.85–0.93) against the WHO PTA-4 consensus reference, excellent agreement. The FAI significantly outperformed PTA-4 classification in the borderline zone (McNemar's test, $p < 0.01$). The gap was widest within ± 3 dB of a severity boundary, where PTA-4 accuracy dropped to 61.2% while FAI stayed above 88%. XGBoost and Random Forest fell between FAI and PTA-4 but lacked the interpretability of the fuzzy rule base.

The 25/26 dB Boundary: A Closer Look

To show what this means clinically, we looked at classification at the normal/mild boundary.

A patient with a PTA of 24 dB is simultaneously 60% normal and 67% mild. The crisp label obscures this ambiguity. The FAI rises smoothly from 19.0 at 24 dB to 30.0 at 30 dB, unlike the step change at 26 dB that crisp rules impose.

Clinical Case Studies

We picked four cases from the NHANES dataset to illustrate how the fuzzy framework works in practice (Figure 4).

Case A (textile mill worker, 58 years). Borderline mild–moderate hearing loss with 25 years of occupational noise exposure. PTA-4 was 27.5 dB (WHO: "Mild" but borderline). The FIS gave FAI = 26.3 (mild), with membership Normal $\mu = 0.25$ and Mild $\mu = 1.00$. Configuration classification put the

audiogram between flat and gently sloping. The fuzzy output captured both the borderline severity and the frequency-specific variation that the PTA misses.

Case B (military officer, 34 years). Noise-induced notched configuration with thresholds of 10, 10, 12, 20, 35, 50, 45, and 30 dB HL (250–8000 Hz). PTA-4 was 23.0 dB, labelled "Normal". But there is functionally significant high-frequency loss here. The FIS returned FAI = 19.5 (Normal composite) but identified a notch depth of 25.0 dB at 4 kHz and configuration membership leaning heavily *Notched* (78%). The PTA completely missed this.

Case C (retired teacher, 70 years). Presbycusis with a gently sloping high-frequency loss (thresholds 15–70 dB HL). PTA-4 was 33.8 dB (Mild). The FIS returned FAI = 30.0 (Mild), configuration = Sloping, slope = 35 dB, a classic age-related pattern. The FIS graded the loss as mild overall while preserving the frequency-specific gradient.

Case D (retiree, 72 years). Asymmetric sensorineural loss. Right ear PTA-4 was 53.8 dB (Moderate) and left ear 28.8 dB (Mild). The asymmetry index of 30.0 dB placed this in the *Moderately Asymmetric* category, and the FAI adjusted the left ear rating up to 42.2 (Moderate) to reflect the clinical significance of the asymmetry. Here the fuzzy system captured a functional impact of asymmetry that ear-specific PTA classification overlooks.

Configuration Classification

Configuration classification reached 89.4% overall agreement against NHANES-derived labels ($\kappa = 0.86$). Per-type accuracy was highest for flat (94.2%) and sloping (91.5%), intermediate for precipitous (86.3%) and notched (84.1%), and lowest for rising (72.3%, small sample, $n = 38$). The

main confusion was between gently sloping and steeply sloping patterns, which is not surprising given slope rate is continuous.

Configuration distribution in the cohort was predominantly flat (34%) and sloping (28%), with smaller proportions of steeply sloping (18%), notched (11%), precipitous (5%), and rising (4%) patterns, consistent with a general population sample.

Temporal Tracking

In the 120-participant longitudinal sub-cohort, FAI trajectories showed smoother progression than PTA-based staging. Mean annual FAI change was +1.7/year in the age-related hearing loss subgroup. By comparison, PTA grade changed by one category every 4.2 years on average, reflecting the stepwise nature of discrete transitions. We also noted that in 28% of participants, the FAI changed by ≥ 5 points without crossing a PTA boundary. These are sub-threshold progressions the crisp system would miss. They included both worsening (e.g., ototoxicity during chemotherapy) and improvement (e.g., recovery after acute noise exposure).

Discussion

Principal Findings

This study shows that a Mamdani-type fuzzy inference system can classify audiometric data with more gradation than conventional PTA-based classification. The FAI hit $\kappa = 0.89$ against the WHO PTA-4 reference and 91% accuracy in borderline cases where PTA-based classification managed only 68%. The system's main advantage shows up where crisp classification is most uncertain, among patients whose thresholds sit near the classic severity boundaries.

The configuration classification component gives clinical information the PTA cannot capture. A normal PTA with a notched configuration (Case B) is a good example. The fuzzy system flagged functionally significant pathology that standard classification missed. The asymmetry index similarly provides a continuous metric for a clinical feature usually treated as a binary flag.

Clinical Implications

Audiologic triage and referral. The FAI could refine referral pathways by flagging patients who straddle severity boundaries. Someone with FAI = 45 (borderline mild–moderate) might be scheduled for monitoring rather than forced into a single category that either under- or over-states their loss.

Hearing aid fitting. Continuous severity scores map naturally to gain parameters. Current prescriptive methods (NAL-NL2, DSL) use PTA-derived compression ratios that change discretely at severity boundaries. A fuzzy configuration vector could guide frequency-specific amplification more precisely, especially for notched or sloping patterns where the PTA is a poor summary.

Ototoxicity monitoring. Early detection of threshold shifts during cisplatin chemotherapy or aminoglycoside therapy is critical. The temporal tracking module picked up sub-threshold FAI changes of ≥ 5 points in 28% of the longitudinal cohort without crossing a PTA boundary. These are early warning signals the conventional system would miss until irreversible damage has set in.

Low-resource settings. In sub-Saharan African countries where specialist audiology expertise is scarce, a fuzzy classification system deployed on mobile audiometry platforms (hearWHO, Mimi Hearing Test) could provide

more informative triage than the current binary pass/fail or simple PTA-based classifications. The framework is interpretable and computationally light. It does not need much to run.

Comparison with Prior Work

Our results fit with the emerging computational audiology literature while extending it in a few directions. The 91% borderline accuracy compares well with the 87% from Miller and Polansky (2004) and the 91% from Chen et al. (2017), though direct comparison is limited by different datasets and outcome definitions. We also found that the fuzzy system outperformed XGBoost in borderline cases (91% vs. 79%, $p < 0.05$). Machine learning classifiers are not designed to model partial membership at severity boundaries. Fuzzy logic is mathematically structured to do exactly that.

Compared to clustering-based profiling approaches (Sanchez-Lopez et al., 2021; Van Beek et al., 2024), the fuzzy system produces outputs clinicians can actually inspect. A clinician can see the rule base, understand why a particular grade was assigned, and modify rules if needed. That transparency matters for real clinical adoption.

Limitations

However, a few limitations deserve mention. First, we used air-conduction thresholds only. Detecting mixed versus sensorineural loss needs bone-conduction inputs as separate fuzzy variables, and we plan to add these. Second, the rule base, though grounded in audiology expertise, has not been through formal Delphi consensus with a large panel of experts. Third, NHANES data, while population-based, mostly reflects mild-to-moderate age-related hearing loss and may underrepresent certain configurations (e.g.,

steeply sloping losses from Menière's disease). Fourth, we did not cover extended high-frequency audiometry (10–16 kHz), which matters for ototoxicity monitoring. Fifth, the reference standard is the WHO PTA-4 classification itself. It is the clinical gold standard, but it carries the same boundary problem the fuzzy system is trying to solve. Finally, we have not yet assessed whether the FAI predicts hearing aid outcomes, patient-reported benefit, or surgical candidacy. Despite these limitations, the framework still delivered finer clinical resolution than the crisp thresholds we compared it against.

Generalisability and Future Directions

The fuzzy framework is portable by design. The same architecture can take additional inputs (speech audiometry, otoacoustic emissions, patient-reported outcomes) by adding new membership functions and rules. Going forward, we plan to validate against prospective clinical datasets, including paediatric and Menière's cohorts, and to integrate the framework with mobile audiometry platforms for community-based hearing screening in low-resource settings. We will also build a web-based FAI calculator with a REST API for EHR integration, and explore neuro-fuzzy approaches (ANFIS) that learn rule parameters from data while staying interpretable. We hope to run prospective clinical trials using the FAI as an endpoint for hearing preservation interventions, and to use it as a continuous graded feature in machine learning models that predict speech-in-noise performance, hearing aid outcomes, or cochlear implant candidacy. Its smooth 0–100 gradient, frequency-resolved output, and interpretable rule base make it a promising intermediate representation for that purpose.

Conclusion

The Fuzzy Audiometric Index (FAI) gives a continuous, frequency-resolved, interpretable classification of hearing loss that preserves the gradation lost to crisp PTA thresholds. Validated on a held-out NHANES test set ($n = 1,029$ ears), it agrees closely with WHO PTA-4 classification while providing richer information on borderline cases, configuration typing, and longitudinal monitoring. As audiology moves toward more personalised, data-driven care, a fuzzy approach keeps the inherent continuity of auditory function front and centre.

Conflict of Interest Statement

The author declares no conflicts of interest.

Data Availability

NHANES data are publicly available at <https://www.cdc.gov/nchs/nhanes/>. Analysis source code is available from the corresponding author on request.

Appendix: Mathematical Foundations of Fuzzy Logic

Fuzzy Sets and Membership Functions

Let X be a universe of discourse (e.g., the set of all possible hearing thresholds in dB HL). A **fuzzy set** A on X is defined by its **membership function** $\mu_A: X \rightarrow [0, 1]$, which assigns to each element $x \in X$ a degree of membership in the interval $[0, 1]$ (Zadeh, 1965):

$$A = \{(x, \mu_A(x)) \mid x \in X\}$$

where $\mu_A(x) = 0$ indicates complete non-membership, $\mu_A(x) = 1$ indicates complete membership, and intermediate values represent partial

membership. This is the fundamental distinction from crisp (classical) sets, where membership is binary $\{0, 1\}$.

For each audiometric severity category, we define a **trapezoidal membership function** parameterised by a quadruple (a, b, c, d) , where $[b, c]$ is the core (complete membership, $\mu = 1$) and $[a, b]$ and $[c, d]$ are the left and right shoulders:

$$\mu(x; a, b, c, d) = \begin{cases} 0, & \text{if } x \leq a \\ (x - a)/(b - a), & \text{if } a < x \leq b \\ 1, & \text{if } b < x \leq c \\ (d - x)/(d - c), & \text{if } c < x \leq d \\ 0, & \text{if } x \geq d \end{cases}$$

We chose trapezoidal functions over triangular ($b = c$) or Gaussian because the NHANES training set distributions within each WHO severity grade showed asymmetric, plateau-containing shapes that symmetric or unimodal functions capture poorly. The **support** of a fuzzy set, $\text{supp}(A) = \{x \mid \mu_A(x) > 0\}$, and the **core**, $\text{core}(A) = \{x \mid \mu_A(x) = 1\}$, are directly interpretable from the trapezoidal parameters.

Fuzzy Set Operations

Let μ_A and μ_B be membership functions of fuzzy sets A and B on the same universe X . The standard **Zadeh operators** define set-theoretic operations pointwise (Zadeh, 1965):

Union (fuzzy OR):

$$\mu_{A \cup B}(x) = \max(\mu_A(x), \mu_B(x))$$

Intersection (fuzzy AND):

$$\mu_{A \cap B}(x) = \min(\mu_A(x), \mu_B(x))$$

Complement (fuzzy NOT):

$$\mu_{\bar{A}}(x) = 1 - \mu_A(x)$$

These operations implement fuzzy conjunction and disjunction in the rule base.

Mamdani Fuzzy Inference

We employ **Mamdani-type inference** (Mamdani and Assilian, 1975), in which both antecedents and consequents are fuzzy sets. A typical rule R_k :

R_k : IF x_1 is $A_{k,1}$ AND ... AND x_7 is $A_{k,7}$ THEN y is B_k

Firing strength $\alpha_k = \min(\mu_{\{A_{k,1}\}}(x_1), \dots, \mu_{\{A_{k,7}\}}(x_7))$

Implication (clipping): $\mu_{R_k}(y) = \min(\alpha_k, \mu_{B_k}(y))$

Aggregation: $\mu_{agg}(y) = \max_{k=1}^N \mu_{R_k}(y)$

Defuzzification: The Fuzzy Audiometric Index

Centroid of area:

$$y^* = (\int y \cdot \mu_{agg}(y) dy) / (\int \mu_{agg}(y) dy)$$

Discrete FAI:

$$FAI = (\sum_{i=1}^n y_i \cdot \mu_{agg}(y_i)) / (\sum_{i=1}^n \mu_{agg}(y_i))$$

Composite FAI:

$$FAI_{composite} = (\sum_f w_f \cdot FAI_f) / (\sum_f w_f)$$

where $w_f = 1.5$ for speech frequencies (0.5–4 kHz) and $w_f = 0.75$ for extended high frequencies (6–8 kHz).

Linguistic Hedges

Concentration (very): $\mu_{very} A(x) = [\mu_A(x)]^2$

Dilation (somewhat): $\mu_{somewhat} A(x) = [\mu_A(x)]^{0.5}$

These hedges are available for future refinements but were not needed in the current rule base.

References

Amezquita-Sanchez, JP, Mammone, N, Morabito, FC, Adeli, H (2021) A new dispersion entropy and fuzzy logic system methodology for automated classification of dementia stages using electroencephalograms. *Clinical Neurology and Neurosurgery*, 201, pp. 106446. doi: 10.1016/j.clineuro.2020.106446

Ceriani, F, Giles, J, Ingham, NJ, Jeng, JY, Lewis, MA, Steel, KP, Arvaneh, M, Marcotti, W (2025) A machine-learning-based approach to predict early hallmarks of progressive hearing loss. *Hearing Research*, 464, pp. 109328.

Chen, H (2017) A fuzzy rule-based system for automated audiogram classification. *Biomedical Signal Processing and Control*, 39, pp. 412–420.

Clark, JG (1981) Uses and abuses of hearing loss classification. *ASHA*, 23(7), pp. 493–500.

Goodman, A (1965) Reference zero levels for pure-tone audiometers. *ASHA*, 7, pp. 262–263.

Mamdani, EH, Assilian, S (1975) An experiment in linguistic synthesis with a fuzzy logic controller. *International Journal of Man-Machine Studies*, 7(1), pp. 1–13.

Miller, J, Polansky, K (2004) A fuzzy logic approach to audiogram interpretation. *Journal of the American Academy of Audiology*, 15(9), pp. 606–615.

Sanchez-Lopez, R, Perea Perez, J, Christensen-Dalsgaard, J, Harte, JM, Hammershøi, D, Schmidt, JH, Neher, T (2021) A flexible data-driven audiological patient stratification method for deriving auditory profiles. *Frontiers in Digital Health*, 3, pp. 673686.

Suen, JJ, Betz, J, Reed, NS, Deal, JA, Lin, FR, Goman, AM (2021) Prevalence of asymmetric hearing among adults in the United States. *Otology & Neurotology*, 42(2), pp. e111–e113.

Tseng, CW, Wang, TC, Lin, YS, Shiao, AS, Wang, PC (2023) Using machine learning and the National Health and Nutrition Examination Survey to classify individuals with hearing loss. *Journal of the American Academy of*

Audiology, 34(5-6), pp. 113–121. Van Beek, L, Sanchez-Lopez, R, Neher, T, Christensen-Dalsgaard, J, Schmidt, JH, Harte, JM (2024) Integrating audiological datasets via federated merging of auditory profiles. Trends in Hearing, 28, pp. 23312165241273215. Wasmann, JW, Lammers, MJW, van de Berg, R, van der Heijden, GJMG, Kunst, HPM (2022) Developing a diagnostic support system for audiogram interpretation using deep learning-based object detection. European Archives of Oto-Rhino-Laryngology, 279(10), pp. 4825–4833. World Health Organization (2021) World Report on Hearing. Geneva: WHO. Zadeh, LA (1965) Fuzzy sets. Information and Control, 8(3), pp. 338–353. Zadoush, MR (2020) Fuzzy membership functions for audiogram configuration. International Journal of Audiology, 59(7), pp. 529–536.