

Pedagogical Guide to fNIRS κ -Correction

Understanding Quantitative Near-Infrared Spectroscopy Measurements

Based on research by Neth Sagara and Amir Ahsan
Department of Physics, Irvine Valley College

March 2026

Abstract

This pedagogical guide introduces undergraduate physics students to functional near-infrared spectroscopy (fNIRS) and the κ -correction framework for recovering quantitative hemoglobin concentration measurements. The guide provides complete derivations, physical intuition, worked examples, and step-by-step implementations. We explain why the Modified Beer-Lambert Law (MBLL) systematically underestimates cortical hemoglobin changes by up to $10\times$ in layered tissues, and how two multiplicative correction factors applied at the optical-density level (κ_{DPF} and the dominant κ_{PV}) compensate for differential pathlength error and partial volume dilution, while a short-separation-regression variance-removal diagnostic (V_{SSR}) is reported per wavelength but never applied as an amplitude factor. The framework is grounded in the diffusion approximation to light transport and the Kienle et al. (1998) semi-analytical solution for two-layer tissue geometry.

Contents

I	Background and Motivation	3
1	What is fNIRS?	3
1.1	How It Works Physically	3
1.2	The Hemodynamic Response	3
1.3	Comparison with fMRI	4
2	The Quantification Problem	4
3	Three Fundamental Biases in fNIRS	4
3.1	Bias 1: Differential Pathlength Error	5
3.2	Bias 2: Partial Volume Dilution (The Dominant Effect)	5
3.3	Bias 3: Superficial Signal Contamination	6
4	From Beer's Law to the Modified Beer-Lambert Law	7
4.1	Step 1: Beer's Law for a Single Absorber	7
4.2	Step 2: Two Absorbers (HbO ₂ and HbR)	7
4.3	Step 3: Adding Scattering, The DPF	8
II	The κ Correction Framework	10
5	Why MBLL Fails in Layered Media: A Detailed Derivation	10
5.1	The Two-Layer Weighted OD Equation	10
5.2	Task-Evoked vs. Systemic Signals	11
5.3	What MBLL Assumes	11
5.4	The Attenuation Bias	11
6	The Multiplicative Correction Framework	12
6.1	Why Multiplicative Correction Works	13
7	Derivation of Each κ Factor	13
7.1	Factor 1: κ_{DPF} , Wavelength-Dependent Path Length	13
7.2	Factor 2: $\kappa_{\text{PV}} = 1/f_{\text{cortex}}$, Partial Volume Correction	13
7.3	Factor 3: $V_{\text{SSR}}(\lambda)$, the SSR Variance-Removal Diagnostic (reported, not applied)	15
7.4	Two correlations that look identical: R vs. ρ	16
8	Processing Order: Why OD-Level Corrections Matter	17
8.1	Why Not Apply Corrections After MBLL?	17
III	Light Transport Theory	19

9	From the Radiative Transfer Equation to the Diffusion Approximation	19
9.1	The Radiative Transfer Equation	19
9.2	The Diffusion Approximation	19
9.3	Validity of the Diffusion Approximation	20
10	The Semi-Infinite Homogeneous Solution	20
10.1	Geometry and Boundary Conditions	20
10.2	The Fluence Solution	21
10.3	Interpretation and Worked Example	21
11	The Two-Layer Kienle Solution	22
11.1	Setup and Boundary Conditions	23
11.2	Analytical Solution in Hankel Space	23
11.3	Numerical Implementation	24
12	Sensitivity (Jacobian) Computation	24
12.1	Definition and Derivation	24
12.2	Cortical Sensitivity Fraction	25
IV	The Analysis Pipeline Step by Step	26
13	Synthetic Data Generation	26
13.1	Hemodynamic Response Model	26
13.2	Physiological Noise	26
13.3	Forward Model in OD Domain	26
13.4	Short-Separation Channel	27
14	The Six-Step Pipeline	27
14.1	Step 1: Raw MBLL Inversion (Biased Baseline)	27
14.2	Step 2: Short-Separation Regression (SSR)	27
14.3	Step 3: Partial Volume Correction (OD Level)	28
14.4	Step 4: MBLL Inversion on Corrected ODs	28
14.5	Step 5: Correction Factor Computation	28
14.6	Step 6: Error Evaluation	29
15	Computational Implementation	29
15.1	Matrix Pseudoinverse	29
15.2	Bandpass Filtering	30
15.3	Vectorized MBLL	30
V	Results and Interpretation	31

16 Correction Factor Results	31
16.1 Summary Table	31
16.2 Physical Interpretation	31
16.3 Noise Performance with Increasing Correction	32
17 HbR Results and Wavelength Dependence	32
17.1 HbR Specific Errors	32
18 Wavelength Dependence of V_{SSR}	32
18.1 Why the Wavelength Dependence?	33
19 Sensitivity and Robustness Analysis	33
19.1 Thickness Sensitivity	33
19.2 Optical Property Sensitivity	34
19.3 Grid Convergence	34
19.4 Robustness to f_{cortex} Errors	34
19.5 Three-Layer CSF Effects	35
VI Practical Guide	37
20 Implementing κ-Correction in Your Lab	37
20.1 Step 1: Obtain Anatomical Information	37
20.2 Step 2: Choose Source-Detector Separation	37
20.3 Step 3: Include Short-Separation Channel	37
20.4 Step 4: Compute Correction Factors Offline	37
20.5 Step 5: Process Data Online or Offline	38
20.6 Step 6: Report with Uncertainty	38
21 Common Pitfalls and How to Avoid Them	38
22 Limitations and Future Directions	39
22.1 Current Limitations	39
22.2 Real-Data Demonstration (Proof of Concept)	40
22.3 Future Directions	40
23 Check Your Understanding	42
A Appendix A: Numerical Parameters	44
A.1 Optical Properties	44
A.2 Tissue Geometry	44
A.3 Monte-Carlo settings and the single source of truth	44
B Appendix B: Glossary of Symbols	45
C Appendix C: Key References	45

Part I

Background and Motivation

1 What is fNIRS?

Functional near-infrared spectroscopy (fNIRS) is an optical neuroimaging technique that non-invasively measures changes in cerebral blood oxygenation. The method exploits a fundamental biophysical principle: hemoglobin, the oxygen-carrying protein in blood, absorbs near-infrared light differently depending on its oxygenation state. Oxygenated hemoglobin (HbO₂) and deoxygenated hemoglobin (HbR) have distinct absorption spectra in the near-infrared window (650–950 nm), allowing us to distinguish them spectroscopically.

1.1 How It Works Physically

An fNIRS system places a light source (LED or laser diode) on the scalp surface. This source emits light at two specific wavelengths (typically 760 nm and 850 nm) chosen because the absorption spectra of HbO₂ and HbR cross near 800 nm (the “isosbestic point”), giving good sensitivity to both chromophores. The light enters the tissue and undergoes two competing processes:

- **Scattering:** Photons bounce off cellular structures, lipid membranes, and organelles, changing direction. In brain tissue, a photon undergoes roughly 100+ scattering events per centimetre. The reduced scattering coefficient $\mu'_s \approx 1.0 \text{ mm}^{-1}$ means photons are effectively redirected every $\sim 1 \text{ mm}$.
- **Absorption:** Photons are absorbed by chromophores (primarily hemoglobin, water, and lipids). The absorption coefficient $\mu_a \approx 0.01\text{--}0.02 \text{ mm}^{-1}$ in the NIR window, which is much smaller than μ'_s . This is precisely why the NIR window is useful: light penetrates deeply because scattering dominates over absorption.

A detector placed 25–40 mm from the source collects photons that have scattered through tissue and returned to the surface. Because of the scattering-dominated transport, these photons follow curved “banana-shaped” paths that dip into the brain cortex. The deeper the cortex, the fewer photons reach it, but at typical source-detector separations (SDS) of 30–35 mm, a significant fraction of detected photons have sampled cortical tissue.

1.2 The Hemodynamic Response

When a brain region becomes active during a cognitive or motor task, local neurons increase their metabolic demand. The brain responds with a hemodynamic response: increased blood flow to the active region, delivering fresh oxygenated hemoglobin. This causes [HbO₂] to rise by typically 1–5 μM and [HbR] to decrease by 0.3–1.5 μM over a timescale of 5–15 seconds. These concentration changes alter the optical density of the tissue, which the fNIRS detector measures as a change in received light intensity.

1.3 Comparison with fMRI

fNIRS and fMRI both measure hemodynamic responses, but they differ in important ways:

Property	fNIRS	fMRI
Spatial resolution	$\sim 1\text{--}3$ cm	$\sim 1\text{--}3$ mm
Temporal resolution	~ 10 ms	$\sim 1\text{--}2$ s
Portability	Wearable	Room-sized
Cost	\$50K–500K	\$1M–3M
Motion tolerance	Good	Poor
Measures	HbO ₂ and HbR	BOLD (mainly HbR)
Depth penetration	Cortex only	Whole brain

2 The Quantification Problem

Despite decades of fNIRS development, a critical limitation persists: most fNIRS studies report results in arbitrary units or relative changes, not absolute concentrations in micromolar (μM). This means:

- You cannot compare [HbO₂] values across different subjects or studies
- You cannot establish clinical diagnostic thresholds (e.g., “brain oxygenation below X μM indicates ischemia risk”)
- You cannot validate fNIRS measurements against gold-standard techniques (MRI, PET)
- You cannot pool data across research sites in meta-analyses

The root cause is that converting optical density changes to concentration changes requires assumptions about light propagation through tissue, assumptions that are violated in the layered human head. The standard conversion tool, the Modified Beer-Lambert Law (MBLL), assumes that tissue is homogeneous and that all concentration changes occur uniformly throughout the sampled volume. Neither assumption is correct.

This is where the κ correction framework comes in. It systematically identifies and corrects for each source of bias, recovering quantitative cortical concentrations from the biased MBLL estimates.

3 Three Fundamental Biases in fNIRS

When fNIRS data are processed using MBLL, three distinct biases cause systematic underestimation of cortical hemoglobin changes.

3.1 Bias 1: Differential Pathlength Error

Beer's Law assumes that light travels in a straight line from source to detector. In real tissue, photons scatter repeatedly, following tortuous paths that are much longer than the geometric source-detector distance. The Differential Pathlength Factor (DPF) quantifies this elongation:

$$\text{effective path length} = L \times \text{DPF}(\lambda) \quad (1)$$

where L is the source-detector separation.

DPF is wavelength-dependent because scattering and absorption both vary with wavelength. Shorter wavelengths scatter more (higher μ'_s), leading to longer effective paths and larger DPF values. For a 25-year-old adult at the forehead:

$$\text{DPF}(760 \text{ nm}) = 6.15 \quad (2)$$

$$\text{DPF}(850 \text{ nm}) = 5.09 \quad (3)$$

This $\sim 17\%$ difference means that if a processing pipeline uses a single scalar DPF (say, 6.0) for both wavelengths, it introduces a wavelength-dependent bias: concentrations are overestimated at one wavelength and underestimated at the other.

3.2 Bias 2: Partial Volume Dilution (The Dominant Effect)

This is the largest and most consequential bias. The human head is layered: scalp ($\sim 3\text{--}5$ mm), skull ($\sim 5\text{--}8$ mm), cerebrospinal fluid ($\sim 1\text{--}3$ mm), and cortex ($\sim 2\text{--}4$ mm of grey matter). When light passes from source to detector, it samples all these layers.

Task-evoked hemodynamic changes occur almost exclusively in the active cortex. The superficial layers (scalp, skull) show continuous systemic physiology (cardiac pulsations, respiration, blood pressure oscillations) but essentially no task-related changes.

The measured optical density is therefore a *weighted average* of cortical and superficial contributions:

$$\Delta\text{OD}_{\text{measured}}(\lambda) = f_{\text{cortex}}(\lambda) \cdot \Delta\text{OD}_{\text{cortex}}(\lambda) + [1 - f_{\text{cortex}}(\lambda)] \cdot \Delta\text{OD}_{\text{superficial}}(\lambda) \quad (4)$$

where $f_{\text{cortex}}(\lambda)$ is the cortical sensitivity fraction, the fraction of total channel sensitivity that originates from the cortex.

The key numbers from our study:

SDS (mm)	f_{cortex}	$\kappa_{\text{PV}} = 1/f_{\text{cortex}}$	What this means
25	0.041	24	Only $\sim 4\%$ of signal from cortex
30	0.063	16	Only $\sim 6\%$ from cortex
35	0.086	12	Only $\sim 9\%$ from cortex
40	0.112	9	$\sim 11\%$ from cortex

These values come from a converged Monte-Carlo light-transport simulation of the two-layer head (and agree with published whole-head “atlas” Monte Carlo). They are an order of magnitude smaller than an under-resolved analytical calculation suggests (a subtle numerical pitfall we discuss in the advanced sections). Since MBLL assumes all tissue is homogeneous, it treats the entire measured OD as coming from a single concentration change, so its estimate is attenuated by f_{cortex} . At 25 mm SDS, MBLL reports only $\sim 4\%$ of the true cortical concentration change) an underestimation by a factor of ~ 23 .

Key Insight

Even with perfect DPF correction and perfect superficial signal removal, MBLL still underestimates cortical concentrations by roughly an order of magnitude, because only a few percent of the channel sensitivity comes from the cortex. The partial-volume correction $\kappa_{\text{PV}} = 1/f_{\text{cortex}}$ is by far the largest factor, ranging from ~ 9 (long SDS) to ~ 23 (short SDS). A crucial consequence (developed later): because the cortical signal is so small, at short separations it can fall below the measurement-noise floor, so κ_{PV} then amplifies noise, which is why the correction is recommended at long separations (harmful at 25 mm, consistently beneficial by ~ 35 mm under the assumed noise, strongest at 38–40 mm; the crossover is instrument-dependent).

3.3 Bias 3: Superficial Signal Contamination

The superficial tissue layer contains blood vessels that exhibit strong physiological oscillations:

- Cardiac pulsation: ~ 1 Hz, amplitude $\sim 0.1 \mu\text{M}$ (HbO₂ equivalent)
- Respiration: ~ 0.25 Hz, amplitude $\sim 0.2 \mu\text{M}$
- Mayer waves (blood pressure regulation): ~ 0.1 Hz, amplitude $\sim 0.3 \mu\text{M}$

These contaminate the measured signal with noise that is *not task-related*.

Short-Separation Regression (SSR) addresses this by placing an additional detector very close to the source (SDS < 15 mm). At this short separation, the detected light barely reaches the cortex; it samples almost exclusively the superficial layer. By regressing the short-channel signal out of the long-channel signal, SSR removes the shared superficial component.

However, SSR has a limitation: if cortical and superficial signals happen to be partially correlated (e.g., task-evoked systemic responses), SSR may inadvertently remove some cortical signal along with the contamination. We quantify how much long-channel variance the short channel removed with a per-wavelength diagnostic, $V_{\text{SSR}}(\lambda) = 1/(1 - R_{\text{SS}}^2)$, but we *report* it rather than applying it as a correction — R_{SS}^2 cannot separate removed cortical signal from removed systemic signal (see Section 7.3).

4 From Beer's Law to the Modified Beer-Lambert Law

4.1 Step 1: Beer's Law for a Single Absorber

Beer's Law (also called the Beer-Lambert-Bouguer Law) describes how light is attenuated as it passes through an absorbing medium:

$$I = I_0 \cdot 10^{-\varepsilon \cdot c \cdot L} \quad (5)$$

where I_0 is the incident intensity, I is the transmitted intensity, ε is the molar extinction coefficient ($\text{cm}^{-1}/(\text{mol/L})$), c is the concentration (mol/L), and L is the path length (cm).

Taking the base-10 logarithm of both sides:

$$\log_{10} \left(\frac{I_0}{I} \right) = \varepsilon \cdot c \cdot L \quad (6)$$

The left-hand side is the **optical density** (OD), also called absorbance:

$$\text{OD} = \varepsilon \cdot c \cdot L \quad (7)$$

Physical intuition. OD = 1 means 90% of light is absorbed (only 10% transmitted). OD = 2 means 99% absorbed. Each unit of OD reduces transmitted light by a factor of 10.

The derivation is straightforward. Starting with Beer's Law:

$$I = I_0 \cdot 10^{-\varepsilon c L} \quad (8)$$

Divide both sides by I_0 :

$$\frac{I}{I_0} = 10^{-\varepsilon c L} \quad (9)$$

Take the logarithm base 10 of both sides:

$$\log_{10} \left(\frac{I}{I_0} \right) = -\varepsilon c L \quad (10)$$

Multiply by -1 and rearrange:

$$\log_{10} \left(\frac{I_0}{I} \right) = \varepsilon c L = \text{OD} \quad (11)$$

This is Beer's Law in logarithmic form, expressing optical density as the product of extinction coefficient, concentration, and path length.

4.2 Step 2: Two Absorbers (HbO_2 and HbR)

In fNIRS, two chromophores absorb simultaneously. At wavelength λ :

$$\text{OD}(\lambda) = [\varepsilon_{\text{HbO}_2}(\lambda) \cdot [\text{HbO}_2] + \varepsilon_{\text{HbR}}(\lambda) \cdot [\text{HbR}]] \cdot L \quad (12)$$

With two unknowns ($[\text{HbO}_2]$ and $[\text{HbR}]$) and one equation, we cannot solve uniquely. But with two wavelengths λ_1 and λ_2 :

$$\begin{pmatrix} \text{OD}(\lambda_1)/L \\ \text{OD}(\lambda_2)/L \end{pmatrix} = \begin{pmatrix} \varepsilon_{\text{HbO}_2}(\lambda_1) & \varepsilon_{\text{HbR}}(\lambda_1) \\ \varepsilon_{\text{HbO}_2}(\lambda_2) & \varepsilon_{\text{HbR}}(\lambda_2) \end{pmatrix} \begin{pmatrix} [\text{HbO}_2] \\ [\text{HbR}] \end{pmatrix} \quad (13)$$

This is a 2×2 linear system $\mathbf{b} = \mathbf{E} \cdot \mathbf{c}$, solved by matrix inversion: $\mathbf{c} = \mathbf{E}^{-1} \cdot \mathbf{b}$.

4.3 Step 3: Adding Scattering, The DPF

In real tissue, photons do not travel in a straight line of length L . They scatter repeatedly, following a much longer effective path. The DPF accounts for this:

$$\text{effective path length at wavelength } \lambda = L \times \text{DPF}(\lambda) \quad (14)$$

Replacing L with $L \cdot \text{DPF}(\lambda)$ in each equation:

$$\Delta\text{OD}(\lambda) = [\varepsilon_{\text{HbO}_2}(\lambda) \cdot \Delta[\text{HbO}_2] + \varepsilon_{\text{HbR}}(\lambda) \cdot \Delta[\text{HbR}]] \cdot L \cdot \text{DPF}(\lambda) \quad (15)$$

This is the **Modified Beer-Lambert Law** (MBLL). The “modified” part is the DPF factor.

Worked Example

Suppose a brain region activates, producing $\Delta[\text{HbO}_2] = +2.5 \mu\text{M}$ and $\Delta[\text{HbR}] = -0.8 \mu\text{M}$. What ΔOD do we expect at each wavelength?

Given:

- $\varepsilon_{\text{HbO}_2}(760) = 0.586$, $\varepsilon_{\text{HbR}}(760) = 1.549$ (cm^{-1}/mM)
- $\varepsilon_{\text{HbO}_2}(850) = 1.058$, $\varepsilon_{\text{HbR}}(850) = 0.691$ (cm^{-1}/mM)
- Convert to $\text{mm}^{-1}/\mu\text{M}$: multiply by 10^{-4}
- $\text{SDS} = 30$ mm, $\text{DPF}(760) = 6.15$, $\text{DPF}(850) = 5.09$

At 760 nm:

$$\Delta\text{OD}(760) = [0.586 \times 10^{-4} \times 2.5 + 1.549 \times 10^{-4} \times (-0.8)] \times 30 \times 6.15 \quad (16)$$

$$= [1.465 \times 10^{-4} - 1.239 \times 10^{-4}] \times 184.5 \quad (17)$$

$$= 0.226 \times 10^{-4} \times 184.5 \quad (18)$$

$$= 0.004170 \quad (19)$$

At 850 nm:

$$\Delta\text{OD}(850) = [1.058 \times 10^{-4} \times 2.5 + 0.691 \times 10^{-4} \times (-0.8)] \times 30 \times 5.09 \quad (20)$$

$$= [2.645 \times 10^{-4} - 0.553 \times 10^{-4}] \times 152.7 \quad (21)$$

$$= 2.092 \times 10^{-4} \times 152.7 \quad (22)$$

$$= 0.03195 \quad (23)$$

So the expected ΔOD values are about 0.0042 and 0.032, small numbers! Note how the 760 nm value is much smaller because the HbO_2 and HbR contributions nearly cancel at that wavelength. This illustrates why fNIRS requires sensitive detectors and why noise is a concern.

Part II

The κ Correction Framework

5 Why MBLR Fails in Layered Media: A Detailed Derivation

The human head is not homogeneous. It consists of discrete layers with different optical properties:

1. Scalp and skin ($d_{\text{scalp}} \approx 3\text{--}5$ mm): highly scattering, some absorption
2. Skull ($d_{\text{skull}} \approx 5\text{--}8$ mm): highly scattering, moderate absorption
3. Cerebrospinal fluid (CSF) ($d_{\text{CSF}} \approx 1\text{--}3$ mm): weakly scattering, very weakly absorbing
4. Cortex (grey matter, $d_{\text{cortex}} \approx 2\text{--}4$ mm): scattering and absorption

For simplicity, we often group the first three into a “superficial” layer and model the geometry as two layers: superficial (depth 0 to $d_{\text{sup}} \approx 12$ mm) and cortex (depth d_{sup} downward).

5.1 The Two-Layer Weighted OD Equation

Light detected by the long-channel detector samples all layers. Each layer contributes to the total optical density proportional to the light flux it receives and the time the light spends there.

Define the spatial sensitivity distribution $J(\mathbf{r})$ as the probability that a photon detected has sampled the infinitesimal volume at position $\mathbf{r}$. This is the Jacobian of the light transport: the product of the source fluence $\Phi_s(\mathbf{r})$ and detector fluence $\Phi_d(\mathbf{r})$ (the fluence field that would be created by a source at the detector location).

The total optical density is an integral over all tissue:

$$\text{OD}(\lambda) = \int_{\text{all space}} J(\mathbf{r}) \cdot \varepsilon(\mathbf{r}, \lambda) \cdot c(\mathbf{r}) d^3\mathbf{r} \quad (24)$$

For a two-layer geometry, we split this integral:

$$\text{OD}(\lambda) = \int_{\text{superficial}} J(\mathbf{r}) \cdot \varepsilon_{\text{sup}}(\lambda) \cdot c_{\text{sup}}(\lambda) d^3\mathbf{r} + \int_{\text{cortex}} J(\mathbf{r}) \cdot \varepsilon_{\text{cort}}(\lambda) \cdot c_{\text{cort}}(\lambda) d^3\mathbf{r} \quad (25)$$

Define the cortical sensitivity fraction as:

$$f_{\text{cortex}}(\lambda) = \frac{\int_{\text{cortex}} J(\mathbf{r}) d^3\mathbf{r}}{\int_{\text{all space}} J(\mathbf{r}) d^3\mathbf{r}} \quad (26)$$

This is a number between 0 and 1 representing what fraction of the total detected light has sampled the cortical region. Similarly, $f_{\text{sup}} = 1 - f_{\text{cortex}}$ is the superficial fraction.

Then:

$$\text{OD}(\lambda) = f_{\text{cortex}}(\lambda) [\varepsilon_{\text{cort}}(\lambda) \cdot c_{\text{cort}}(\lambda)] + f_{\text{sup}}(\lambda) [\varepsilon_{\text{sup}}(\lambda) \cdot c_{\text{sup}}(\lambda)] \quad (27)$$

More usefully, for changes around a baseline:

$$\Delta\text{OD}(\lambda) = f_{\text{cortex}}(\lambda) \cdot \varepsilon_{\text{cort}}(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) + f_{\text{sup}}(\lambda) \cdot \varepsilon_{\text{sup}}(\lambda) \cdot \Delta c_{\text{sup}}(\lambda) \quad (28)$$

5.2 Task-Evoked vs. Systemic Signals

This is the crucial distinction:

- **Task-evoked signals** occur only in active cortex. When a subject performs a cognitive task, $\Delta c_{\text{cort}} \neq 0$ but $\Delta c_{\text{sup}} \approx 0$ (there is no localized task-related hemodynamic response in superficial layers).
- **Systemic signals** occur everywhere. Cardiac pulsations, respiration, and slow blood pressure fluctuations cause concentration changes throughout the head. Both Δc_{cort} and Δc_{sup} are nonzero, but they reflect system-wide physiology, not localized brain activation.

For task-evoked changes, equation (28) becomes:

$$\Delta\text{OD}_{\text{task}}(\lambda) = f_{\text{cortex}}(\lambda) \cdot \varepsilon_{\text{cort}}(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) \quad (29)$$

The superficial term drops out.

5.3 What MBLL Assumes

MBLL assumes tissue is homogeneous with a single extinction coefficient $\varepsilon(\lambda)$ and uniform concentration change $\Delta c(\lambda)$:

$$\Delta\text{OD}_{\text{MBLL assumed}}(\lambda) = \varepsilon(\lambda) \cdot \Delta c_{\text{MBLL}}(\lambda) \cdot L \cdot \text{DPF}(\lambda) \quad (30)$$

Rearranging for concentration:

$$\Delta c_{\text{MBLL}}(\lambda) = \frac{\Delta\text{OD}(\lambda)}{\varepsilon(\lambda) \cdot L \cdot \text{DPF}(\lambda)} \quad (31)$$

5.4 The Attenuation Bias

Now compare. The measured ΔOD equals equation (29):

$$\Delta\text{OD}_{\text{measured}}(\lambda) = f_{\text{cortex}}(\lambda) \cdot \varepsilon(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) \quad (32)$$

But MBLL inverts this as if all the signal came from a homogeneous medium:

$$\Delta c_{\text{MBLL}}(\lambda) = \frac{f_{\text{cortex}}(\lambda) \cdot \varepsilon(\lambda) \cdot \Delta c_{\text{cort}}(\lambda)}{\varepsilon(\lambda) \cdot L \cdot \text{DPF}(\lambda)} = \frac{f_{\text{cortex}}(\lambda) \cdot \Delta c_{\text{cort}}(\lambda)}{L \cdot \text{DPF}(\lambda)} \quad (33)$$

Wait, but in the two-wavelength system we solve a matrix equation. Let's be more careful.

For two wavelengths, the true cortical signals are:

$$\Delta \text{OD}_{\text{meas}}(760) = f_{\text{cortex}}(760) \cdot [\varepsilon_{\text{HbO}_2}(760) \cdot \Delta[\text{HbO}_2]_{\text{cort}} + \varepsilon_{\text{HbR}}(760) \cdot \Delta[\text{HbR}]_{\text{cort}}] \quad (34)$$

$$\Delta \text{OD}_{\text{meas}}(850) = f_{\text{cortex}}(850) \cdot [\varepsilon_{\text{HbO}_2}(850) \cdot \Delta[\text{HbO}_2]_{\text{cort}} + \varepsilon_{\text{HbR}}(850) \cdot \Delta[\text{HbR}]_{\text{cort}}] \quad (35)$$

But MBLL performs the spectroscopic inversion as if the f_{cortex} factors are 1 everywhere:

$$\begin{pmatrix} \Delta \text{OD}_{\text{meas}}(760)/(L \cdot \text{DPF}(760)) \\ \Delta \text{OD}_{\text{meas}}(850)/(L \cdot \text{DPF}(850)) \end{pmatrix} = \mathbf{E} \begin{pmatrix} \Delta[\text{HbO}_2]_{\text{MBLL}} \\ \Delta[\text{HbR}]_{\text{MBLL}} \end{pmatrix} \quad (36)$$

where $\mathbf{E}$ is the extinction matrix. Inverting:

$$\begin{pmatrix} \Delta[\text{HbO}_2]_{\text{MBLL}} \\ \Delta[\text{HbR}]_{\text{MBLL}} \end{pmatrix} = \mathbf{E}^{-1} \begin{pmatrix} f_{\text{cortex}}(760) \cdot [\text{equation terms}]/(L \cdot \text{DPF}(760)) \\ f_{\text{cortex}}(850) \cdot [\text{equation terms}]/(L \cdot \text{DPF}(850)) \end{pmatrix} \quad (37)$$

The result is that both HbO₂ and HbR are attenuated, but not uniformly because f_{cortex} differs at the two wavelengths. Additionally, if MBLL uses a single scalar DPF for both wavelengths instead of wavelength-specific values, this introduces further attenuation errors.

The net effect: MBLL underestimates the true cortical concentrations by factors that depend on f_{cortex} , DPF accuracy, and the spectroscopic inversion.

6 The Multiplicative Correction Framework

The correction framework applies three multiplicative factors at the OD level before MBLL inversion:

$$\Delta \text{OD}_{\text{corrected}}(\lambda) = \kappa_{\text{DPF}} \times \kappa_{\text{PV}}(\lambda) \times \Delta \text{OD}_{\text{SSR}}(\lambda) \quad (38)$$

where $\Delta \text{OD}_{\text{SSR}}(\lambda)$ is the short-separation-regression-denoised optical density and:

- κ_{DPF} : corrects for DPF mismatch (wavelength-independent factor, often = 1 in modern pipelines)
- $\kappa_{\text{PV}}(\lambda) = 1/f_{\text{cortex}}(\lambda)$: corrects for partial volume dilution (wavelength-dependent; the dominant factor)
- $V_{\text{SSR}}(\lambda) = 1/(1 - R_{\text{SS}}^2(\lambda))$: an *inverse residual-variance ratio*, computed *per wavelength* and **reported as a variance-removal diagnostic only** — it is *never* applied to the data. It is not an amplitude factor: the residual standard deviation (amplitude) scales as $\sqrt{1 - R_{\text{SS}}^2}$, and R_{SS}^2 alone cannot identify how much of the removed variance was cortical (see the V_{SSR} box later in this guide). Only $\kappa_{\text{DPF}} \times \kappa_{\text{PV}}$ is applied.

Then MBLL inversion is applied to the corrected $\Delta \text{OD}_{\text{corrected}}(\lambda)$ using wavelength-specific DPF values.

6.1 Why Multiplicative Correction Works

Because Beer-Lambert law is linear in concentration, the corrections can be applied as multiplicative factors at the OD level. If the true cortical OD is ΔOD_{true} and MBLL returns an attenuated estimate $\Delta OD_{\text{MBLL}} = A \cdot \Delta OD_{\text{true}}$ (where $A < 1$ is an attenuation factor), then multiplying ΔOD_{MBLL} by $1/A$ recovers the true value.

7 Derivation of Each κ Factor

7.1 Factor 1: κ_{DPF} , Wavelength-Dependent Path Length

The DPF is wavelength-dependent. Scholkmann and Wolf (2013) give a general equation for the frontal-head DPF as a function of wavelength λ (nm) and age A (years):

$$\text{DPF}(\lambda, A) = 223.3 + 0.05624 A^{0.8493} - 5.723 \times 10^{-7} \lambda^3 + 1.245 \times 10^{-3} \lambda^2 - 0.9025 \lambda \quad (39)$$

At age $A = 25$ years and the two wavelengths used here:

$$\text{DPF}(760) = 6.15 \quad (40)$$

$$\text{DPF}(850) = 5.09 \quad (41)$$

The ratio is $6.15/5.09 \approx 1.21$, i.e. DPF is about 17% smaller at 850 nm than at 760 nm.

If a practitioner uses a scalar $\text{DPF} = 6.0$ for both wavelengths, the MBLL inversion incorrectly normalizes each wavelength by the wrong DPF. The errors are:

$$\text{Error at 760 nm : uses 6.0 instead of 6.15} \Rightarrow \text{overestimate by } 6.15/6.0 = 1.025 \quad (42)$$

$$\text{Error at 850 nm : uses 6.0 instead of 5.09} \Rightarrow \text{underestimate by } 5.09/6.0 = 0.848 \quad (43)$$

The spectroscopic inversion mixes these wavelengths; the net error cannot be corrected by a single scalar κ_{DPF} but requires wavelength-specific normalization. However, for simplicity and when using empirical DPF formulas (Scholkmann equation) consistently for both forward modeling and inversion, $\kappa_{\text{DPF}} = 1$.

7.2 Factor 2: $\kappa_{\text{PV}} = 1/f_{\text{cortex}}$, Partial Volume Correction

This is the dominant correction. From our analysis:

$$\Delta OD_{\text{measured}}(\lambda) = f_{\text{cortex}}(\lambda) \cdot \varepsilon(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) + f_{\text{sup}}(\lambda) \cdot \varepsilon_{\text{sup}}(\lambda) \cdot \Delta c_{\text{sup}}(\lambda) \quad (44)$$

For task-evoked signals, the superficial term is negligible. To recover the true cortical concentration, divide the measured OD by f_{cortex} :

$$\Delta OD_{\text{corrected}}(\lambda) = \frac{\Delta OD_{\text{measured}}(\lambda)}{f_{\text{cortex}}(\lambda)} = \varepsilon(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) + \frac{f_{\text{sup}}(\lambda)}{f_{\text{cortex}}(\lambda)} \cdot \varepsilon_{\text{sup}}(\lambda) \cdot \Delta c_{\text{sup}}(\lambda) \quad (45)$$

If superficial signals are already removed by SSR, the second term vanishes:

$$\Delta OD_{\text{PV-corrected}}(\lambda) = \frac{\Delta OD_{\text{SSR-cleaned}}(\lambda)}{f_{\text{cortex}}(\lambda)} = \varepsilon(\lambda) \cdot \Delta c_{\text{cort}}(\lambda) \quad (46)$$

The correction factor is:

$$\kappa_{\text{PV}}(\lambda) = \frac{1}{f_{\text{cortex}}(\lambda)} \quad (47)$$

Worked Example

Suppose $f_{\text{cortex}}(760) = 0.063$ at 30 mm SDS (the converged Monte-Carlo value). What is κ_{PV} ?

$$\kappa_{\text{PV}}(760) = \frac{1}{0.063} = 15.9 \quad (48)$$

This means that MBLL underestimates the cortical HbO₂ change by a factor of ~ 16 . If the true cortical change is $2.5 \mu\text{M}$, MBLL reports only $2.5/15.9 \approx 0.16 \mu\text{M}$. Multiplying by $\kappa_{\text{PV}} = 15.9$ recovers the true value, provided that tiny $0.16 \mu\text{M}$ signal is itself above the measurement noise, which (as we will see) it may not be at short separations.

Intuitive analogy: the group quiz. To build intuition for κ_{PV} and f_{cortex} , consider the following classroom analogy. A professor assigns a group quiz (out of 100 points) to a team of five students. Four of the students studied hard and each independently knows the correct answers, but the fifth student (call them the “slacker”) did not study and contributes nothing original. The quiz is administered as a group effort: each student writes answers for their assigned portion, and the group submits a single paper.

Here the **slacker represents the superficial tissue layer** (it carries blood and produces physiological signals, but contributes no task-relevant information. The **four diligent students represent the cortex**) where the real cognitive “work” happens. The **group quiz score is the fNIRS measurement**: a weighted average of all layers’ contributions.

Suppose the four diligent students each score perfectly on their sections, yielding a combined raw score of 80/100. The slacker’s section earns 0 additional points. But because the professor grades the group as a whole, the score is reported as **80 out of 100**, not the perfect 100 that reflects the diligent students’ true ability.

In this analogy:

- The fraction of the score attributable to the studying students is $f_{\text{cortex}} = 80/100 = 0.80$.
- The “dilution factor” is exactly analogous to the partial volume effect: the slacker’s dead weight pulls down the observed score.
- The correction factor is $\kappa_{\text{PV}} = 1/f_{\text{cortex}} = 100/80 = 1.25$. Multiplying the group score by κ_{PV} recovers the true score: $80 \times 1.25 = 100$.

Now consider the real fNIRS case at 30 mm SDS, where $f_{\text{cortex}} = 0.063$. This is like a group of *sixteen* students where only one studied: just $\sim 6\%$ of the answers come from the diligent student, and the rest are random guesses. The observed group score might be only $6/100$, and the correction factor $\kappa_{\text{PV}} = 1/0.063 = 16$ would recover the true individual score: $6 \times 16 \approx 100$. But notice the catch: if the diligent student's 6 points are barely distinguishable from the lucky guesses of the others (the “noise”), multiplying by 16 also multiplies that ambiguity. This is exactly why, at short separations where the cortical contribution is so small, the correction can amplify noise instead of signal, and why we apply it only at long separations where the cortical share is larger ($\sim 11\%$ at 40 mm, $\kappa_{\text{PV}} \approx 9$).

7.3 Factor 3: $V_{\text{SSR}}(\lambda)$, the SSR Variance-Removal Diagnostic (reported, not applied)

Short-separation regression removes superficial contamination by subtracting a scaled short-channel signal from the long-channel signal:

$$\Delta\text{OD}_{\text{long, SSR-cleaned}}(\lambda) = \Delta\text{OD}_{\text{long}}(\lambda) - \beta(\lambda) \cdot \Delta\text{OD}_{\text{short}}(\lambda) \quad (49)$$

where $\beta(\lambda)$ is the regression coefficient minimizing the residual power.

Define $R_{\text{SS}}^2(\lambda)$ as the fraction of variance in the *uncorrected* long-channel ΔOD explained by the short-channel regressor. Equivalently, R_{SS}^2 is the coefficient of determination of the SSR fit itself, *not* of a post-hoc regression on the SSR residual. (By construction of ordinary least squares, the SSR residual is orthogonal to the short-channel regressor, so regressing the residual on the short channel would always return $R^2 \approx 0$.) If $R_{\text{SS}}^2 = 0$, the short and long channels are uncorrelated (ideal case); if $R_{\text{SS}}^2 = 1$, they are perfectly correlated.

We report the *inverse residual-variance ratio*

$$V_{\text{SSR}}(\lambda) = \frac{1}{1 - R_{\text{SS}}^2(\lambda)} \quad (50)$$

as a per-wavelength **diagnostic of how much long-channel variance the short channel removed**. It is deliberately *not* a correction, and it is never multiplied into the data.

Key Insight

V_{SSR} is a variance ratio, not an amplitude factor. Under the ordinary-least-squares variance decomposition, $1 - R_{\text{SS}}^2$ is the ratio of the *residual variance* to the original variance. The corresponding change in *amplitude* (standard deviation) is therefore $\sqrt{1 - R_{\text{SS}}^2}$, not $1 - R_{\text{SS}}^2$ — so $1/(1 - R_{\text{SS}}^2)$ is the reciprocal of a *variance* ratio and has the wrong units to be an amplitude restoration. More fundamentally, R_{SS}^2 pools superficial variance, cortical-systemic covariance, measurement noise, and any short-channel cortical sensitivity; it *cannot* identify how much of the removed variance was cortical. There is no way to convert V_{SSR} into a cortical-amplitude correction without a physiological or forward model of the regressed components, which we do not assume.

Worked Example

In the five-subject finger-tapping data (upgraded contralateral analysis), the nearest short channel explains $R_{SS}^2(760) \approx 0.44$ of the long-channel variance at 760 nm, so

$$V_{SSR}(760) = \frac{1}{1 - 0.44} \approx 1.78, \quad (51)$$

$$\text{residual/original SD ratio} = \sqrt{1 - 0.44} \approx 0.75. \quad (52)$$

The interpretation is a *report*: “at 760 nm the short channel removed about 44% of the long-channel variance, leaving roughly 75% of the amplitude.” At 850 nm the systemic oscillations dominate the long-channel variance, so $R_{SS}^2(850) \approx 0.57$ and $V_{SSR}(850) \approx 2.34$ — a much larger diagnostic value, which is precisely why V_{SSR} is reported *per wavelength* and never averaged across wavelengths. Neither number is multiplied into the recovered concentration.

Key Insight

Why we report V_{SSR} but never apply it. $\kappa_{PV} = 1/f_{\text{cortex}}$ corrects the optical dilution of the cortical component that *survives* SSR; it does not, and cannot, reconstruct any cortical signal SSR already subtracted. V_{SSR} flags channels where SSR removed a large share of the long-channel variance (so the surviving cortical estimate should be read with caution), but converting it into an amplitude multiplier would require assuming the entire removed variance was cortical — which is false, since SSR is designed to remove the *superficial* systemic signal. Earlier drafts of this study labelled the quantity κ_{SSR} and speculated that applying it would roughly double the recovered amplitude; that counterfactual has been removed as unjustified. The applied correction is $\kappa_{DPF} \times \kappa_{PV}$ only.

7.4 Two correlations that look identical: R vs. ρ

A point that trips up almost everyone the first time: the regression statistic R in $V_{SSR} = 1/(1 - R^2)$ is *not* the same as the cortical-systemic correlation, even though both are computed with the identical Pearson formula. The formula is the same; the two signals you feed into it are different.

A Pearson correlation is just a normalized dot product, geometrically, the cosine of the angle between two signal vectors:

$$\text{Corr}(\mathbf{a}, \mathbf{b}) = \frac{\mathbf{a} \cdot \mathbf{b}}{\|\mathbf{a}\| \|\mathbf{b}\|} = \cos \theta_{ab}. \quad (53)$$

Write the long channel as $y = c + s$ (true cortical signal c plus systemic/superficial signal s) and the short channel as $x \approx s$ (it mostly samples the superficial layer). Then:

- R is the correlation between the two things we *measure*: the long channel y and the short channel x . (For simple regression, R_{SS}^2 is exactly this Pearson correlation squared.)

- ρ is the correlation between the two things we *cannot measure separately*: the cortical part c and the systemic part s .

Here is the crux. y contains s , and x *is* (approximately) s . So correlating y with x means correlating $(c + s)$ with s , and s appears in both, so you get a large R *even if c and s have nothing to do with each other* ($\rho = 0$). In words: R mostly measures “how much of the long channel is systemic,” while ρ measures “do the cortical and systemic responses rise and fall together.” Different questions, different numbers, same formula.

Key Insight

This is exactly why $V_{\text{SSR}} = 1/(1 - R^2)$ must *not* be read as an amount of cortex to hand back. In real data $R^2 \approx 0.4\text{--}0.6$, but that is large because the long channel is *systemic-dominated* (lots of s), not because SSR removed that fraction of the cortex. The cortex SSR actually nicks is governed by ρ (small for a well-designed, externally paced task), so a hypothetical $1/(1 - R^2)$ amplitude restoration would hand back far more than was lost — which is precisely why we report V_{SSR} as a variance diagnostic and never apply it. If you remember one thing: *R is between the two channels you record; ρ is between the two tissue signals you can't separate.*

8 Processing Order: Why OD-Level Corrections Matter

The order of operations is critical:

1. **Raw OD measurement:** $\Delta\text{OD}_{\text{measured}}(\lambda, t)$ at each time point
2. **Short-separation regression** (at OD level): Regress long-channel OD against short-channel OD to remove superficial contamination
3. **Partial volume correction** (at OD level): Divide by $f_{\text{cortex}}(\lambda)$ (this is the only amplitude correction applied)
4. $V_{\text{SSR}}(\lambda)$ **diagnostic** (computed, not applied): Record $1/(1 - R_{\text{SS}}^2(\lambda))$ as a per-wavelength flag of how much long-channel variance the short channel removed
5. **MBLL inversion:** Apply the extinction matrix and wavelength-specific DPF to convert corrected ODs to concentrations

8.1 Why Not Apply Corrections After MBLL?

A tempting alternative is to apply all corrections after MBLL inversion, directly on the concentration estimates. This fails because:

1. MBLL is a spectroscopic inversion mixing the two wavelengths
2. The wavelength-dependent correction $\kappa_{\text{PV}}(\lambda)$ (and the per-wavelength $V_{\text{SSR}}(\lambda)$ diagnostic) are computed

separately at each wavelength 3. If applied after inversion, the wavelength dependence is lost and the corrections cannot properly disentangle HbO₂ and HbR

Worked Example

Illustrate the problem with concrete numbers. Suppose at 30 mm SDS:

- $f_{\text{cortex}}(760) = 0.063$, $f_{\text{cortex}}(850) = 0.066$ (converged Monte-Carlo values; wavelengths have slightly different sensitivities)
- True cortical: $\Delta[\text{HbO}_2]_{\text{true}} = +2.5 \mu\text{M}$, $\Delta[\text{HbR}]_{\text{true}} = -0.8 \mu\text{M}$
- Extinction coefficients (in $\text{mm}^{-1}/\mu\text{M}$): $\varepsilon_{\text{HbO}_2}(760) = 0.0000586$, $\varepsilon_{\text{HbR}}(760) = 0.0001549$, etc.

Correct procedure (OD-level corrections):

Measured OD (before any correction) from partial volume dilution:

$$\Delta\text{OD}(760) = 0.063 \times [\text{extinction terms}] \quad (54)$$

$$\Delta\text{OD}(850) = 0.066 \times [\text{extinction terms}] \quad (55)$$

Apply partial volume correction:

$$\Delta\text{OD}_{\text{corrected}}(760) = \frac{\Delta\text{OD}(760)}{0.063} = [\text{extinction terms}] \quad (56)$$

$$\Delta\text{OD}_{\text{corrected}}(850) = \frac{\Delta\text{OD}(850)}{0.066} = [\text{extinction terms}] \quad (57)$$

MBLL inversion recovers $\Delta[\text{HbO}_2] \approx +2.5 \mu\text{M}$ and $\Delta[\text{HbR}] \approx -0.8 \mu\text{M}$. ✓

Incorrect procedure (post-MBLL corrections):

MBLL inversion on uncorrected ODs gives attenuated estimates:

$$\Delta[\text{HbO}_2]_{\text{MBLL}} \approx +0.71 \mu\text{M} \quad (58)$$

$$\Delta[\text{HbR}]_{\text{MBLL}} \approx -0.23 \mu\text{M} \quad (59)$$

Now, to apply correction, one might compute an average $\kappa_{\text{PV}} = (15.9 + 15.2)/2 \approx 15.5$ and multiply both by this scalar. But this is wrong because HbO₂ and HbR have different wavelength dependencies:

The extinction coefficients satisfy $\varepsilon_{\text{HbO}_2}(760)/\varepsilon_{\text{HbO}_2}(850) \neq \varepsilon_{\text{HbR}}(760)/\varepsilon_{\text{HbR}}(850)$. The MBLL matrix inverts wavelengths, not chromophores. Applying a scalar post-MBLL correction cannot undo the wavelength-mixing and produces biased estimates.

Part III

Light Transport Theory

9 From the Radiative Transfer Equation to the Diffusion Approximation

9.1 The Radiative Transfer Equation

The transport of photons through tissue is governed by the radiative transfer equation (RTE):

$$\frac{\partial L(\mathbf{r}, \hat{s})}{\partial t} + c \hat{s} \cdot \nabla L(\mathbf{r}, \hat{s}) + c \mu_t(\mathbf{r}) L(\mathbf{r}, \hat{s}) = c \mu_s(\mathbf{r}) \int_{\hat{s}'} L(\mathbf{r}, \hat{s}') p(\hat{s}' \rightarrow \hat{s}) d\Omega' + S(\mathbf{r}, \hat{s}) \quad (60)$$

where:

- $L(\mathbf{r}, \hat{s})$: radiance (power per unit area per unit solid angle) at position $\mathbf{r}$ in direction $\hat{s}$
- c : speed of light in the medium
- $\mu_t = \mu_a + \mu_s$: total attenuation coefficient (absorption + scattering)
- $p(\hat{s}' \rightarrow \hat{s})$: scattering phase function
- $S(\mathbf{r}, \hat{s})$: source term

This equation is accurate but extremely difficult to solve for realistic head geometry and wavelengths because the angular dependence $\hat{s}$ makes it 5-dimensional (3 spatial + 2 angular).

9.2 The Diffusion Approximation

When scattering dominates absorption ($\mu_s' \gg \mu_a$, the “diffuse limit”), the radiance becomes nearly isotropic (independent of direction). In this limit, we can replace the directional radiance with the fluence (total radiance integrated over all directions):

$$\Phi(\mathbf{r}) = \int_{\hat{s}} L(\mathbf{r}, \hat{s}) d\Omega \quad (61)$$

The diffusion approximation states that the radiance is approximately isotropic with a small anisotropic component proportional to the gradient of fluence:

$$L(\mathbf{r}, \hat{s}) \approx \frac{\Phi(\mathbf{r})}{4\pi} + \frac{3}{4\pi} \hat{s} \cdot \nabla \Phi(\mathbf{r}) \times \text{correction} \quad (62)$$

Substituting into the RTE and integrating over solid angle, we obtain the diffusion equation:

$$-\nabla \cdot [D(\mathbf{r})\nabla\Phi(\mathbf{r})] + \mu_a(\mathbf{r})\Phi(\mathbf{r}) = S(\mathbf{r}) \quad (63)$$

where $D(\mathbf{r}) = 1/[3(\mu_a(\mathbf{r}) + \mu'_s(\mathbf{r}))]$ is the diffusion coefficient. In the steady state (time-independent case):

$$-\nabla \cdot [D\nabla\Phi] + \mu_a\Phi = S \quad (64)$$

This is a 3D partial differential equation, still challenging but far more tractable than the RTE.

9.3 Validity of the Diffusion Approximation

The diffusion approximation is valid when:

1. $\mu'_s \gg \mu_a$ (scattering dominates absorption)
2. Domain size much larger than transport mean free path $\ell_t = 1/(\mu_a + \mu'_s)$
3. Not too close to boundaries or sources
4. Not in highly absorbing regions

For brain tissue in the NIR window:

- $\mu_a \approx 0.01\text{--}0.02 \text{ mm}^{-1}$
- $\mu'_s \approx 0.7\text{--}1.5 \text{ mm}^{-1}$
- Ratio: $\mu'_s/\mu_a \approx 50\text{--}100 \Rightarrow$ diffusion approximation is excellent

10 The Semi-Infinite Homogeneous Solution

10.1 Geometry and Boundary Conditions

For a semi-infinite medium (tissue occupying $z > 0$ with air at $z < 0$) with constant optical properties, the diffusion equation in steady state is:

$$-D\nabla^2\Phi(\mathbf{r}) + \mu_a\Phi(\mathbf{r}) = S(\mathbf{r}) \quad (65)$$

A point source at depth z_s below the surface deposits power P :

$$S(\mathbf{r}) = P\delta(\mathbf{r} - \mathbf{r}_s) \quad (66)$$

The boundary condition at the tissue surface ($z = 0$) is the extrapolated boundary condition:

$$\Phi(x, y, 0) = 0 \quad \text{at} \quad z = z_b = 2AD \quad (67)$$

where $A \approx 2.95$ for refractive index mismatch (typical tissue-to-air mismatch).

10.2 The Fluence Solution

The solution is obtained using the image source method. The Green's function for the semi-infinite geometry is:

$$\Phi(\mathbf{r}) = \frac{P}{4\pi D} \left[\frac{e^{-\mu_{\text{eff}} r_1}}{r_1} - \frac{e^{-\mu_{\text{eff}} r_2}}{r_2} \right] \quad (68)$$

where:

- $r_1 = \sqrt{\rho^2 + (z - z_s)^2}$ is the distance from the source at depth z_s
- $r_2 = \sqrt{\rho^2 + (z + z_s + 2z_b)^2}$ is the distance from the image source at depth $-(z_s + 2z_b)$ (below the surface)
- $\mu_{\text{eff}} = \sqrt{3\mu_a(\mu_a + \mu'_s)}$ is the effective attenuation coefficient
- $\rho = \sqrt{x^2 + y^2}$ is the radial distance from the source

The first term represents direct light spreading from the source. The second term (with opposite sign) represents light reflected by the tissue surface, enforcing the boundary condition.

10.3 Interpretation and Worked Example

The fluence decays exponentially with distance due to the $e^{-\mu_{\text{eff}} r}$ terms. Near the source ($r \rightarrow 0$), the fluence is large. Far from the source, it decays exponentially.

Worked Example

Compute the fluence at a point 30 mm from the source on the tissue surface. Use:

- $\mu_a = 0.015 \text{ mm}^{-1}$
- $\mu'_s = 1.0 \text{ mm}^{-1}$
- $D = 1/[3(\mu_a + \mu'_s)] = 1/[3(1.015)] = 0.328 \text{ mm}$
- $\mu_{\text{eff}} = \sqrt{3 \times 0.015 \times 1.015} = \sqrt{0.0456} = 0.214 \text{ mm}^{-1}$
- $z_s = 1/\mu'_s = 1 \text{ mm}$ (source depth)
- $A = 2.95$, so $z_b = 2 \times 2.95 \times 0.328 = 1.94 \text{ mm}$
- $P = 1$ (normalized)
- Point is at $\rho = 30 \text{ mm}$, $z = 0$

Compute distances:

$$r_1 = \sqrt{30^2 + (0 - 1)^2} = \sqrt{900 + 1} = \sqrt{901} = 30.017 \text{ mm} \quad (69)$$

$$r_2 = \sqrt{30^2 + (0 + 1 + 2 \times 1.94)^2} = \sqrt{900 + (4.88)^2} = \sqrt{900 + 23.81} = \sqrt{923.81} = 30.394 \text{ mm} \quad (70)$$

Compute exponentials:

$$e^{-\mu_{\text{eff}} r_1} = e^{-0.214 \times 30.017} = e^{-6.424} = 0.00159 \quad (71)$$

$$e^{-\mu_{\text{eff}} r_2} = e^{-0.214 \times 30.394} = e^{-6.504} = 0.00149 \quad (72)$$

Fluence:

$$\Phi = \frac{1}{4\pi \times 0.328} \left[\frac{0.00159}{30.017} - \frac{0.00149}{30.394} \right] \quad (73)$$

$$= \frac{1}{1.309} [0.0000529 - 0.0000490] \quad (74)$$

$$= 0.764 \times 0.0000039 \quad (75)$$

$$= 2.98 \times 10^{-6} \quad (76)$$

The fluence is very small at 30 mm distance because most photons have been absorbed or scattered away.

11 The Two-Layer Kienle Solution

The human head is not semi-infinite and homogeneous. It has distinct layers (scalp, skull, CSF, cortex) with different optical properties. For a two-layer geometry:

- Layer 1 (superficial): 0 to d_1 mm, optical properties $\{\mu_{a,1}, \mu'_{s,1}, n_1\}$
- Layer 2 (deep/cortex): d_1 to ∞ mm, optical properties $\{\mu_{a,2}, \mu'_{s,2}, n_2\}$

Kienle et al. (1998) solved the diffusion equation for this geometry using the Hankel transform method. The solution is semi-analytical: expressed in the Hankel domain (which can be solved analytically in each layer), then numerically transformed back to real space via Hankel inversion.

11.1 Setup and Boundary Conditions

In cylindrical coordinates (ρ, z) with azimuthal symmetry, the diffusion equation becomes:

$$-D \left[\frac{\partial^2 \Phi}{\partial \rho^2} + \frac{1}{\rho} \frac{\partial \Phi}{\partial \rho} + \frac{\partial^2 \Phi}{\partial z^2} \right] + \mu_a \Phi = S(\rho, z) \quad (77)$$

The Hankel transform (zeroth-order) transforms $\Phi(\rho, z)$ to $\tilde{\Phi}(q, z)$:

$$\tilde{\Phi}(q, z) = \int_0^\infty \Phi(\rho, z) J_0(q\rho) \rho d\rho \quad (78)$$

In Hankel space, derivatives with respect to ρ become multiplication by $-q^2$, and the diffusion equation becomes an ordinary differential equation in z (much simpler!):

$$-D \left[-q^2 \tilde{\Phi} + \frac{\partial^2 \tilde{\Phi}}{\partial z^2} \right] + \mu_a \tilde{\Phi} = \tilde{S}(q, z) \quad (79)$$

Rearranging:

$$\frac{\partial^2 \tilde{\Phi}}{\partial z^2} - (\mu_{\text{eff}}^2 + q^2) \tilde{\Phi} = -\frac{1}{D} \tilde{S} \quad (80)$$

where $\mu_{\text{eff}}^2 = 3\mu_a(\mu_a + \mu'_s)$.

11.2 Analytical Solution in Hankel Space

Each layer i has its own effective attenuation $\alpha_i = \sqrt{\mu_{\text{eff},i}^2 + q^2}$. The general solution is:

$$\tilde{\Phi}_i(q, z) = A_i e^{\alpha_i z} + B_i e^{-\alpha_i z} \quad (81)$$

Boundary conditions:

- At $z = 0$ (extrapolated surface): $\tilde{\Phi}_1(q, 0) = 0$ (from $\Phi = 0$ at extrapolated surface)
- At $z = d_1$ (layer interface): Continuity of Φ and of flux $-D \frac{\partial \Phi}{\partial z}$
- At $z \rightarrow \infty$ (deep tissue): $\tilde{\Phi}_2 \rightarrow 0$ (fluence vanishes at infinity)
- Source term: A point source at $z = z_s$ in layer 1 deposits power P

Solving these conditions yields expressions for A_1 , B_1 , A_2 , B_2 in terms of q , source location, and a reflection coefficient at the layer interface that depends on the refractive index mismatch (the “Da reflection coefficient”).

The final solution is inverted via:

$$\Phi(\rho, z) = \int_0^\infty \tilde{\Phi}(q, z) J_0(q\rho) q dq \quad (82)$$

This integral is evaluated numerically using Gaussian quadrature (e.g., `scipy.integrate.quad` in Python).

11.3 Numerical Implementation

The computation proceeds as:

1. For each spatial point (ρ, z) :
 - (a) For each q value (typically 0 to $\sim 1 \text{ mm}^{-1}$):
 - i. Compute layer 1 and layer 2 solutions in Hankel space using boundary conditions
 - ii. Extract $\tilde{\Phi}(q, z)$
 - iii. Multiply by $q J_0(q\rho)$ (integrand)
 - (b) Numerically integrate over q to get $\Phi(\rho, z)$
2. Repeat for all points in a 3D grid (or 2D in cylindrical symmetry)

Modern implementations use adaptive quadrature with relative error tolerances of 10^{-10} to ensure accuracy.

12 Sensitivity (Jacobian) Computation

The Jacobian $J(\mathbf{r})$ is the spatial sensitivity distribution: it quantifies which tissue regions contribute most to detected photons.

12.1 Definition and Derivation

The Jacobian is the product of source fluence and detector fluence:

$$J(\mathbf{r}) = \Phi_{\text{source}}(\mathbf{r}) \cdot \Phi_{\text{detector}}(\mathbf{r}) \quad (83)$$

where:

- $\Phi_{\text{source}}(\mathbf{r})$: fluence field created by the actual source at the source location
- $\Phi_{\text{detector}}(\mathbf{r})$: fluence field that would be created by a source at the detector location (often called the “adjoint” or “reciprocal” field)

The Jacobian is large in regions where both the source and detector “see” the tissue; these are the regions that most influence the detected signal. The shape is banana-like: highest near the source-detector midpoint at intermediate depths.

12.2 Cortical Sensitivity Fraction

Integrating the Jacobian over the cortical region:

$$J_{\text{cortex}} = \int_{z > d_{\text{sup}}} J(\mathbf{r}) d^3\mathbf{r} \quad (84)$$

and dividing by the total integral:

$$f_{\text{cortex}} = \frac{J_{\text{cortex}}}{J_{\text{total}}} = \frac{\int_{z > d_{\text{sup}}} J(\mathbf{r}) d^3\mathbf{r}}{\int_{\text{all space}} J(\mathbf{r}) d^3\mathbf{r}} \quad (85)$$

Worked Example

For $\text{SDS} = 30$ mm and superficial layer thickness $d_{\text{sup}} = 12$ mm, suppose numerical integration yields:

$$J_{\text{cortex}} = \int_{z > 12} J(\mathbf{r}) d^3\mathbf{r} = 63 \text{ (arbitrary units)} \quad (86)$$

$$J_{\text{total}} = \int_{\text{all}} J(\mathbf{r}) d^3\mathbf{r} = 1000 \quad (87)$$

Then:

$$f_{\text{cortex}} = \frac{63}{1000} = 0.063 \quad (88)$$

And $\kappa_{\text{PV}} = 1/0.063 = 15.9$. (Computing this integral reliably is subtle: a naive Hankel-transform evaluation is not numerically converged and can overstate J_{cortex} by an order of magnitude, which is why we calibrate f_{cortex} against Monte Carlo, see the box below.)

Key Insight

A numerical cautionary tale. The two-layer Jacobian is computed via a Hankel (Bessel) transform, evaluated as a sum over spatial frequencies s . The integrand contains a term $e^{\alpha(s)z}$ that *grows* with s ; at the large s needed to resolve the source/detector, it overflows and is then cancelled against a decaying term. The fluence itself is fine, but the depth-weighted *ratio* f_{cortex} is not: as you add more quadrature points it drifts ($0.0005 \rightarrow 0.62 \rightarrow 0.29 \rightarrow 0.19 \rightarrow 0.10 \rightarrow 0.085 \dots$), only slowly approaching the true ~ 0.06 – 0.09 . An “apparently converged” value at a fixed coarse setting (e.g. 0.285) can be an order of magnitude too high. The lesson: always check convergence of the *quantity you actually report*, not just the underlying field, and cross-check an analytical kernel against Monte Carlo. We use a converged two-layer Monte Carlo (`mc_2layer.py`) for all f_{cortex} values in this guide.

Part IV

The Analysis Pipeline Step by Step

13 Synthetic Data Generation

13.1 Hemodynamic Response Model

The canonical hemodynamic response function (HRF) for a task is modeled as a sum of two gamma distributions:

$$\text{HRF}(t) = \frac{t^{a_1} e^{-t/b_1}}{b_1^{a_1} \Gamma(a_1)} - c \frac{t^{a_2} e^{-t/b_2}}{b_2^{a_2} \Gamma(a_2)} \quad (89)$$

Parameters: $a_1 = 6$, $b_1 = 1$ s, $a_2 = 16$, $b_2 = 1$ s, $c = 0.1$.

This produces an HRF that rises over 4–6 seconds, peaks around 6 seconds, and then decays over 15–20 seconds.

For a block-design task (e.g., 20 seconds on, 20 seconds off), the hemodynamic response is the convolution of the block regressor with the HRF.

Peak cortical concentrations are set to:

- $\Delta[\text{HbO}_2]_{\text{peak}} = +2.5 \mu\text{M}$
- $\Delta[\text{HbR}]_{\text{peak}} = -0.8 \mu\text{M}$

These are realistic values for sensorimotor or cognitive tasks.

13.2 Physiological Noise

Superficial layers exhibit continuous oscillations:

$$\Delta c_{\text{sup}}(t) = A_{\text{cardiac}} \sin(2\pi \times 1.0 \text{ Hz} \times t) + A_{\text{respir}} \sin(2\pi \times 0.25 \text{ Hz} \times t) + A_{\text{mayer}} \sin(2\pi \times 0.1 \text{ Hz} \times t) \quad (90)$$

with amplitudes:

- Cardiac: $A_{\text{cardiac}} = 0.1 \mu\text{M}$ (HbO₂ equivalent)
- Respiration: $A_{\text{respir}} = 0.2 \mu\text{M}$
- Mayer waves: $A_{\text{mayer}} = 0.3 \mu\text{M}$

13.3 Forward Model in OD Domain

For each subject, layer, and wavelength:

$$\Delta\text{OD}(\lambda, t) = \sum_i \varepsilon_i(\lambda) \times \Delta c_i(\lambda, t) \times L_{\text{eff}}(i, \lambda) \quad (91)$$

where the sum is over chromophores (HbO₂, HbR) and layers (superficial, cortex), and $L_{\text{eff}}(i, \lambda)$ is the effective path length through layer i .

For two layers:

$$\Delta\text{OD}(\lambda, t) = f_{\text{cortex}}(\lambda) \times [\varepsilon_{\text{HbO}_2}(\lambda)\Delta[\text{HbO}_2]_{\text{cort}}(t) + \varepsilon_{\text{HbR}}(\lambda)\Delta[\text{HbR}]_{\text{cort}}(t)] \quad (92)$$

$$+ f_{\text{sup}}(\lambda) \times [\varepsilon_{\text{HbO}_2}(\lambda)\Delta[\text{HbO}_2]_{\text{sup}}(t) + \varepsilon_{\text{HbR}}(\lambda)\Delta[\text{HbR}]_{\text{sup}}(t)] \quad (93)$$

Noise is added at the OD level: $\Delta\text{OD}_{\text{noisy}} = \Delta\text{OD}_{\text{clean}} + \mathcal{N}(0, \sigma_{\text{OD}}^2)$, where $\sigma_{\text{OD}} \approx 10^{-4}$ (realistic instrumental noise).

13.4 Short-Separation Channel

A parallel short-channel (SDS ≈ 10 mm) is also generated with much higher superficial weight:

$$\Delta\text{OD}_{\text{short}}(\lambda, t) \approx 0.95 \times [\text{superficial terms}] + 0.05 \times [\text{cortical terms}] \quad (94)$$

This short-channel will be used for SSR.

14 The Six-Step Pipeline

We illustrate the pipeline at 30 mm SDS for one representative subject.

14.1 Step 1: Raw MBLI Inversion (Biased Baseline)

Apply MBLI directly to the measured $\Delta\text{OD}(\lambda)$ without any correction:

$$\begin{pmatrix} \Delta[\text{HbO}_2]_{\text{MBLI}} \\ \Delta[\text{HbR}]_{\text{MBLI}} \end{pmatrix} = \mathbf{E}^{-1} \begin{pmatrix} \Delta\text{OD}(760)/(L \times \text{DPF}(760)) \\ \Delta\text{OD}(850)/(L \times \text{DPF}(850)) \end{pmatrix} \quad (95)$$

Results at 30 mm SDS (peak task-evoked signal):

- $\Delta[\text{HbO}_2]_{\text{MBLI, peak}} = 0.71 \mu\text{M}$ (true: $2.5 \mu\text{M}$) $\Rightarrow$ 71.5% error
- $\Delta[\text{HbR}]_{\text{MBLI, peak}} = -0.23 \mu\text{M}$ (true: $-0.8 \mu\text{M}$) $\Rightarrow$ 71.2% error

The underestimation is consistent across both chromophores at this SDS.

14.2 Step 2: Short-Separation Regression (SSR)

Regress the short-channel OD against the long-channel OD at each wavelength independently:

$$\Delta\text{OD}_{\text{long, SSR-cleaned}}(\lambda) = \Delta\text{OD}_{\text{long}}(\lambda) - \beta(\lambda) \times \Delta\text{OD}_{\text{short}}(\lambda) \quad (96)$$

Compute the regression coefficient $\beta(\lambda)$ by minimizing the residual variance. A measure of systemic contamination is:

$$R_{\text{SS}}^2(\lambda) = 1 - \frac{\text{Var}[\Delta\text{OD}_{\text{long, SSR-cleaned}}]}{\text{Var}[\Delta\text{OD}_{\text{long}}]} \quad (97)$$

For our synthetic data:

- $R_{\text{SS}}^2(760) = 0.658$
- $R_{\text{SS}}^2(850) = 0.712$

These values indicate that SSR removes about 65–71% of the signal variance due to shared superficial contamination.

14.3 Step 3: Partial Volume Correction (OD Level)

Divide the SSR-cleaned OD by $f_{\text{cortex}}(\lambda)$ at each wavelength:

$$\Delta\text{OD}_{\text{PV-corrected}}(\lambda) = \frac{\Delta\text{OD}_{\text{SSR-cleaned}}(\lambda)}{f_{\text{cortex}}(\lambda)} \quad (98)$$

At 30 mm SDS: $f_{\text{cortex}}(760) = 0.063$, $f_{\text{cortex}}(850) = 0.066$ (converged Monte Carlo).

$$\kappa_{\text{PV}}(760) = \frac{1}{0.063} = 15.9, \quad \kappa_{\text{PV}}(850) = \frac{1}{0.066} = 15.2 \quad (99)$$

14.4 Step 4: MBLI Inversion on Corrected ODs

Apply MBLI to the corrected ODs using wavelength-specific DPF values:

$$\begin{pmatrix} \Delta[\text{HbO}_2]_{\text{corrected}} \\ \Delta[\text{HbR}]_{\text{corrected}} \end{pmatrix} = \mathbf{E}^{-1} \begin{pmatrix} \Delta\text{OD}_{\text{PV-corrected}}(760)/(L \times 6.15) \\ \Delta\text{OD}_{\text{PV-corrected}}(850)/(L \times 5.09) \end{pmatrix} \quad (100)$$

Results (peak task-evoked signal):

- $\Delta[\text{HbO}_2]_{\text{corrected, peak}} = 2.47 \mu\text{M}$ (true: $2.5 \mu\text{M}$) $\Rightarrow$ 1.2% error
- $\Delta[\text{HbR}]_{\text{corrected, peak}} = -0.78 \mu\text{M}$ (true: $-0.8 \mu\text{M}$) $\Rightarrow$ 2.5% error

Tremendous improvement!

14.5 Step 5: Correction Factor Computation

Compute diagnostic factors:

$$\kappa_{\text{DPF}} = 1.000 \quad (\text{by design, using true DPF for both forward and inverse}) \quad (101)$$

$$\kappa_{\text{PV}} = \frac{1}{f_{\text{cortex}}} \quad \text{mean across wavelengths} = \frac{15.9 + 15.2}{2} = 15.55 \quad (102)$$

$$V_{\text{SSR}}(\lambda) = \frac{1}{1 - R_{\text{SS}}^2(\lambda)} \quad (103)$$

$$\text{At 760 nm: } V_{\text{SSR}}(760) = \frac{1}{1-0.16} = \frac{1}{0.84} = 1.19$$

$$\text{At 850 nm: } V_{\text{SSR}}(850) = \frac{1}{1-0.902} = \frac{1}{0.098} = 10.16$$

The two wavelengths differ by nearly an order of magnitude (1.19 vs 10.16), so *no cross-wavelength mean is formed*: their average would not be a physically meaningful quantity. $V_{\text{SSR}}(\lambda)$ is reported per wavelength as a variance-removal diagnostic only.

Applied correction (the V_{SSR} diagnostics are *not* multiplied in):

$$\kappa_{\text{applied}} = \kappa_{\text{DPF}} \times \kappa_{\text{PV}} = 1.000 \times 15.55 = 15.55 \quad (104)$$

14.6 Step 6: Error Evaluation

Compute root-mean-square error (RMSE) between estimated and ground-truth peak concentrations:

$$\text{RMSE}_{\text{MBLL}} = 1.592 \text{ } \mu\text{M} \quad (105)$$

$$\text{RMSE}_{\text{corrected}} = 0.161 \text{ } \mu\text{M} \quad (106)$$

Improvement:

$$\frac{\text{RMSE}_{\text{MBLL}} - \text{RMSE}_{\text{corrected}}}{\text{RMSE}_{\text{MBLL}}} \times 100\% = \frac{1.592 - 0.161}{1.592} \times 100\% = 89.9\% \quad (107)$$

15 Computational Implementation

15.1 Matrix Pseudoinverse

The MBLL inversion is a least-squares problem solved via matrix pseudoinverse. Define:

$$\mathbf{b}(\lambda_i) = \frac{\Delta\text{OD}(\lambda_i)}{L \times \text{DPF}(\lambda_i)}, \quad i = 1, 2 \quad (108)$$

The extinction matrix is:

$$\mathbf{E} = \begin{pmatrix} \varepsilon_{\text{HbO}_2}(760) & \varepsilon_{\text{HbR}}(760) \\ \varepsilon_{\text{HbO}_2}(850) & \varepsilon_{\text{HbR}}(850) \end{pmatrix} \quad (109)$$

The solution is:

$$\begin{pmatrix} \Delta[\text{HbO}_2] \\ \Delta[\text{HbR}] \end{pmatrix} = \mathbf{E}^{-1} \begin{pmatrix} b(760) \\ b(850) \end{pmatrix} \quad (110)$$

For 2 wavelengths and 2 chromophores, $\mathbf{E}$ is square and invertible (assuming the two wavelengths are not degenerate).

In Python:

```
c = np.linalg.solve(E, b)
```

For time-series data, we stack all time points and solve vectorized:

```
C = np.linalg.solve(E, B) where B is  $2 \times N_{\text{time}}$ , yielding C also  $2 \times N_{\text{time}}$ .
```

15.2 Bandpass Filtering

Task-evoked signals occupy frequencies 0.01–0.1 Hz. Physiological noise spans 0.1–2 Hz. A bandpass filter removes high-frequency noise:

$$H(f) = \begin{cases} 1 & 0.01 < f < 0.1 \text{ Hz} \\ 0 & \text{otherwise} \end{cases} \quad (111)$$

Implementation: Butterworth filter (order 4) with cutoff frequencies $f_{\text{low}} = 0.01$ Hz and $f_{\text{high}} = 0.1$ Hz.

15.3 Vectorized MBLL

For efficiency, MBLL inversion is vectorized in the time domain:

$$\mathbf{C}(t) = \mathbf{E}^{-1}\mathbf{B}(t) \quad (112)$$

where $\mathbf{B}$ is a $2 \times N_{\text{time}}$ matrix and $\mathbf{C}$ is also $2 \times N_{\text{time}}$.

Runtime: For 3000 time points, this takes ~ 1 ms in Python/NumPy on a modern laptop.

Part V

Results and Interpretation

16 Correction Factor Results

16.1 Summary Table

SDS (mm)	f_{cortex}	κ_{PV}	$V_{\text{SSR}}(760)$	$V_{\text{SSR}}(850)$
25	0.043	23.50	1.17	9.83
30	0.064	15.69	1.19	10.16
35	0.088	11.57	1.22	10.54
38	0.101	10.04	1.22	10.65
40	0.110	9.22	1.21	10.52

(Values are the converged Monte-Carlo fractions from `fcortex_production.json`; see the single-source-of-truth note in Appendix A.)

Note on V_{SSR} . Across the study *only κ_{PV} (and κ_{DPF}) is applied to the data*; $V_{\text{SSR}}(\lambda)$ is reported *per wavelength* as a variance-removal diagnostic of residual superficial contamination, never multiplied into the corrected signal. It is listed separately at 760 and 850 nm because the two differ by nearly an order of magnitude, so no cross-wavelength mean (and no composite κ_{total}) is formed. This holds for both the synthetic validation and the real finger-tapping analysis later in this guide, for the reason given in the V_{SSR} box above (R_{SS}^2 cannot identify the cortical component of the removed variance).

16.2 Physical Interpretation

Short separations (25 mm): Very small cortical sensitivity ($\sim 4\%$), requiring a very large partial-volume factor ($\kappa_{\text{PV}} = 23$). The long channel is dominated by superficial signal, so the SSR diagnostic is large at 850 nm ($V_{\text{SSR}}(850) \approx 9.8$). *This is precisely the regime where the correction fails*: the cortical signal is so small it sits at or below the measurement-noise floor, and multiplying by 23 amplifies the noise. Short separations should be used for SSR, not for quantitative amplitude estimation.

Moderate-to-long separations (35–40 mm): The cortical fraction rises to ~ 9 – 11% , so κ_{PV} falls to ~ 9 – 11 and, crucially, the cortical signal climbs above the noise floor. This is the regime where the correction is reliable (HbO₂ RMSE reduced by $\sim 44\%$ at 38–40 mm; a 30-cohort synthetic sweep is positive in every cohort from 35 mm up). The recommended operating range is $\text{SDS} \gtrsim 35$ mm, strongest at 38–40 mm, with the exact onset set by the instrument noise.

Long separations (40 mm): The cortical fraction is largest here ($\sim 11\%$), so the partial-volume factor is smallest ($\kappa_{\text{PV}} = 9.1$) and the cortical SNR is highest, the most favourable

regime for the correction. Even so, the cortex still contributes only about a tenth of the channel sensitivity, so κ_{PV} remains an order-of-magnitude correction.

16.3 Noise Performance with Increasing Correction

Larger corrections amplify measurement noise. However, the multi-step pipeline (SSR first, then partial volume correction) manages this:

1. SSR removes systemic noise before amplification
2. Partial volume correction then amplifies the cleaned signal

This ordering is crucial for achieving the reported low RMSE.

17 HbR Results and Wavelength Dependence

17.1 HbR Specific Errors

SDS (mm)	RMSE _{MBLL}	RMSE _{corr}	Error reduction
25	0.681	0.395	42.0%
30	0.669	0.222	66.8%
35	0.653	0.163	75.0%
38	0.643	0.104	83.8%
40	0.635	0.112	82.4%

With the converged sensitivity fractions, the overall HbR RMSE improves from 0.657 to 0.226 μM (65.6%), and unlike HbO₂ the HbR correction stays positive across the full separation range, rising from 42.0% at 25 mm to 83.8% at 38 mm and holding at 82–84% at 38–40 mm. The apparent HbR advantage at short separations is a secondary noise-propagation artifact of the 2×2 MBLL inverse (the extinction matrix maps the two-wavelength OD noise onto HbO₂ and HbR with different gains, and here the propagated noise is smaller in the HbR estimate), *not* a reason to use short separations. Because the framework is scoped to long SDS by the HbO₂ noise-floor limit, we report the long-SDS HbR improvement (82–84%), where both chromophores genuinely benefit.

18 Wavelength Dependence of V_{SSR}

A critical finding: the V_{SSR} diagnostic differs significantly between wavelengths (synthetic validation, means across subjects):

SDS (mm)	$V_{\text{SSR}}(760)$	$V_{\text{SSR}}(850)$	Ratio
25	1.17	9.83	8.40
30	1.19	10.16	8.54
35	1.22	10.54	8.64
38	1.22	10.65	8.73
40	1.21	10.52	8.69

18.1 Why the Wavelength Dependence?

This pattern is driven by the wavelength-dependent extinction coefficients rather than by penetration depth alone. The systemic (superficial) signal has a characteristic HbO₂/HbR composition, and how strongly it appears in the long-channel ΔOD (and hence how much of that variance the short channel can explain) depends on the absorption weighting at each wavelength:

- At 850 nm, where HbO₂ absorption dominates ($\varepsilon_{\text{HbO}_2} > \varepsilon_{\text{HbR}}$), the systemic signal accounts for a large fraction of the long-channel ΔOD , so the short channel explains most of its variance ($R_{\text{SS}}^2 \approx 0.90$, giving $V_{\text{SSR}} \approx 10$).
- At 760 nm, where HbR absorption dominates ($\varepsilon_{\text{HbR}} \gg \varepsilon_{\text{HbO}_2}$), the short channel explains comparatively little of the long-channel variance ($R_{\text{SS}}^2 \approx 0.16$, giving $V_{\text{SSR}} \approx 1.2$).

Because the contamination structure differs so strongly between the two wavelengths, no single scalar $\bar{V}_{\text{SSR}}$ (and no cross-wavelength mean) is formed; the diagnostic is reported per wavelength.

Critical implication: Because V_{SSR} differs so much between wavelengths, collapsing it to a single cross-wavelength number would be meaningless. This is why $V_{\text{SSR}}(\lambda)$ is reported per wavelength as a diagnostic — and, for the same reason its two-wavelength values are not comparable as amplitudes, it is not applied to the data at all.

19 Sensitivity and Robustness Analysis

19.1 Thickness Sensitivity

The cortical sensitivity fraction f_{cortex} is very sensitive to the assumed superficial layer thickness. Varying d_{sup} from 8 mm to 16 mm:

Thickness (mm)	f_{cortex} (30 mm SDS)	κ_{PV}	% Change
8	0.239	4.2	−74%
10	0.122	8.2	−48%
12	0.063	15.9	0.0% (baseline)
14	0.030	33.3	+109%
16	0.015	66.2	+316%

These (converged Monte-Carlo) values show the dependence is even *stronger* than an analytical estimate suggested: a 2 mm error (12→14 mm) more than doubles κ_{PV} (+109%). Superficial-layer thickness is the single most important input to the correction, strongly motivating subject-specific anatomy (MRI, ultrasound, or an age-matched atlas).

19.2 Optical Property Sensitivity

Varying the cortical optical properties by $\pm 30\%$ (converged two-layer Monte Carlo, `mc_2layer.py`; the μ_a column uses correlated reweighting of a single photon ensemble, so it is low-variance):

Property Change	μ_a effect on κ_{PV}	μ'_s effect	Combined
-30%	-32%	-12%	-32%
Nominal	0%	0%	0%
+30%	+39%	+20%	+39%

Absorption is the dominant optical input (because $\kappa_{PV} = 1/f_{\text{cortex}}$ tracks the survival-weighted cortical pathlength). Optical-property uncertainty is therefore not negligible, but it is still smaller than the effect of superficial-layer thickness (more than +100% for a +2 mm error above), which remains the single most important input. Note that an earlier, naively evaluated analytical kernel understated this optical sensitivity (it suggested only $\pm 13\text{--}15\%$); the converged Monte Carlo is the reliable estimate.

19.3 Grid Convergence

The 3D integration for computing f_{cortex} uses an adaptive mesh. Convergence analysis:

Grid Resolution	$\Delta\rho$	f_{cortex}	Relative Error
Coarse	2.0 mm	0.2814	+1.4%
Medium	1.0 mm	0.2851	+0.2%
Fine	0.5 mm	0.2849	+0.0%
Very fine	0.25 mm	0.2849	< 0.01%

A medium resolution (1 mm) provides sufficient accuracy *in the spatial grid*. Note carefully, however, that this table varies only the (ρ, z) *spatial* mesh; it holds the Hankel-transform quadrature fixed. As the cautionary box earlier explained, it is that quadrature (not the spatial grid) that dominates the error in f_{cortex} . Spatial-grid convergence to ~ 0.285 therefore does *not* mean the value is correct; the converged answer (Monte Carlo) is ~ 0.06 at 30 mm. This is the trap the present study originally fell into.

19.4 Robustness to f_{cortex} Errors

A natural concern is: what happens if f_{cortex} is not known precisely? In practice, anatomical uncertainty guarantees some error in the computed correction factors. Our study systematically evaluated the framework’s tolerance to such errors by introducing deliberate f_{cortex} perturbations and measuring the resulting recovery accuracy.

The results are reassuring:

Error in f_{cortex}	RMSE improvement (HbO ₂)	RMSE improvement (HbR)
$\pm 10\%$	63–71%	29–43%
$\pm 30\%$	34–66%	–3–47%

Key Insight

Even with $\pm 30\%$ error in f_{cortex} (substantial anatomical uncertainty), the correction still improves HbO₂ RMSE by 34–66% at long SDS (HbR is more sensitive, turning marginally negative only at the -30% extreme). Performance is best when f_{cortex} is accurate and degrades on either side, faster for underestimates (where the inflated κ_{PV} overcorrects and amplifies noise). The practical message: once a *converged* forward model places f_{cortex} in the right band, an approximate value is far better than none, so researchers who lack subject-specific MRI can still benefit from population-averaged anatomy, *provided they work at long separations*.

19.5 Three-Layer CSF Effects

A more realistic three-layer model inserts a thin cerebrospinal fluid (CSF) layer between the skull and the cortex, evaluated within the *same* production Monte Carlo (`mc_production.py`) that supplies the two-layer fractions, using the wavelength-specific scalp and cortex optical properties (Appendix A) and inserting a CSF layer:

- Layer 1 (scalp + skull): 10 mm, wavelength-specific (760 nm: $\mu_a = 0.0128$, $\mu'_s = 1.08 \text{ mm}^{-1}$; 850 nm: $\mu_a = 0.0165$, $\mu'_s = 0.86 \text{ mm}^{-1}$)
- Layer 2 (CSF): 2 mm, $\mu_a = 0.004 \text{ mm}^{-1}$, $\mu'_s = 0.25 \text{ mm}^{-1}$ (very low scattering; 0.24 at 850 nm)
- Layer 3 (cortex): semi-infinite, wavelength-specific (760 nm: $\mu_a = 0.0192$, $\mu'_s = 0.82 \text{ mm}^{-1}$; 850 nm: $\mu_a = 0.0212$, $\mu'_s = 0.76 \text{ mm}^{-1}$)

(The total superficial depth is held at 12 mm so the cortical boundary is unchanged; a 1 mm CSF thickness check is also emitted, giving a smaller $\gamma \approx 1.3$ –1.4.)

Because CSF scatters very weakly ($\mu'_s \approx 0.25 \text{ mm}^{-1}$, versus 1.0 mm^{-1} for skull), it behaves as a low-scattering channel that lets light traverse it with little attenuation and reach cortex that would otherwise be absorbed superficially, the classic CSF “light-piping” effect. A white Monte Carlo of the three-layer geometry (the diffusion approximation is invalid in the near-transparent CSF layer) shows that inserting the CSF layer *increases* the cortical sensitivity fraction:

SDS (mm)	$f_{\text{cortex}}^{2\text{L}}$	γ (MC, 760 nm)	$f_{\text{cortex}}^{3\text{L}}$	Change
25	0.042	1.84	0.077	+84%
30	0.064	1.66	0.107	+66%
35	0.088	1.57	0.138	+57%
40	0.110	1.44	0.158	+44%

The Monte Carlo finds that the explicit CSF layer *increases* f_{cortex} by 44–84% relative to the two-layer approximation ($\gamma = f^{3\text{L}}/f^{2\text{L}} \approx 1.44$ –1.84, decreasing with SDS), in agreement with the light-piping literature.

Caution

Warning: ignoring the CSF layer degrades (but at long SDS does not reverse) the correction.

If the true tissue contains a CSF layer but corrections are computed with a two-layer model, the two-layer model *underestimates* f_{cortex} (e.g., ≈ 0.10 versus ≈ 0.15 at 38 mm). Its $\kappa_{\text{PV}} = 1/f_{\text{cortex}}$ is therefore too *large* and partially overcorrects. At long separations the net effect is a reduced but still positive benefit (HbO₂ improvement +34.0% vs. +69.1% matched; HbR +14.7% vs. +37.8%). The framework degrades gracefully rather than catastrophically, but including CSF (via the ratio γ) is recommended when anatomy is known.

Caution

Modelling note. The cortical-sensitivity *ratio* γ above is robust to the CSF Monte Carlo's isotropic-scattering approximation, so we use only the ratio and apply it to the converged Henyey–Greenstein two-layer fractions ($f^{3\text{L}} = \gamma f^{2\text{L}}$). Even with light-piping the cortical fraction stays small (< 0.15), so κ_{PV} remains large. For reference, an earlier diffusion finite-difference estimate gave a spurious $\sim 99\%$ *reduction*; that was a numerical artifact, because diffusion theory fails in the near-transparent CSF layer, which is exactly why Monte Carlo is needed here.

Part VI

Practical Guide

20 Implementing κ -Correction in Your Lab

20.1 Step 1: Obtain Anatomical Information

Ideally, acquire or estimate the superficial layer thickness for your subject population:

1. **MRI:** Measure scalp, skull, and CSF thicknesses from structural images. Most reliable but expensive.
2. **Age-based tables:** Use population-averaged values from literature (e.g., scalp = 4 mm, skull = 7 mm for adults).
3. **Ultrasound:** Portable and cheap, can measure scalp and skull thickness in clinic.

For a first study, use standard values: $d_{\text{sup}} = 12$ mm (scalp 4 + skull 7 + CSF 1 mm).

20.2 Step 2: Choose Source-Detector Separation

- **If cortical sensitivity is critical:** Use $\text{SDS} \geq 35$ mm. Requires larger corrections but better cortical specificity.
- **If noise and SNR are limiting:** Use $\text{SDS} = 30$ mm. Good balance.
- **If only relative changes matter:** $\text{SDS} = 25$ mm is acceptable but requires larger corrections and is more sensitive to depth assumptions.

20.3 Step 3: Include Short-Separation Channel

Place a detector at $\text{SDS} \approx 10$ mm for every measurement location. This is essential for SSR.

20.4 Step 4: Compute Correction Factors Offline

For each SDS and anatomical configuration:

1. Solve the two-layer Kienle equation via Hankel transform
2. Compute fluence fields for source and detector
3. Integrate Jacobian over cortical and total volumes
4. Derive $f_{\text{cortex}}(\lambda)$ and $\kappa_{\text{PV}}(\lambda) = 1/f_{\text{cortex}}(\lambda)$

This computation takes ~ 0.1 s per channel on a modern laptop and can be done once, offline, before data acquisition.

20.5 Step 5: Process Data Online or Offline

1. Measure raw OD at two wavelengths, long and short channels
2. Apply bandpass filter (0.01–0.1 Hz) to remove high-frequency noise
3. Perform SSR: regress long-channel OD against short-channel OD at each wavelength
4. Apply wavelength-specific κ_{PV} at OD level (the only amplitude correction)
5. Compute and record the per-wavelength $V_{SSR}(\lambda)$ diagnostic (do *not* apply it)
6. Invert with MBLL using wavelength-specific DPF
7. Extract HbO₂ and HbR concentrations and peak values

20.6 Step 6: Report with Uncertainty

Always report:

- Mean $\pm$ SD peak HbO₂ and HbR concentrations in micromolar
- Assumed anatomical parameters (thickness, DPF values)
- SDS used
- Acknowledgment of sensitivity to thickness and optical properties

21 Common Pitfalls and How to Avoid Them

Caution

Pitfall 1: Using scalar DPF for both wavelengths.

Many commercial systems apply a single $DPF = 6.0$ to both 760 nm and 850 nm. This introduces $\sim 17\%$ wavelength-dependent errors.

Fix: Use Scholkmann formula or wavelength-specific empirical DPF values:

$$DPF(760) = 6.15 \quad (113)$$

$$DPF(850) = 5.09 \quad (114)$$

Normalize each wavelength independently in MBLL.

Caution

Pitfall 2: Applying corrections after MBLL inversion.

Post-MBLL corrections lose wavelength information and fail to properly unmix HbO₂ and HbR.

Fix: Apply all corrections at the OD level before MBLL inversion.

Caution**Pitfall 3: Ignoring partial volume without short-separation regression.**

Some groups use SSR alone, thinking it solves the partial volume problem. SSR removes superficial noise but does NOT correct for the attenuation of cortical signal.

Fix: Combine SSR (removes systemic noise) and PV correction (corrects attenuation of cortical signal).

Caution**Pitfall 4: Using incorrect layer thickness.**

Assuming $d_{\text{sup}} = 8$ mm when the true value is 12 mm causes 30% error in corrections.

Fix: If possible, measure thickness from MRI or ultrasound for each subject. If not, perform sensitivity analysis showing how results change with ± 4 mm variations in thickness.

Caution**Pitfall 5: Treating the SSR R^2 as an amplitude correction, or averaging it across wavelengths.**

$V_{\text{SSR}}(\lambda) = 1/(1 - R_{\text{SS}}^2)$ is an inverse residual-*variance* ratio, not an amplitude factor (the residual-to-original standard-deviation ratio is $\sqrt{1 - R_{\text{SS}}^2}$, still not a task-response amplitude ratio), and its value differs by nearly an order of magnitude between 760 and 850 nm, so a cross-wavelength average is meaningless.

Fix: Report $V_{\text{SSR}}(\lambda)$ per wavelength as a variance-removal *diagnostic* and never multiply it into the recovered concentration; apply only $\kappa_{\text{DPF}} \times \kappa_{\text{PV}}$.

22 Limitations and Future Directions

22.1 Current Limitations

1. **Tissue model:** Two-layer model is simplified. Real anatomy includes CSF with very different optical properties. Three-layer corrections are needed for highest accuracy.
2. **Assumes homogeneous layers:** Real tissue has spatial variation (e.g., bone density variations in skull). Monte Carlo simulation would capture this but is much slower.
3. **Assumes task-evoked, not systemic signals:** If a cognitive task causes global blood pressure changes (task-evoked systemic response), SSR may remove some cortical signal. The V_{SSR} diagnostic flags channels where this is likely (high R_{SS}^2), but cannot by itself disentangle the ambiguity.
4. **Requires anatomical knowledge:** Thickness and optical properties must be known or estimated. Subject-specific anatomical imaging (MRI, ultrasound) is not always available.

22.2 Real-Data Demonstration (Proof of Concept)

Application of the pipeline to the publicly available BIDS-NIRS-Tapping finger-tapping dataset (Luke et al., 2021; SDS $\approx$ 38 mm) demonstrates the framework on real data as a proof of concept. There is no concurrent ground truth here, so this tests whether the recovered amplitudes are physiologically plausible, not whether they are quantitatively accurate. Importantly, the data sit at 38 mm, inside the long-separation regime where the correction is reliable. The pipeline proceeds in ordered stages at the OD level:

1. **Short-separation regression (SSR):** Per-wavelength OD-level regression against the short-channel reference removes systemic (extracerebral) signal components.
2. **$V_{\text{SSR}}(\lambda)$ diagnostic:** The regression R^2 yields $V_{\text{SSR}}(\lambda) = 1/(1 - R_{\text{SS}}^2(\lambda))$ (group means 760: 1.78, 850: 2.34 for this dataset). This is *reported as a variance-removal diagnostic of residual superficial contamination but never applied* (see the V_{SSR} box earlier): R_{SS}^2 cannot identify the cortical component of the removed variance, so it is not converted into an amplitude factor.
3. **Partial-volume (κ_{PV}) correction (the applied amplitude correction):** The two-layer Monte Carlo gives $f_{\text{cortex}} \approx 0.10$ at 38 mm. But a *real* adult head also contains a thin cerebrospinal-fluid (CSF) layer, which scatters light very weakly and “pipes” it down toward the cortex (see the CSF section). This raises the cortical fraction by the light-piping factor $\gamma \approx 1.5$, to a three-layer value $f_{\text{cortex}}^{\text{3L}} \approx 0.15\text{--}0.16$. Because these are real heads, we use this CSF-augmented value, giving a per-channel $\kappa_{\text{PV}} \approx 6.8$ (760 nm: $1/0.151 \approx 6.6$; 850 nm: $1/0.148 \approx 6.8$).
4. **MBLL inversion:** Standard dual-wavelength conversion from corrected ΔOD to $\Delta[\text{HbO}_2]$ and $\Delta[\text{HbR}]$.

Applying the per-channel κ_{PV} to the SSR-denoised data recovers physiologically plausible amplitudes. For Subject 01 (contralateral, window-mean) the uncorrected HbO_2 of $+0.144 \mu\text{M}$ rises to $+0.649 \mu\text{M}$ ($\kappa_{\text{PV}} = 6.71$) and HbR from -0.030 to $-0.181 \mu\text{M}$; across all five subjects the group-mean corrected HbO_2 is $0.92 \pm 0.34 \mu\text{M}$ ($\sim 7\times$ the uncorrected $0.124 \mu\text{M}$) and HbR $-0.047 \rightarrow -0.343 \mu\text{M}$. These magnitudes are physiologically plausible for a finger-tapping task. (Using the two-layer $f_{\text{cortex}} \approx 0.10$ instead (ignoring CSF) gives the larger $\kappa_{\text{PV}} \approx 9.7$ and $\approx 1.3 \mu\text{M}$, an upper bound; the physiological estimate is thus bracketed $\approx 0.9\text{--}1.3 \mu\text{M}$.) This is the payoff of using a *converged*, CSF-augmented f_{cortex} : an under-converged ≈ 0.78 would make κ_{PV} do almost nothing. Consistent with the statistical clarification, V_{SSR} is reported as a variance diagnostic and is *not* applied.

22.3 Future Directions

1. **Three-layer Kienle solution** with CSF layer for full anatomical accuracy.
2. **Hybrid CSF correction:** A particularly promising approach is to compute f_{cortex} rapidly using the two-layer Hankel transform (~ 0.1 s per channel) and then multiply by an SDS-dependent correction factor $\gamma(d_{\text{SDS}}, d_{\text{CSF}})$ from a precomputed lookup table.

The adjusted sensitivity $f_{\text{cortex}}^{\text{adj}} = \gamma \times f_{\text{cortex}}^{2\text{L}}$ would retain the speed of the analytical two-layer solution while accounting for the CSF light-guiding effect. Our three-layer analysis shows that γ varies systematically: approximately 1.84 at 25 mm, 1.66 at 30 mm, 1.57 at 35 mm, and 1.44 at 40 mm SDS (760 nm).

3. **Adaptive thickness estimation** from temporal statistics of SSR.
4. **Bayesian inference** combining MBLL data with prior distributions on tissue properties.
5. **Machine learning models** trained on Monte Carlo forward models to predict f_{cortex} given measurement geometry and partial optical property measurements.
6. **In vivo validation** with concurrent fMRI or invasive recordings as ground truth to establish the framework's clinical performance.
7. **Real-time feedback:** Deploy corrections online during experiments for closed-loop protocols.

23 Check Your Understanding

These problems work through the core ideas of the framework. Try each before reading the solution.

1. From f_{cortex} to κ_{PV} . A channel at 30 mm has $f_{\text{cortex}} = 0.06$. (a) What is κ_{PV} ? (b) If the raw MBLL estimate of cortical $\Delta[\text{HbO}_2]$ is $0.15 \mu\text{M}$, what is the corrected estimate?

Solution: (a) $\kappa_{\text{PV}} = 1/0.06 \approx 16.7$. (b) $0.15 \times 16.7 \approx 2.5 \mu\text{M}$.

2. Why not always use short separations? Short channels are cheap and common. Using the noise-floor idea, explain in one or two sentences why you should *not* apply κ_{PV} at 25 mm.

Solution: At 25 mm $f_{\text{cortex}} \approx 0.04$, so the cortical optical-density signal ($\sim 10^{-3}$) sits at or below the measurement-noise floor ($\text{SNR} < 1$). Multiplying by $\kappa_{\text{PV}} \approx 24$ amplifies the noise as much as the signal, so the “corrected” estimate is dominated by amplified noise rather than recovered cortex.

3. The CSF correction. The two-layer model gives $f_{\text{cortex}} = 0.10$ at 38 mm, and the CSF light-piping ratio is $\gamma = 1.5$. (a) What is the three-layer f_{cortex} ? (b) How does κ_{PV} change? (c) Does the recovered amplitude go up or down?

Solution: (a) $0.10 \times 1.5 = 0.15$. (b) κ_{PV} drops from $1/0.10 = 10$ to $1/0.15 \approx 6.7$. (c) Down: a smaller κ_{PV} multiplies the signal by less, so the recovered amplitude decreases (in the paper, the two-layer $\approx 1.3 \mu\text{M}$ upper bound drops to the CSF-augmented $\approx 0.9 \mu\text{M}$).

4. R vs. ρ . A subject’s long and short channels have $R_{\text{SS}}^2 = 0.64$. A classmate concludes that SSR must have removed 64% of the *cortical* signal. Are they right?

Solution: No. $R_{\text{SS}}^2 = 0.64$ means the short channel explains 64% of the *long-channel* variance, mostly because the long channel is dominated by the same systemic signal the short channel sees. The cortical loss is set by ρ (the cortical–systemic correlation), not by R^2 , and is usually much smaller. This is exactly why $V_{\text{SSR}} = 1/(1 - R^2)$ is reported as a variance-removal diagnostic and never applied as an amplitude restoration.

5. Wavelength bookkeeping. Why must $\kappa_{\text{PV}}(\lambda)$ be applied to each wavelength’s ΔOD *before* the MBLL inversion, rather than to the final $\Delta[\text{HbO}_2]$?

Solution: The MBLL inversion mixes the two wavelengths to separate HbO_2 from HbR . Because f_{cortex} differs at 760 and 850 nm, the correction is wavelength-specific; applying it *after* the inversion would discard the per-wavelength information and could not properly disentangle the two chromophores (Section 8 on processing order).

6. An order-of-magnitude sanity check. You read that a 25 mm channel has $f_{\text{cortex}} = 0.55$. Based on converged Monte Carlo and atlas studies, is this plausible?

Solution: No. Converged Monte Carlo and atlas studies (e.g. Strangman et al. 2013) put the brain sensitivity fraction at $\sim 4\text{--}8\%$ at 25–30 mm, an order of magnitude smaller. A value

of 0.55 is a red flag for an under-converged analytical calculation (the trap discussed in the numerical cautionary tale).

7. Putting it together. In one paragraph, summarize the full corrected pipeline for a single long channel: what is removed, what is multiplied in, what is only reported, and why the result is trustworthy only at long SDS.

Solution: Band-pass the optical density; regress out the short channel (SSR) to remove systemic contamination; multiply the SSR-cleaned ΔOD by $\kappa_{PV} = 1/f_{\text{cortex}}$ (CSF-augmented) at each wavelength to undo partial-volume dilution; invert with the MBLL using wavelength-specific DPFs. V_{SSR} is computed and reported as a variance-removal diagnostic of residual contamination but *not* applied. The result is trustworthy at long SDS (consistently beneficial by ~ 35 mm, strongest at 38–40 mm), because there the cortical signal is above the noise floor; at short SDS the same correction amplifies noise. The exact onset is set by the instrument noise.

A Appendix A: Numerical Parameters

A.1 Optical Properties

Extinction coefficients and reduced scattering coefficients at 760 and 850 nm (tissue at body temperature):

Chromophore/Tissue	$\varepsilon(760)$	$\varepsilon(850)$	Unit	Source
HbO ₂	0.586	1.058	cm ⁻¹ /mM	Prahl/OMLC
HbR	1.549	0.691	cm ⁻¹ /mM	Prahl/OMLC
Scalp/skin	0.0120	0.0090	mm ⁻¹	Jacques 2013
Skull	0.0140	0.0110	mm ⁻¹	Okada 2003
CSF	0.0010	0.0008	mm ⁻¹	Simpson 1998
μ'_s (scalp/skull)	1.0–1.2	0.85–1.0	mm ⁻¹	Jacques 2013
μ'_s (brain grey matter)	1.0–1.2	0.85–1.0	mm ⁻¹	Okada 2003
μ'_s (CSF)	0.01	0.01	mm ⁻¹	Simpson 1998

A.2 Tissue Geometry

Standard adult head geometry for frontal measurement:

Layer	Thickness (mm)	Refractive Index	Composition
Scalp	3–5	1.40	Skin, hair, connective tissue
Skull	5–8	1.40	Bone (compact and trabecular)
CSF (dura/arachnoid)	1–3	1.33	Cerebrospinal fluid
Cortex (grey matter)	2–4	1.40	Neuronal tissue
Total superficial	10–14	,	Typically 12 mm for adults

A.3 Monte-Carlo settings and the single source of truth

All cortical sensitivity fractions used anywhere in this study come from *one* converged production Monte Carlo, written to a single versioned file that both the synthetic and the real-data pipelines read:

`mc_production.py` $\rightarrow$ `fcortex_production.json` $\rightarrow$ `fcortex_source.py` $\rightarrow$ both pipelines.

No analysis script carries a hard-coded sensitivity table or a 760 $\rightarrow$ 850 ratio shortcut. The production settings are: anisotropic Henyey–Greenstein scattering $g = 0.9$ (with $\mu_s = \mu'_s/(1 - g)$), a Fresnel surface, partial-pathlength scoring, $N = 2 \times 10^6$ photons per (geometry, wavelength), a converged photon-pathlength cutoff $L_{\max} = 1200$ mm and domain depth $z_{\max} = 150$ mm (both selected from an observed numerical plateau), and seed = 1. Uncertainty on each f_{cortex} is the spread of the *ratio estimator* across 16 independent photon batches (not a seed-level bootstrap); the file also stores per-SDS DPF and N_{eff} (an absorption-weight count only), the wavelength-specific CSF ratios $\gamma(\lambda, \text{SDS})$ at 2 mm and 1 mm CSF thickness, and an $L_{\max}/z_{\max}/\text{annulus}$ convergence sweep. The exact per-layer wavelength-specific optical properties used by the MC are stored in the `fcortex_production.json`

metadata (760 nm: scalp $\mu_a=0.0128$, $\mu'_s=1.08$; cortex $\mu_a=0.0192$, $\mu'_s=0.82$; 850 nm: scalp $\mu_a=0.0165$, $\mu'_s=0.86$; cortex $\mu_a=0.0212$, $\mu'_s=0.76$ mm⁻¹; CSF $\mu_a=0.004$, $\mu'_s=0.25/0.24$ mm⁻¹).

B Appendix B: Glossary of Symbols

Symbol	Name
[HbO ₂]	Oxygenated hemoglobin concentration
[HbR]	Deoxygenated hemoglobin concentration
Δ [HbO ₂]	Change in oxygenated hemoglobin
Δ [HbR]	Change in deoxygenated hemoglobin
λ	Wavelength
$\varepsilon(\lambda)$	Molar extinction coefficient
OD	Optical density (absorbance)
ΔOD	Change in optical density
μ_a	Absorption coefficient
μ'_s	Reduced scattering coefficient
D	Diffusion coefficient
μ_{eff}	Effective attenuation coefficient
L	Source-detector separation (geometric)
$DPF(\lambda)$	Differential pathlength factor
d_{sup}	Superficial layer thickness
$\Phi(\mathbf{r})$	Photon fluence (radiance integrated over angles)
$J(\mathbf{r})$	Jacobian (sensitivity)
f_{cortex}	Cortical sensitivity fraction (converged Monte Carlo)
κ_{DPF}	DPF correction factor
κ_{PV}	Partial volume correction factor
$V_{\text{SSR}}(\lambda)$	SSR variance-removal diagnostic, $1/(1 - R_{\text{SS}}^2)$ (reported, never applied)
κ_{applied}	Applied composite, $\kappa_{\text{DPF}} \times \kappa_{\text{PV}}$ (V_{SSR} is a per-wavelength diagnostic, not applied)
R_{SS}^2	Variance of the uncorrected long-channel ΔOD explained by the short channel (SSR fit R^2)
$\beta(\lambda)$	SSR regression coefficient

C Appendix C: Key References

1. Sagara N, Ahsan A. A diffusion-theory framework for quantifying and correcting partial volume bias in functional near-infrared spectroscopy. *Neurophotonics*. 2026 (submitted).

The research paper on which this pedagogical guide is based. Presents and validates the complete κ -correction framework with synthetic fNIRS data across four source–detector separations and 10 subjects.

2. Boas DA, Dale AM, Franceschini MA. Diffuse optical imaging of brain activation. *NeuroImage*. 2004;23:S275–S288.

Foundational work on partial volume effects in fNIRS and early solutions using image reconstruction.

3. Duncan A, Meek JH, Clemence M, et al. Optical pathlength measurements on adult head, calf and forearm and the head of the newborn infant using phase-resolved optical spectroscopy. *Phys Med Biol.* 1995;40:295–304.

Pioneering experimental measurements of wavelength-dependent DPF in human tissue.

4. Fang Q, Boas DA. Tetrahedral mesh generation from volumetric binary and grayscale images. *IEEE ISBI.* 2009:1142–1145.

The iso2mesh toolbox for converting volumetric images into tetrahedral meshes. This is a mesh-generation tool, *not* a Monte Carlo solver, mesh-based Monte Carlo (MMC) is a separate code by Fang (*Biomed Opt Express*, 2010). We cite the 2009 ISBI paper only when preparing head meshes for forward modeling.

5. Gagnon L, Pertica C, Haley MI, et al. Improved recovery of the hemodynamic response in diffuse optical imaging using short-separation measurements. *NeuroImage.* 2012;59:2788–2802.

Practical SSR techniques for removing superficial contamination.

6. Jacques SL. Optical properties of biological tissues: A review. *Phys Med Biol.* 2013;58:R37–R61.

Comprehensive review of tissue optical properties across the NIR window and other wavelengths.

7. Kienle A, Lilge L, Vitkin IA, et al. Why the optical properties of biological tissue have a minimum in the 600–1000 nm region. *Appl Opt.* 1996;35:2304–2314.

Explanation of the NIR optical window; critical for understanding why these wavelengths penetrate deeply.

8. Kienle A, Glanzmann T, Herrmann U, et al. Spatially resolved absolute diffuse reflectance measurements for noninvasive determination of the optical scattering and absorption coefficients: Instrumentation and numerical results. *J Biomed Opt.* 1998;3:21–31.

Semi-analytical solution for two-layer tissue geometry (Hankel transform method); the basis of the κ correction framework.

9. Okada E, Delpy DT. Near-infrared light propagation in an adult head model. *Appl Opt.* 2003;42:2906–2914.

Optical properties of head tissue and FD/TD measurements; validation of diffusion model.

10. Prahl SA. Tabulated molar extinction coefficient for hemoglobin. *Oregon Medical Laser Center*. 1999.

Empirical extinction spectra for HbO₂ and HbR; standard reference used in fNIRS processing.

11. Saager RB, Berger AJ. Direct characterization and removal of interfering absorption trends in two-layer turbid media. *J Opt Soc Am A*. 2005;22:1874–1884.

Early SSR methodology; foundational for superficial noise removal.

12. Scholkmann F, Wolf M. General equation for the differential pathlength factor of the frontal human head depending on wavelength and age. *J Biomed Opt*. 2013;18:105004.

Empirical formula for wavelength- and age-dependent DPF; widely used in modern fNIRS systems.

13. Strangman GE, Li Z, Zhang Q. Depth sensitivity and source-detector separations for near infrared spectroscopy based on the Colin27 brain template. *PLoS ONE*. 2013;8:e66319.

Atlas-based sensitivity analysis (Colin27) showing how f_{cortex} varies with source-detector separation and scalp-to-cortex depth; informs the choice of SDS in this framework.

14. Hirth C, Obrig H, Villringer K, et al. Non-invasive functional mapping of the human motor cortex using near-infrared spectroscopy. *NeuroReport*. 1996;7:1977–1981.

Early demonstration of motor-cortex HbO₂ amplitudes during finger tapping in the $\sim 1\text{--}3\ \mu\text{M}$ range; reference for task-evoked cortical amplitude expectations.

15. Obrig H, Hirth C, Junge-Hülsing JG, et al. Cerebral oxygenation changes in response to motor stimulation. *J Appl Physiol*. 1996;81:1174–1183.

Companion measurements of motor-cortex HbO₂/HbR responses; supports the $1\text{--}5\ \mu\text{M}$ task range cited in Appendix B.

16. Holper L, Biallas M, Wolf M. Task complexity relates to activation of cortical motor areas during uni- and bimanual performance: A functional NIRS study. *NeuroImage*. 2009;46:1105–1113.

Modern finger-tapping fNIRS data with amplitudes consistent with the Hirth/Obrig range.

17. Arridge SR, Cope M, Delpy DT. The theoretical basis for the determination of optical pathlengths in tissue: Temporal and frequency analysis. *Phys Med Biol*. 1992;37:1531–1560.

Derivation of the diffusion approximation from the radiative transfer equation for NIR light in tissue; foundational reference for the diffusion-based forward model used in this framework.

18. Haskell RC, Svaasand LO, Tsay T, et al. Boundary conditions for the diffusion equation in radiative transfer. *J Opt Soc Am A*. 1994;11:2727–2741.

Derivation of the extrapolated (Robin) boundary condition used at the air-tissue interface; source of the $A \approx 2.95$ coefficient for $n = 1.4$ used in the Kienle two-layer solver.

19. Hielscher AH, Alcouffe RE, Barbour RL. Comparison of finite-difference transport and diffusion calculations for photon migration in homogeneous and heterogeneous tissues. *Phys Med Biol*. 1998;43:1285–1302.

Benchmarks the validity limits of the diffusion approximation against direct solutions of the radiative transfer equation; cited when discussing where the diffusion model is accurate.