



PassFail Criteria Report

Printed on 19-Mar-2020 at 13:51:43 by u1

Study Name: manual 1 min sampling	Company: KAYE	ValProbe RT Version: 1.0.0.51
SOP / Protocol #: 1	Run#: 1	ValProbe RT Reports Version: 1.0.0.36
Asset Name: Asset 10050	Asset ID: 10050	BS Firmware Version: 0.7.0
Started By: a	Date: 18-Mar-2020 13:12:41	Stopped By: a
Setup Created By: a	Setup Created Date: 18-Mar-2020 13:05:45	Study File Comments: 1 min sampling with lethality
Setup Comments: 1	Setup Description: 1	

Comments

Lethality has been selected with the following criteria: Base Temperature : 121.1 °C D Value : 1.00 Z Value : 10.0 TimeScale : Minutes
Lethality Calculated: During Entire Cycle

Performed By: _____

Date: _____

Reviewed By: _____


Date: _____

PassFail Criteria Report

Printed on 19-Mar-2020 at 13:51:43 by u1

Study Name: manual 1 min sampling

Temperature_1		
Process Temperature: 138.00 °C	SamplingRate: 1 minute	Template Name: pf1
Asset ID: 10050	SOP Protocol: 1	Description: 1

Input Criteria	Criteria	Value	Sensors Label	Time	Status
Process Temperature Band Min. (°C)	Process Temp - 130.00	25.60			 PASS
Process Temperature Band Max. (°C)	Process Temp + 140.00	140.13			

Performed By: _____

Date: _____

Reviewed By: _____


Date: _____

PassFail Criteria Report

Printed on 19-Mar-2020 at 13:51:43 by u1

Study Name: manual 1 min sampling

Temperature_2		
Process Temperature: 138.00 °C	SamplingRate: 1 minute	Template Name: pf1
Asset ID: 10050	SOP Protocol: 1	Description: 1

Input Criteria	Criteria	Value	Sensors Label	Time	Status
Process Temperature Band Min. (°C)	Process Temp - 130.00	29.48			 PASS
Process Temperature Band Max. (°C)	Process Temp + 140.00	139.74			

Performed By: _____

Date: _____

Reviewed By: _____

Date: _____

Study Name: manual 1 min sampling

Comments:**System Messages**

18-Mar-2020 13:12:41	***Study Start***	Manual	a
18-Mar-2020 13:12:41	ZA46_Start	Manual	
18-Mar-2020 13:12:41	YZ90_Start	Manual	
18-Mar-2020 13:12:42	ZD72_Start	Manual	
18-Mar-2020 13:12:42	YI07_Start	Manual	
18-Mar-2020 15:56:00	ZA46_End	Manual	
18-Mar-2020 15:56:00	ZD72_End	Manual	
18-Mar-2020 15:56:00	YZ90_End	Manual	
18-Mar-2020 15:56:00	YI07_End	Manual	
18-Mar-2020 15:56:00	***Study End***	Manual	a

Performed By: _____

Date: _____

Reviewed By: _____

Date: _____