

Temperature Logger U88A-A			Verification 23-Jun-2020 by HTC_Admin1		
Company: KAYE		Logger Firmware Version: 1.4.1		Software Version: 1.1.0.1	
SOP / Protocol #: VS020BC-XS		Battery life: 30 %		Manual Verification	
Temperature Std: Kaye IRTD		Serial No: J1129		Cal Date: 14-Oct-2019	

**Summary Verification Results: Failed**

SP1: Passed	SP2: Passed	SP3: Passed	SP4: Passed	SP5: Failed	SP6: Failed
-------------	-------------	-------------	-------------	-------------	-------------

**Start Time: 10:55:39**

Setpoint 1: -25.00°C Passed		
<b>Stability Criteria:</b>	IRTD Stability: 0.012°C for 3 minutes	Additional time of Stability: 5 minutes
<b>Deviation Criteria:</b>	0.50°C for 3 minutes	

**Stability Evaluation:**

Time of Stability: 11:46:10	Sensor Temperature: -23.57°C	IRTD Temperature: -23.859°C	Sensor Deviation From IRTD: 0.29°C
	Sensor Stability: 0.01°C	IRTD Stability: 0.010°C	

Deviation Evaluation:				Criteria: 0.50°C for 3 minutes				Maximum Deviation: 0.30			
Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
11:46:20	-23.860	-23.57	0.29	11:46:30	-23.861	-23.57	0.29	11:46:40	-23.862	-23.57	0.29
11:46:50	-23.862	-23.57	0.29	11:47:00	-23.862	-23.57	0.29	11:47:10	-23.862	-23.57	0.29
11:47:20	-23.862	-23.57	0.29	11:47:30	-23.862	-23.57	0.29	11:47:40	-23.864	-23.57	0.29
11:47:50	-23.865	-23.58	0.28	11:48:00	-23.865	-23.58	0.28	11:48:10	-23.865	-23.57	0.30
11:48:20	-23.867	-23.57	0.30	11:48:30	-23.868	-23.58	0.29	11:48:40	-23.868	-23.58	0.29
11:48:50	-23.868	-23.57	0.30	11:49:00	-23.869	-23.57	0.30	11:49:10	-23.868	-23.57	0.30

Setpoint 2: -10.00°C Passed		
<b>Stability Criteria:</b>	IRTD Stability: 0.012°C for 3 minutes	Additional time of Stability: 5 minutes
<b>Deviation Criteria:</b>	0.50°C for 3 minutes	

**Stability Evaluation:**

Time of Stability: 12:21:50	Sensor Temperature: -9.81°C	IRTD Temperature: -10.048°C	Sensor Deviation From IRTD: 0.24°C
	Sensor Stability: 0.01°C	IRTD Stability: 0.009°C	

Deviation Evaluation:				Criteria: 0.50°C for 3 minutes				Maximum Deviation: 0.24			
Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
12:22:00	-10.047	-9.81	0.24	12:22:10	-10.047	-9.81	0.24	12:22:20	-10.046	-9.81	0.24
12:22:30	-10.045	-9.81	0.23	12:22:40	-10.044	-9.81	0.23	12:22:50	-10.044	-9.80	0.24
12:23:00	-10.044	-9.80	0.24	12:23:10	-10.043	-9.80	0.24	12:23:20	-10.042	-9.80	0.24
12:23:30	-10.043	-9.80	0.24	12:23:40	-10.041	-9.80	0.24	12:23:50	-10.041	-9.80	0.24
12:24:00	-10.040	-9.80	0.24	12:24:10	-10.041	-9.80	0.24	12:24:20	-10.040	-9.80	0.24
12:24:30	-10.040	-9.80	0.24	12:24:40	-10.039	-9.80	0.24	12:24:50	-10.039	-9.80	0.24

Setpoint 3: 30.00°C Passed		
<b>Stability Criteria:</b>	IRTD Stability: 0.012°C for 3 minutes	Additional time of Stability: 5 minutes
<b>Deviation Criteria:</b>	0.50°C for 3 minutes	

**Stability Evaluation:**

Time of Stability: 12:43:20	Sensor Temperature: 29.98°C	IRTD Temperature: 30.008°C	Sensor Deviation From IRTD: -0.03°C
	Sensor Stability: 0.01°C	IRTD Stability: 0.005°C	

Performed by \_\_\_\_\_

Date \_\_\_\_\_

Reviewed by \_\_\_\_\_

Date \_\_\_\_\_

## Deviation Evaluation:

Criteria: 0.50°C for 3 minutes

Maximum Deviation: -0.03

Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
12:43:30	30.007	29.98	-0.03	12:43:40	30.007	29.99	-0.02	12:43:50	30.007	29.99	-0.02
12:44:00	30.006	29.99	-0.02	12:44:10	30.006	29.99	-0.02	12:44:20	30.006	29.99	-0.02
12:44:30	30.004	29.99	-0.01	12:44:40	30.005	29.99	-0.01	12:44:50	30.004	29.98	-0.02
12:45:00	30.004	29.98	-0.02	12:45:10	30.004	29.99	-0.01	12:45:20	30.004	29.99	-0.01
12:45:30	30.004	29.99	-0.01	12:45:40	30.003	29.99	-0.01	12:45:50	30.004	29.98	-0.02
12:46:00	30.004	29.99	-0.01	12:46:10	30.003	29.99	-0.01	12:46:20	30.003	29.99	-0.01

## Setpoint 4: 45.00°C Passed

Stability Criteria: IRTD Stability: 0.012°C for 3 minutes Additional time of Stability: 5 minutes

Deviation Criteria: 0.50°C for 3 minutes

## Stability Evaluation:

Time of Stability: 12:59:30 Sensor Temperature: 44.92°C IRTD Temperature: 45.012°C Sensor Deviation From IRTD: -0.09°C  
 Sensor Stability: 0.01°C IRTD Stability: 0.008°C

## Deviation Evaluation:

Criteria: 0.50°C for 3 minutes

Maximum Deviation: -0.09

Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
12:59:40	45.011	44.92	-0.09	12:59:50	45.010	44.93	-0.08	13:00:00	45.010	44.92	-0.09
13:00:10	45.010	44.93	-0.08	13:00:20	45.009	44.92	-0.09	13:00:30	45.008	44.92	-0.09
13:00:40	45.008	44.93	-0.08	13:00:50	45.007	44.92	-0.09	13:01:00	45.007	44.93	-0.08
13:01:10	45.006	44.92	-0.09	13:01:20	45.006	44.92	-0.09	13:01:30	45.006	44.92	-0.09
13:01:40	45.006	44.92	-0.09	13:01:50	45.005	44.92	-0.09	13:02:00	45.006	44.92	-0.09
13:02:10	45.005	44.93	-0.08	13:02:20	45.005	44.92	-0.09	13:02:30	45.004	44.92	-0.08

## Setpoint 5: 90.00°C Failed Did Not Meet Calibrated Deviation Criteria

Stability Criteria: IRTD Stability: 0.012°C for 3 minutes Additional time of Stability: 5 minutes

Deviation Criteria: 0.50°C for 3 minutes

## Stability Evaluation:

Time of Stability: 13:29:20 Sensor Temperature: 92.01°C IRTD Temperature: 89.997°C Sensor Deviation From IRTD: 2.01°C  
 Sensor Stability: 0.01°C IRTD Stability: 0.007°C

## Deviation Evaluation:

Criteria: 0.50°C for 3 minutes

Maximum Deviation: 2.02

Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
13:29:30	89.997	92.01	2.01	13:29:40	89.997	92.01	2.01	13:29:50	89.996	92.01	2.01
13:30:00	89.996	92.01	2.01	13:30:10	89.995	92.01	2.01	13:30:20	89.994	92.00	2.01
13:30:30	89.994	92.00	2.01	13:30:40	89.993	92.00	2.01	13:30:50	89.993	92.00	2.01
13:31:00	89.993	92.01	2.02	13:31:10	89.993	92.01	2.02	13:31:20	89.992	92.01	2.02
13:31:30	89.991	92.00	2.01	13:31:40	89.992	92.00	2.01	13:31:50	89.991	92.01	2.02
13:32:00	89.991	92.00	2.01	13:32:10	89.990	92.00	2.01	13:32:20	89.990	92.01	2.02

## Setpoint 6: 140.00°C Failed Did Not Meet Calibrated Deviation Criteria

Stability Criteria: IRTD Stability: 0.012°C for 3 minutes Additional time of Stability: 5 minutes

Deviation Criteria: 0.50°C for 3 minutes

## Stability Evaluation:

Time of Stability: 14:07:30 Sensor Temperature: 144.62°C IRTD Temperature: 139.979°C Sensor Deviation From IRTD: 4.64°C  
 Sensor Stability: 0.01°C IRTD Stability: 0.007°C

Performed by \_\_\_\_\_

Date \_\_\_\_\_

Reviewed by \_\_\_\_\_

Date \_\_\_\_\_



# ValProbe RT Temperature Verification Report

Printed on 15-Jul-2020 at 16:40:08 by User1

## Deviation Evaluation:

Criteria: 0.50°C for 3 minutes

Maximum Deviation: 4.66

Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)	Time	IRTD(°C)	Sensor(°C)	Dev(°C)
14:07:40	139.979	144.62	4.64	14:07:50	139.978	144.62	4.64	14:08:00	139.978	144.63	4.65
14:08:10	139.977	144.62	4.64	14:08:20	139.977	144.62	4.64	14:08:30	139.976	144.63	4.65
14:08:40	139.976	144.62	4.64	14:08:50	139.975	144.62	4.64	14:09:00	139.975	144.62	4.64
14:09:10	139.975	144.63	4.65	14:09:20	139.974	144.62	4.65	14:09:30	139.973	144.62	4.65
14:09:40	139.974	144.62	4.65	14:09:50	139.973	144.62	4.65	14:10:00	139.972	144.63	4.66
14:10:10	139.972	144.62	4.65	14:10:20	139.972	144.62	4.65	14:10:30	139.972	144.63	4.66

Performed by \_\_\_\_\_

Date \_\_\_\_\_

Reviewed by \_\_\_\_\_

Date \_\_\_\_\_