



Diabetotech

AID Systems	MiniMed 780G	Tandem Control-IQ	CamAPS FX/MyLife Loop	Diabeloop	Omnipod 5	Beta Bionics iLet	TouchCare Nano System	Twist AID System	DIY Loop	AndroidAPS	iAPS/Trio
Photo											
Labeling	CE + FDA-label	CE + FDA-label	CE-label	CE-label	CE + FDA-label	FDA-label	CE-label	FDA-label	/	/	/
Age limit	≥ 7 years	≥ 6 years and ≥ 25 kg	≥ 1 year and ≥ 10 kg	≥ 18 years	≥ 2 years	≥ 6 years	≥ 6 years and ≥ 22 kg	≥ 6 years	/	/	/
Insulin requirement	8-250 U per day	10-100 U per day	≥ 5-350 U per day	8-90 U per day	≥ 6 U per day	no	≥ 10 U per day	no	/	/	/
CGM	Guardian 3 & 4, Simplera Sync	Dexcom G6 & G7, FreeStyle Libre 2 Plus depending on your region, future: FreeStyle Libre 3 Plus	Dexcom G6 or FreeStyle Libre 3	Dexcom G6	Dexcom G6 & G7, FreeStyle Libre 2 Plus depending on region, future: FreeStyle Libre 3 Plus	Dexcom G6 & G7, FreeStyle Libre 3 Plus	TouchCare Nano CGM	FreeStyle Libre 3 Plus	Dexcom, FreeStyle Libre, Enlite, Nightscout as BG Source	Dexcom, FreeStyle Libre, Enlite, Nightscout as BG Source	Dexcom, FreeStyle Libre, Enlite, Nightscout as BG Source
Pump	MiniMed 780G	Tandem t:slim X2 + Mobi pump	Ypsopump (Dana-pump via CamDiab)	Kaleido, Dana-i (Accu-Chek Insight discontinued, future: Medisafe WITH)	Omnipod	iLet ACE Pump	TouchCare Nano pump	twist Pump	Omnipod Eros/DASH, older Medtronic pumps, future: Dana-i	AccuCheck Combo/Insight, Dana R/RS/i, older Medtronic pumps, Omnipod Eros/DASH, DiaconG8, Medtrum Nano/300U, EoPatch2	Omnipod Eros/DASH, older Medtronic pumps, future: Dana-i
Algorithm	SmartGuard	Control-IQ technology	CamAPS FX	DBLG1	SmartAdjust	iLet Dosing Decision Software	APGO	twist Loop	Loop	OpenAPS	OpenAPS
CALCULATE	Insulin delivery is calculated based on total insulin delivery of the last 2-6 days + autocorrections 1x/5 minutes	Preset basal insulin rate is modulated based on predicted glycemia, total insulin dose and weight + autocorrections 1x/hour	Delivers extended insulin boluses, calculated using predicted glycemia, total daily insulin dose, and weight.	Calculates microboluses based on predicted glycemia, total daily insulin dose, weight & average carbohydrate intake + autocorrections	Delivers microboluses based on predicted glycemia and total insulin dose of the past three days.	Delivers microboluses based on predicted glycemia and weight.	Delivers insulin every 2 minutes based on previous, current and predicted glucose and insulin delivery.	Preset basal insulin rate is modulated based on predicted glycemia, insulin and carbohydrates on board.	Preset basal insulin rate is modulated based on predicted glycemia, insulin and carbohydrates on board.	Preset basal insulin rate is modulated based on predicted glycemia, insulin and carbohydrates on board.	Preset basal insulin rate is modulated based on predicted glycemia, insulin and carbohydrates on board.
ADJUST											
Target Value (mg/dl)	100, 110 or 120 mg/dl	112.5-160 mg/dl	80-200 mg/dl	100-150 mg/dl	110-150 mg/dl	110-130 mg/dl	100, 110 or 120 mg/dl	87-180 mg/dl	87-180 mg/dl	80-200 mg/dl	72-180 mg/dl
Target Value (mmol/l)	5.6, 6.1 or 6.7 mmol/l	6.25-8.8 mmol/l	4.4-11 mmol/l	6.1-8.3 mmol/l	6.1-8.3 mmol/l	6.1-7.2 mmol/l	5.6, 6.1 or 6.7 mmol/l	4.8-10 mmol/l	4.8-10 mmol/l	4.4-11 mmol/l	4-10 mmol/l
Additional Target value	Temp Target (target value 150 mg/dl or 8.3 mmol/l + no autocorrections)	Sleep (target range 112.5-140 mg/dl or 6.25-6.7 mmol/l + no autocorrections) & Exercise Activity (target range 140-160 mg/dl or 7.8-8.9 mmol/l)	Ease-Off (target value + 45 mg/dl or 2.5 mmol/l and insulin sensitivity +50%) & Boost (35% more insulin delivery)	Activity Mode (target values +70 mg/dl or 3.9 mmol/l) & Zen Mode (target values + 10-40 mg/dl or 0.6-2.2 mmol/l)	Activity Feature (target value 150 mg/dl or 8.3 mmol/l + reduced insulin delivery)	No	Exercise Activity (less insulin delivery, no specific glucose target)	Workout Range (can be set to 87-250 mg/dl or 4.8-13.8 mmol/l) and Pre-Meal Range (can be set to 67-130 mg/dl or 3.7-7.2 mmol/l)	Customisable target and % insulin delivery (overrides)	Customisable target and % insulin delivery	Customisable target and % insulin delivery
Basal insulin rate	No	Yes (different profiles available)	No	No	No	No	No	Yes	Yes	Yes (different profiles available)	Yes (different profiles available)
Carb ratio	Yes	Yes	Yes	No ("Aggressiveness at mealtime")	Yes	No	Yes	Yes	Yes	Yes	Yes
Insulin Sensitivity	No	Yes	No	No ("Aggressiveness at hyper/normoglycemia")	Yes (Only effect on boluses)	No	Yes	Yes	Yes	Yes	Yes
Duration of Insulin Action	Yes (Active Insulin Time)	No	No	No	Yes	No	No	No	Yes (Insulin model)	Yes (Insulin model)	Yes (Insulin model)
Hypoglycemia threshold	No	No	No	Yes	No	No	No	Yes	Yes	No	No
Other adjustable parameters	Autocorrections on/off	No	No	Total daily insulin dose	Correct Above	No	Auto Meal Handling on/off	Max basal rate	Delivery limits, Algorithm type	Delivery limits, Algorithm type, Autosens, UAM	Delivery limits, SMB / Autosens / Dynamic options, Autotune
REVERT	System will revert to manual mode if no connection with pump, no sensor data or concerns about sensor integrity										
EDUCATE	General education for AID systems + device specific education										
SENSOR	Fingerpick is needed when values do not align with your expectations. Most sensor can be calibrated.										
SHARE											
Webbased platform	CareLink	Glooko, Tandem Source	Glooko	YourLoops	Glooko	Beta Bionics Cloud	EasyView Web	Tidepool	Nightscout, Tidepool	Nightscout, Tidepool	Nightscout, Tidepool
Follow app	CareLink Connect	Dexcom Follow	Companion app	Dexcom Follow	Dexcom Follow	Bionic Circle app	EasyFollow	insight	Loop Follow/Caregiver	NSClient app	Loop Follow