

Salient Characteristics:

The specifications for the medication cart should meet the following:

- mobile workstation with drawers for the secure storage of medication
- portable with lockable casters
- weigh no more than 160lbs including batteries, power and accessories
- lockable bins for medication storage
- maximum of 4 different tiers or levels of storage
- customizable bin storage configuration options for medication storage that allow:
 - 3 small and 2 medium bins (21)
 - 9 small bins (4)
- minimum bin dimensions according to size:
 - Small: 11" x 5" x 2"
 - Medium: 11" x 7" x 2"
- articulating arm for monitor that allows the vertical movement rotation of the arm
 - articulating arm should allow at minimum three different positioning options for adjustment
 - movement of the monitor is not sufficient, entire arm should be moveable/rotatable
- minimum tilt rotation of keyboard is 30°
- maximum base dimensions (W x D) is 17" x 20"
- minimum cart height adjustment of 17"
- lowest height of cart is at least of 32"
- Minimum Writing Surface (W x D) of 18" x 11"
- Electronic mechanism for controlling the security of the cart
- Ability to differentiate user access to medication through Single Sign On
- Interactive touch display to control or display medication cart functions
 - access to medication
 - height adjustment
 - PC restart
 - Battery
- Multiple ports to allow direct power connection from PC, monitor, and peripheral to the cart.
- Additional mount/bracket options for:
 - barcode scanner
 - storage bin
 - storage for waste bin
 - storage for cups
 - monitor
- Include medical grade parts including power cord

The computer/workstation should:

- be a small or micro form factor able to store inside cart
- have an Windows 11 Operating System
- contain a solid state hard drive with a minimum size of 1TB
- contain a minimum of 16 GB RAM
- Support two-factor authentication sign on that supports PIV card
- be located on VA Office of Information Technology (OIT) list of approved devices

Peripherals should include:

- PIV card reader can either be external or incorporated as an element in the keyboard
- Monitor: 19" (or larger) LCD that is splash proof and cleanable
- Privacy Filter for Monitor
- Keyboard should be waterproof, sealed, medical grade or better
- Mouse, waterproof and washable or better
- Barcode Scanner