

THIS IS A SOURCES SOUGHT ANNOUNCEMENT, A MARKET SURVEY FOR WRITTEN INFORMATION ONLY. THIS IS NOT A SOLICITATION ANNOUNCEMENT FOR PROPOSALS AND NO CONTRACT WILL BE AWARDED FROM THIS ANNOUNCEMENT. NO REIMBURSEMENT WILL BE MADE FOR ANY COSTS ASSOCIATED WITH PROVIDING INFORMATION IN RESPONSE TO THIS ANNOUNCEMENT OR ANY FOLLOW-UP INFORMATION REQUESTS. NO TELEPHONE CALLS WILL BE ACCEPTED REQUESTING AN RFQ PACKAGE OR SOLICITATION. THERE IS NO RFQ PACKAGE OR SOLICITATION. IN ORDER TO PROTECT THE PROCUREMENT INTEGRITY OF ANY FUTURE PROCUREMENT, IF ANY, THAT MAY ARISE FROM THIS ANNOUNCEMENT, INFORMATION REGARDING THE TECHNICAL POINT OF CONTACT WILL NOT BE GIVEN AND NO APPOINTMENTS FOR PRESENTATIONS WILL BE MADE.

The U.S. Food and Drug Administration (FDA), Office of Acquisitions and Grants Services (OAGS) issues this Sources Sought Notice (SSN) on behalf of the Center for Drug Evaluation and Research (CDER), Office of Testing and Research in order to determine if there are existing small business sources capable of providing an Alberta Idealized Nasal Inlet (AINI), or equivalent instrument, with an Alberta Nasal Leak (or equivalent) test kit. The AINI will be installed on Copley Scientific Next Generation Impactor (NGI) and Fast Screening Impactor (FSI) testing devices to be used for nasal deposition testing of spray dried powders for intranasal drug delivery. The AINI instrument shall meet or exceed the requirements listed below.

If your firm is considered a small business source under North American Industry Classification System (NAICS) 334516 – Analytical Laboratory Instrument Manufacturing; with a Small Business Size Standard of 1,000 employees, and you believe that your firm would be able to provide the FDA with the product described below, please submit an email to suzanne.martella@fda.hhs.gov.

Other than small business concerns, *especially* vendors who may be **able to offer domestic products**, are also encouraged to **submit a capability statement** that provides company information, as well as information demonstrating that it can meet all of the minimum specifications included below—**to include the country of manufacture of products offered**.

Minimum Requirements/Specifications:

1. The Alberta Idealized Nasal Inlet (AINI)(or equivalent) shall:
 - a. Seamlessly integrate with the Copley Scientific Next Generation Impactor (NGI) and Fast Screening Impactor (FSI) systems for daily operation.
 - b. Accurately mimic deposition behavior in each region, allowing the collection of drug samples that reflect the corresponding fraction of the dose for analysis.
 - c. Easily be separated into its component parts to enable drug recovery and assay for each individual area (the vestibule, the turbinates, the olfactory region, and the nasopharynx)
 - d. Come complete with mensuration and test certificates.
 - e. Include an Alberta Nasal Leak (or equivalent) Test Kit that shall:

- i. Verify the integrity and proper sealing of the AINI system
- ii. Detect and quantify any air leakage that could affect test accuracy
- iii. Include all necessary components, fixtures, and instructions for performing leak testing
- iv. Be compatible with the AINI system and associated impactor devices.

Delivery, Installation, Training and Additional System Requirements:

1. The system must be a newly manufactured unit, not used and refurbished or previously used for demonstration.
2. Offered systems shall be a turn-key solution i.e., the Contractor shall be responsible for providing all hardware, components, instruments, computers, software, and that otherwise required to meet these specifications and the FDA's stated need.
3. The Contractor shall provide all labor, travel, and tools to install the equipment at the address provided below, to include shipping. The system shall be delivered with all necessary supplies, accessories, and standard kits required for installation, start-up, and qualification tests.
4. Installation shall be conducted by factory certified service personnel, who shall demonstrate that all modules start up properly, are correctly installed and connected.
5. The systems shall include onsite factory acceptance test, validation documentation and execution support. Validation/Qualifications tests shall be reviewed and approved by an FDA representative before/after installation of the equipment and execution of protocols. Validation/Qualification reports shall include the standard operating procedures and the allowed manufacturer limits for the instrument. All documentation shall be provided in electronic format.
6. The system and associated accessories shall include operations and maintenance manuals covering proper operation, routine maintenance, and troubleshooting for the system and controlling software. All manuals and documentation shall be provided in hard copy and/or electronic format.
7. Upon acceptance of the system, the Contractor shall provide on-site operator training/familiarization for 3-5 users. Such familiarization shall include system operations, calibration, optimization, troubleshooting and basic operational maintenance procedures.
8. Systems shall be warranted for a minimum of one (1) year from FDA acceptance of the system(s) to include on-site training. Warranty shall be included with the equipment and not separately priced. Warranty services shall include troubleshooting capabilities based on complete knowledge of the entire system, immediate access to replacement parts, immediate access to system improvements and updates, labor and travel costs. Phone and email technical support shall be included for a minimum of one (1) year.

Delivery Terms: Delivery, installation, verification/qualification tests, and onsite training shall be completed within 90 days of contract award at the below address:

US Food and Drug Administration (FDA)
Center for Drug Evaluation and Research (CDER)
Office of Testing and Research
10903 New Hampshire Ave, Lab WO 64, Rm 1083
Silver Spring MD 20993

POC: *will be appointed and contact information will be given upon contract award.*

Responses to this Sources Sought shall unequivocally demonstrate that the respondent is regularly engaged in the sale of same or substantially similar services. Though the target audience is small business vendors or small businesses capable of supplying a U.S. product/service of a small business vendor or producer, all interested parties may respond. At a minimum, responses shall include the following:

- Business name and bio, Unique Entity ID (SAM) number, business address, business website, business size status (i.e., SB, VOSB, SDVOSB, HUBZone SB, SDB, WOSB, LB), point of contact name, mailing address (if different from business address), phone number and email address. Provide this same information again if responding to provide a service offered by another firm.
- Past Performance (within the last 3 years) information for the sale of same or substantially similar product to include date of sale, description of services provided, contract dollar value, client name, client address, client contact name, client point of contact mailing address (if different from that provided for client), client point of contact phone number, client point of contact email address, and name of the manufacturer (to include Unique Entity ID (SAM) number and size status) if not the respondent.
- Descriptive literature, brochures, marketing material, etc. detailing the nature of what the responding firm is regularly engaged in manufacturing or selling.
- The offeror shall furnish sufficient technical information necessary for the Government to conclusively determine that the offered services meet the technical requirements identified above.
- If applicable, identification of the firm's GSA Schedule contract(s) by Schedule number and SINS that are applicable to this potential requirement are also requested.
- If applicable, identification of Best in Class (BIC) contract information or other Government-wide or HHS-wide contracts that the equipment/service is available on.
- If a large business, provide whether subcontracting opportunities exist for small business concerns.
- Standard commercial warranty and payment terms.
- **Provide place of manufacture or any other applicable information to enable review and analysis pertaining to the Buy American statute and requirements relating to Made in America in the event a nonavailability waiver request through the MIAO Digital Waiver Portal is needed.**
- Though this is not a request for quote, informational pricing for the products would be helpful for market research purposes.

The Government is not responsible for locating or securing any information not identified in the response.

Interested Contractors must respond with capability statements which are to be submitted by email only to the point of contact listed below on or before **3:00 pm** (Central Time in Jefferson AR) on **April 8, 2026** to Suzanne Martella, email suzanne.martella@fda.hhs.gov. **Reference: 75F40126SSN133298.**

Disclaimer and Important Notes

This notice does not obligate the Government to award a contract or otherwise pay for the information provided in response. The Government reserves the right to use information provided by respondents for any purpose deemed necessary and legally appropriate. Any organization responding to this notice should ensure that its response is complete and sufficiently detailed to allow the Government to determine the organization's qualifications to perform the work.

Responses with weblinks will not be opened or viewed.

Responses must demonstrate capability, not merely affirm the respondent's capability (i.e. the response must go beyond the statement that, "XYZ company can provide the instrument, technical support etc.").

Respondents are advised that the Government is under no obligation to acknowledge receipt of the information received or provide feedback to respondents with respect to any information submitted. After a review of the responses received, a pre-solicitation synopsis or combined synopsis/solicitation may be published in SAM.gov – Contracting Opportunities. However, *responses to this notice will not be considered adequate responses to a solicitation.*

Confidentiality and Proprietary Information

No proprietary, classified, confidential, or sensitive information should be included in your response. The Government reserves the right to use any non-proprietary technical information in any resultant solicitation(s).

Additional Notes:

If the stated requirements appear restrictive, please submit comments detailing the concern. Failure to provide information indicating that the listed specifications are too restrictive will be interpreted as confirmation that the specifications are adequate for a competitive environment.