

SALIENT CHARACTERISTICS

PHYSICAL CHARACTERISTICS: Equipment must have the physical characteristics to enable the government to perform the capabilities as shown below:

M5551 – Tomography System, Optical Coherence (OCT)

- a) Must be capable of core technologies: cSLO, SD-OCT, Active Eye Tracking, Noise Reduction, Auto Rescan
- b) Must be capable of minimum Multimodal Imaging: Retina, Glaucoma, Infrared-Reflectance (IR)
- c) Light Sources: IR laser 815 nm (infrared) and OCT-SLD 880 nm (infrared) Standard Objective Lenses: Max. scan field (visual angle) $30^\circ \times 30^\circ$ (ca. 9×9 mm) and ART (image averaging) up to 100 single images
- d) Widefield Imaging Objective Lens: Max. scan field (visual angle) $\varnothing 55^\circ$ (ca. 16,5 mm); High resolution mode (digital) ≈ 10 $\mu\text{m}/\text{pixel}$ lateral and ART (image averaging) up to 100 single images
- e) Ultra-Widefield Imaging Objective Lens: Max. scan field $\varnothing 102^\circ$ (ref. pupil), $\varnothing 135^\circ$ (ref. eye center); High resolution mode (digital) ≈ 20 $\mu\text{m}/\text{pixel}$ lateral and ART (image averaging) up to 100 single images
- f) OCT – Retinal imaging: A-scan rate: 40 kHz / 85 kHz, Axial digital image size: 496 pixel, Axial resolution: 3,9 $\mu\text{m}/\text{pixel}$, Lateral resolution: 5,7 $\mu\text{m}/\text{pixel}$, Scan angle: up to 55° (ca. 16,5 mm), Scan depth into the tissue: 1,9 mm, Scan patterns: Line scan, volume scan, radial scan, circle scan, ONH-RC scan, individual
- g) OCT – Anterior imaging: cSLO image width: 8,3 / 11 / 16,6 mm, OCT scan depth: 8,3 / 11 / 16,6 mm, ART (image averaging): up to 100 single images
- h) OCT Angiography imaging: High speed mode: $30^\circ \times 15^\circ$ with 11 μm lateral resolution, High resolution mode: $10^\circ \times 10^\circ$ with 5,7 μm lateral resolution, Max. number of pixel: 768×384 -pixel, ART (image averaging): up to 7 single images
- i) Must include table
- j) EMR connectivity
- k) Dimensions: 25.20”D x 43.40”W x 42”H
- l) Quantity: 1

EQUIPMENT LINE ITEMS:

- 1) M5551 - Spectralis Tracking OCT System - 0226-HRAOCT2-ADV
- 2) Ultra-Widefield Angiography Lens - 0226-085
- 3) OCT Angiography, OCTA SHIFT (85 & 125 KHz) - 88025A
- 4) Spectralis Anterior Segment Module - 0226-ASM
- 5) Widefield Imaging Module - 0226-WFIMAGE
- 6) GMPE Software - 0226-962B
- 7) Installation & Training