



Radiation Isotope Identification Detector Functionality Checklist

Page 1

Information	
Date	
Destination (if known)	
Verified by	
Model	
Serial Number (e.g. XXXXXX)	

Equipment & Accessories		
Verify shipping case has no damage		☐
Verify shipping case latch has no damage		☐
Verify case and device serial number match		☐
Verify the following components are present in the case.		
	A. Certificate of conformance	☐
	B. Battery info sheet	☐
	C. Quick start guide	☐
	D. Power supply	☐
	E. Battery charger power supply	☐
	F. AA battery caddy	☐
	G. Batteries (2)	☐
	H. Battery charger	☐
	I. RIID	☐
	J. Shoulder strap	☐

Notes:



Radiation Isotope Identification Detector Functionality Checklist

Page 2

Equipment & Accessories (cont.)		
	K. Shuko plug to C13 (2) (NSDD added)	☐
	L. IEC cables (2)	☐
	M. Car charger	☐
	N. Mini-USB to USB data cable	☐
	O. Mini-B to USB adapter	☐

NSDD Maintenance Toolbox (NMT)	
Instrument successfully connects to NMT	☐
• Power on device (verify battery has at least 2 bars) • Wait for device to complete startup sequence • Open NMT on PC • Connect device to PC with mini-USB to USB data cable (N) • Click on Handheld Menu (HATS icon, bottom right) • Select RIID • Select DASHBOARD • Click Search • Click on device with matching serial number	
Serial number and model in NMT match instrument	☐
Click NSDD Standardize Device button (if shown)	☐

Notes:



Radiation Isotope Identification Detector Functionality Checklist

Page 3

NSDD Maintenance Toolbox (NMT) (cont.)	
Verify all status categories are OK (green check) • Settings • Software version • Internal Stabilization Source <i>Return to main menu (Click back arrow lower left corner)</i>	☐ ☐ ☐
Collect new background • Select Utilities • Verify there are NO radioactive sources in the vicinity • Click Collect New Background • Wait for measurement to complete <i>Return to main menu (Click back arrow lower left corner)</i>	☐
<i>Start an Assessment with default options and no sources nearby. Verify results.</i>	
Assessment results are OK (green check) • Select ASSESSMENT • Click Start Assessment • Wait for measurement to complete • Click Show Results • Observe results <i>Return to main menu (Click back arrow lower left corner)</i>	☐
Export NMT report • Click UP arrow in lower right corner	☐
Disconnect device • Select DASHBOARD • Click Disconnect (bottom) • Disconnect device from computer	☐

Notes:



Radiation Isotope Identification Detector Functionality Checklist

Page 4

Functional Check	
RIID charger functional (D&L)	☐
Battery charger functional (E&L)	☐
Car charger functional (M)	☐
Instrument powers on with battery ONLY (G)	☐
Verify all indicators • Hold RIID left button to enable vibrate and audio • Obtain radioactive gamma source • Place RIID next to radioactive source • Audible sound • Visual lights & messages • Vibration	☐ ☐ ☐
Verify Identification results are correct • Obtain radioactive gamma source • Place RIID next to radioactive gamma source • Press check button to start Identification • Wait for measurement to complete • Press check button to Acknowledge • Verify identification result matches gamma source isotope	☐
Verify identification results can be offloaded to USB • Connect mini-USB to USB adapter (O) • Connect USB drive to adapter • Press check button for Actions • Select Save to USB drive • Select Report (PDF) • Wait for process to complete • Verify process is successful • Disconnect USB adapter cable and USB drive • Return to main RIID menu	☐
Both batteries show full charge after charging (G)	☐

Notes:

--



Radiation Isotope Identification Detector Functionality Checklist

Page 5

Outgoing Deployment Final Preparation	
Verify advanced mode password is correct • Press left arrow to open menus • Press down arrow to select settings (gear icon) • Select Advanced mode (3rd option) • Use Down arrow and check key to enter password (31431) • If successful, Advanced settings menu will appear • If unsuccessful, Incorrect password – Access denied will appear • Stay in Advanced settings menu for next steps	☐
Set clock for destination (if known) • Access Advanced settings menu • Scroll down and highlight Select clock (20th option) • Use buttons to set correct date and time • Select Back when complete	☐
Set destination time zone (if known, NOT REQUIRED) • Access Advanced settings menu • Scroll down and highlight Set time zone (21st option) • Use buttons to set correct date and time Select Back when complete	☐
Set destination Language • Access Advanced settings menu • Scroll down and highlight Set time zone (23rd option) • Highlight desired language • Select Back when complete	☐
Delete all events on the instrument • Access Advanced settings menu • Scroll down and highlight Delete all event data (24th option) • Press check button to select • Highlight Yes • Press check button to select • Press check button to Acknowledge	☐

Notes:



Radiation Isotope Identification Detector Functionality Checklist

Page 6

Outgoing Deployment Final Preparation (cont.)	
Return to main RIID screen • Press back arrow several times until main screen appears	☐
Power off device (hold check button until device powers off)	☐
Remove battery	☐
Return device and accessories to case	☐
Verify batteries are ~30% charged	☐
RECOMMEND: Use Inspired Energy Multi-bay Smart Charge and Calibration System	
Return batteries to case	☐

Notes:

--