

PERFORMACNE WORK STATEMENT (PWS)

WARRANTY AND MAINTENANCE OF LAERDAL SIMMAN 3G

at the 60th Aeromedical Evacuation Squadron
Travis Air Force Base, California

2026

DRAFT

1. DESCRIPTION OF SERVICES

1.1. The Contractor shall provide non-personal services consisting of management, labor, tools, supplies, material, transportation, parts, and replacement parts, diagnostic equipment and other equipment necessary to maintain and provide preventive, corrective maintenance service for Laerdal SimMan 3G located at the 60th Aeromedical Evacuation Squadron, Travis Air Force Base, California. Contractor travel to perform preventative maintenance services, unscheduled/corrective or repair services, and any necessary parts and replacement shall be included in the contract and at no additional cost to the government. The Contractor shall submit reports and documentation as identified throughout this Performance Work Statement (PWS).

1.1.1. Period of Performance:

Base Year: 08 August 2026 – 7 August 2027

1st Option Year: 08 August 2027 – 7 August 2028

2nd Option Year: 08 August 2028 – 7 August 2029

3rd Option Year: 08 August 2029 – 7 August 2030

4th Option Year: 08 August 2030 – 7 August 2031

1.2. All work under this contract shall be performed in compliance with Federal Drug Administration (FDA) standards, National Fire Protection Association Standard NFPA 99 (Electrical Equipment in Healthcare Facilities), and Original Equipment Manufacturer (OEM) specifications as required.

1.3. The Contractor shall be authorized/qualified by the OEM in the repair of all equipment listed under this contract. Contractor shall maintain authorization/qualifications throughout entire contract period. Qualification includes but not limited to the certification of all repairpersons and repair facilities by the OEM as being authorized/qualified to perform required repairs. The Government shall maintain the right to seek proof of qualification prior to award and anytime during contract period of performance. Only OEM authorized/OEM trained personnel shall perform services under this contract.

1.4. Equipment covered under the contract. Equipment listed in this contract will be maintained to meet the OEM's specifications. Appendix A: Equipment List provides the location, quantity and other pertinent information of the SimMan 3G.

1.5. Scheduled Maintenance Inspections. The Contractor shall provide one scheduled preventive maintenance inspection per year performed by OEM authorized/OEM trained personnel. The Contractor shall complete scheduled maintenance to the current published OEM specifications. Service shall consist of: review of system operation, inspection, and functional testing checks of all equipment in accordance with the latest OEM established service procedures as well as full refurbishment/replacement of consumables, some bladders, pad sets, refurbishment of IV arm skin/veins, cleaning, and documentation of findings/recommendations. Scheduled maintenance will include Microsoft patch management if patches exist at the time of the scheduled service.

1.5.1. Schedules shall be developed and approved within 90 days after contract award. Contractor shall finalize the date and time of preventive maintenance services with Clinical Management at least two weeks prior to performing services. All services shall be performed

during normal duty hours 8:00 AM to 5:00 PM, Monday through Friday, excluding Federal holidays as listed in paragraph 2.4.1. The Contractor shall check in with Clinical Management staff both upon arrival and upon completion of on-site work in accordance with paragraph 2.5. of this PWS.

1.6. **Unscheduled/Corrective/Repair Services.** The Contractor shall provide one-year extended warranty to include the coverage of materials, manufacturing, or workmanship issues, return to bench service only, packaging materials for return to Laerdal, shipping costs for repairs (Ground freight from/to customer's facility). The Contractor shall complete corrective maintenance to the current published OEM specifications in order to return the equipment/system to fully operational status within current OEM specifications.

1.7. **Loaner Simulator Service.** The Contractor will provide a one-year loaner service program which includes the shipment of a loaner simulator (on request prior to 1500hrs EST), transportation case for shipping to/from Customer's facility/site. Pre-paid Ground Service will be utilized for all shipping to/from the customer's facility. This service will also cover the expenses to inspect / clean loaner.

1.8. **On-Site Equipment Installation.** The Contractor will provide a one-time on-site installation during base-year period (01 October 2025 – 30 September 2026) by a Laerdal-Certified Field Service Engineer (FSE). FSE will travel to the customer's site and unpack and assemble the human patient simulator/manikin at the customer's designated location. During the installation, the FSE will provide an informal block of instruction on the assemblage of the manikin and the FSE is available for question regarding the set-up of the manikin and basic powering on/shutting down procedures.

1.9. **Telephone Support.** The Contractor shall provide access to OEM telephone support in accordance with OEM specifications and standards.

- Tech Support: 24x7
- Clinical Apps Support: Monday – Friday, 8 am – 5 pm

1.10. **Software Support.** The Contractor shall provide software updates, patches, and field corrections in accordance with the OEM's latest established service procedures. The Contractor shall receive advance approval for the installation of all software updates and revisions from the Government and shall be coordinated through Clinical Management. Defective software shall be replaced at no additional cost to the government.

1.11. **Service Reports.** A service report is required for any maintenance service performed at any time. The original service report must be presented to Clinical Management upon completion of the work prior to departure or must be emailed to Clinical Management at 60AES.AEC.CLINICALMANAGEMENT@us.af.mil within five (5) business days. The service report must include the following as a minimum:

- Date and time of service.
- Company and service representative names.
- Contract number.

- A detailed description of any services performed including labor and any parts used.
- OEM PM/Calibration checklists with results may serve as the service report.

1.12. OEM Replacement Parts. The Contractor shall provide all replacement parts, excluding consumables, necessary for scheduled and corrective services. The Contractor shall have ready access to all OEM replacement parts. All parts supplied shall be compatible with the existing system. All parts must be either new OEM parts or parts refurbished by the OEM of the covered medical equipment. In the event that replacements parts are required to be shipped, shipping shall be performed in the fastest reasonable means possible (i.e., next day air) at no additional cost to the Government. Parts used during scheduled and corrective services shall become the property of the government. Replaced parts shall become the property of the Contractor and shall be removed from the facility and disposed of at no additional cost to the government.

1.13. In the event that items/sub-assemblies /accessories are repaired via "Repair by Replacement," a detailed description of replaced items is to be included upon return of repaired components. All charges resulting from a Contractor-determined requirement to transport Government-owned property covered by this contract to and from an alternate repair location shall be the responsibility of the Contractor.

2. GENERAL INFORMATION

2.1. The Contractor shall not accept any instructions issued by any person employed by the US Government, other than: the Contracting Officer (CO), or Clinical Management, all acting within the limits of their authority. **The CO is the only authorized individual who has the authority to modify or make changes to performance under this contract.**

2.2. The Contractor Point of Contact (POC). The Contractor shall provide in writing the name and telephone number of a primary and alternate individual to act as their representative for the scheduling and coordination of service calls, and to be responsible for the coordination of the contract with the Government.

2.3. The Contractor shall be responsible for the repair/replacement of damaged Government owned equipment and property due to the negligence of the Contractor or his representatives. All such replacement or repair shall be at the Contractor's expense and shall be inspected to the satisfaction of the CO or appointed representative.

2.4 Hours of Operation. Contractor Personnel shall be readily available five (5) days a week to work during normal business hours 8:00 AM - 5:00 PM, Monday through Friday. The Contractor is not required to provide preventative, routine or corrective services on the eleven (11) Federal Holidays or weekends.

2.4.1. Federal holidays are as follows:

New Year's Day - 1 January

Martin Luther King Day - 3rd Monday in January

Washington's Birthday - 3rd Monday in February

Memorial Day - last Monday in May
Juneteenth – 19 June
Independence Day - 4 July
Labor Day - 1st Monday in September
Columbus Day - 2nd Monday in October
Veteran's Day - 11 November
Thanksgiving Day - 4th Thursday in November
Christmas Day - 25 December

If a scheduled service falls on an observed Federal Holiday, service will be performed on the next workday, (i.e., if a holiday falls on Monday, service will be provided on Tuesday). Due to mission requirements, work will be scheduled in advance.

2.5. The Contractor is encouraged to request a pre-maintenance inspection prior to the onset of the contract. As all apparent discrepancies must be identified upon submittal of quote, any and all claims must be received within 30 days of contract award. All non-apparent/hidden discrepancies must be identified no later than the first scheduled preventative maintenance. Any claims resulting from hidden defects must be received within 30 days of first scheduled preventative maintenance. Any equipment found to be inoperable during this pre-maintenance inspection would be repaired using a separate purchase order.

2.6. The government certifies that the equipment to be maintained under this contract will be in good operating condition on the effective date of this contract. For the purpose of this contract, the clause, "good operating condition" means the conditions necessary for the equipment to function as intended without corrective maintenance. The Contractor agrees to leave the equipment in good operating condition at the expiration of this contract.

2.7. During the final month of this contract, the Government will make final inspection of the equipment. Any correction of deficiencies noted during this inspection shall be resolved prior to contract end. The Government reserves the right to request the repair (at no additional cost to the Government) of items identified with latent defects after contract termination, when it has been determined that defects either are a result of contractor's performance or should have been discovered during normal performance of work under this contract.

Appendix A

Equipment List

Manufacturer/Model	Description	Product #	Serial #	Location
Lateral SimMan 3G	High Fidelity Adult simulator that displays neurological and physiological symptoms		21224259945	Bldg 241