

Purchase Description

Procurement of Optical Spectrum Analyzer

The U.S. Army TMDE Activity (USATA) has a requirement to purchase a brand name purchase of one (1) Optical Spectrum Analyzer High Performance model or equivalent that will be a suitable replacement for an aging equipment. This item shall be delivered to the Army Primary Standards Laboratory (APSL), Building 5435 Fowler Road, Redstone Arsenal, AL. The Optical Spectrum Analyzer shall include the following specifications or equivalent (if not brand name):

Requirements:

Yokogawa AQ6370E-20 Optical Spectrum Analyzer	Qty:1
Performance Model	
Wavelength range : 600 to 1700 nm	
Wavelength accuracy : ± 0.01 nm	
Wavelength resolution : 0.02 nm	
Close-in Dynamic range : 78 dB typ.	
Level range : +20 to -90 dBm	
Up to 2x faster measurement: SMSR mode	
Sharper spectrum measurement: HCDR mode	
Large touchscreen LCD	

1.0 Delivery Requirements

Shipping Address: U.S. Army TMDE Activity, Attn: Jonathan Larson, Building 5435, Redstone Arsenal, AL 35898