

Surgical LED Lighting System - Salient Characteristics:

Minimum Clinical Performance Requirements:

- **Illumination and Focus:** Provides deep-cavity illumination with minimal shadows, thanks to unique reflector technology. The center sterile handle allows for easy adjustments in intensity, focus, and positioning.
- **Color Temperature:** white light and 4500K for bright, white light-tailored to different procedures.
- **Energy Efficiency:** Equipped with energy-efficient LEDs that have a lifespan of 40,000 hours.
- **Mobility and Stability:** Features 110° vertical travel, 360° rotation, and drift-free stability, making it versatile for various procedures.
- **Control Options:** Includes wall-mounted controls for intensity and color temperature adjustments.

Source Considerations

- Power source: 100-240VAC, 148-240VA, 50/60 Hz, 1 phase
- Voltage requirements: N/A

Biomedical Tech Considerations

Operator Training requirements: NO
Consumable/Spare/Repair parts inventory considerations: NO
Special Calibration Equipment: NO
Quality history/Supportability considerations: YES
Ease of Maintenance/repair considerations: YES
Regional Support and Service considerations: YES

Certifications

- FDA certifications or 510K numbers that the equipment/system may hold: YES
- Patient Movement Item (PMI) designation if applicable: N/A
- JECETC/JAC certification that the equipment/system may hold: YES
- Conformance to National/International standards such as ASTM, ISO, CSA, UL, NFPA, etc: YES
- Conformance to MIL-STDs if they are required