

STATEMENT OF WORK

1. Objective/Requirements

The information listed below is required to complete this purchase.

2. Characteristics, Scope, and Specs

LINE	PART NO.	DESCRIPTION	QTY.
1	60102	Faraday Defense 48"x46"x36" RF Tent Enclosure - Constructed with two layers of high-shielding Faraday CYBER Fabric (Nickel, Copper, Silver) and lab tested and certified to shield effectiveness standards MIL STD 188-125 and IEEE 299-2006 - Blocks and isolates WiFi (2.4 & 5GHz), Bluetooth, cell signals including 5G networks, RFID, NFC, and radio signals from low MHz to 40GHz - Stitched construction to reduce tears (compared to non-stitch enclosures) - Adjustable Velcro tension straps for precise tent tightness and positioning within the frame - Patented magnetic door system with a magnetic door frame for a tight RF seal - Heavy duty vinyl floor, external outer layer of Black Sports Nylon for extra protection of Farady CYBER Fabric - Internal Faraday Tent Dimensions: 48" x 36" x 36" - External Frame Dimensions: 54" L x 39" W x 39" H	2
2	60124	Ventilation system for tabletop enclosure with (2) 120mm x 120mm dual honeycomb vent panels. Includes (1) 120mm 12VDC muffin fan with 90-240VAC/12VDC power supply with US plug (US plug clip can be changed for international operation) Air flow: 204CFM	2
3	60138	Wall mounted high performance IO plate. Includes 250VAC/500VDC/30A facility filter (-100 dB 100Hz to 40GHz) w/ (1) USA 120VAC/15 5 outlet power strips rated for max 15A and a 5-15P 120V/15A line cord	2
4	60148	6 Terminal DC Filter	2
5	60145	SMA Female Bulkhead w/Dust Cap	6
6	60160	Dual RJ-45 Ethernet-Filtered	2
7	60168	Dual HDMI 1.4 Interface Filter	2
8	60159	Dual USB 3.0 HP Filter	2
9	60164	(1) 75 Ohm BNC RF Bulkhead for SDI Camera and (1) 50 Ohm BNC RD Bulkhead	4
10	60154	JRE UFI-1 Universal Interface HP Filter for the thermocouple and digital/analog IO test	4
11	60139	Blank wall-mounted removable plate with leg extenders	4

3. Place of Performance


 NASA Johnson Space Center
 2101 E. NASA Pkwy

Houston, TX 77058

4. Period of Performance

8 weeks ARO