

The proposed offeror must be able to provide **fifteen (15) CADD-Solis Ambulatory Infusion Pumps with lock boxes, extended description follows:**

Salient Requirements:

- Seller must provide 15 each infusion Pumps as per characteristics listed below.
- Seller must provide 15 each Lock boxes which can store infusion pump and medication bag and provide protection against unauthorized access and theft.
- Seller must provide system that has been approved by the Food and Drug Administration (FDA) and must not have any unresolved FDA recalls.
- Seller shall provide Infusion Pumps, Patient Controlled Analgesics capable of allowing patients the ability to administering medication.
- Seller shall provide device with delivery mode of epidural and intravenous.
- Seller shall provide device with pump mechanism of linear peristaltic.
- Seller shall provide device within its life expectancy. Device identified for discontinuation within the device's life expectancy will be deemed non-acceptable.
- Seller shall provide device with indicators, alarms and dose error reduction.
- Seller shall provide device with stream occlusion, air in line, system malfunction, empty reservoir, low battery alarm/ indicator.
- Seller shall provide device with weight within the range of at least 2-3 pounds for ease of use and transport.
- Seller shall provide device with a flow range of at least 0-100 Continuous/0-500 Bolus and be able to be adjusted by at least 0.05ml/hr. and have a reservoir that is a proprietary cassette/ IV bag.
- Seller shall provide a device capable of providing event log.
- Seller shall provide Pump with capabilities key and/or security code.
- Seller shall provide pump must be compatible with central lines.
- Seller shall provide two copies of operator/user manual.
- Seller shall provide reprocessing instructions.
- Seller must be an authorized dealer, distributor, or reseller for the proposed system.
- Seller must provide a warranty for the system in accordance with Original Equipment Manufacturer (OEM) guidance.