

Fuming chamber with UV contamination module
Minimum Technical Requirements

Summary

This procurement is for twenty-two (22) superglue fuming chamber with UV contamination. The vendor must meet the minimum Technical Requirements listed below which must be brand new and be certified to the manufacturer's Original Equipment Manufacturer (OEM) specification. OEM is Foster & Freeman :

Fuming chamber capabilities:

- Up to twenty-two (22) FFLEX fuming chambers (MVC/FFLEX)
- Approximately 28 cubic feet internal volume (approximate Internal dimensions: 50 inches H x 31.6 inches W x 31.4 inches D)
- Digital variable control for temperature, humidity, fuming and purge cycle times
- Customizable saved processes and logs usage
- Automatic purge
- Adjustable humidity controls from 75%-90% minimum
- Dual temperature hot plate with a minimum 120° C temperature (230° C capability desirable for micro-bursting but not required)
- Internal air circulation with fan assistance
- Integrated filtration system for CA fume containment, user serviceable (no external ducting to facility)
- UV-C (254nm output) decontamination module (MVC/FFLEX/UVKIT)
 - Includes storage for tube
 - Integrated hanging hook
 - Self-capping connectors
- Two (2) per unit equals forty-four (44) Carbon filters should be included
- Variable shelf locations and various positions to suspend items
- Cabinet on wheels
- 120 V power
- Powder coat, aluminum or stainless-steel construction
- Glass windows or door aperture
- Floor standing or with elevated stand or cart for ease of loading and unloading
- Twenty-two (2) Stainless steel user configurable racking/shelving with hanging rails (MVC/FFLEX/M/EXTENDED)
 - With four (4) large shelves
 - One (1) corner unit
 - One (1) corner unit with integrated support
 - Includes all poles for shelving
- Must be tested and certified compliant with current occupational Exposure Standards (<.03ppm)
- Shipping and handling included
- Includes one (1) product training upon first delivery

FSL-A Dual Wavelength Forensic Laser and Imaging System
Minimum Technical Requirements

Delivery Locations and Schedule:

- The equipment shall be delivered within approximately thirty (30) days of receipt of the contract.
- The first call-order will include training and installation by manufacturer to the FCGIL location below. All other equipment shall be installed by the laboratory.
- Twenty-two (22) superglue fuming chambers with UV contamination module, shelving and filters will be delivered to any of the following locations to be determined at the time of order:
 - Forensic Crime Gun Intelligence Laboratory (FCGIL) Bureau of Alcohol, Tobacco, Firearms and Explosives, 4504 E 18th ST N, Wichita, KS 67208
 - FSL-A, 2600 Century Parkway NE, Suite 400, Atlanta, GA 30345
 - FSL-W 6000 Ammendale BLVD, Beltsville MD 20705