

Specifications for solid-state NMR chiller:

Vendor shall provide NASA GRC with a chiller assembly for maintaining low sample and probe temperatures for solid-state nuclear magnetic resonance measurements. It shall meet the following requirements:

1. Low-temperature chiller that integrates with current Bruker solid-state NMR probes and Bruker solid-state NMR system using compatible adaptors and feedthroughs.
2. Non-ambient measurement chamber capable of performing solid-state nuclear magnetic resonance measurements at temperatures down to -40 °C.
3. Requires footprint external to the chamber of approximately 60 cm x 90 cm.
4. Up to 3600 liter per hour to maintain proper sample cooling
5. On-site installation into existing equipment