

**Central Arkansas Veterans Health Care System
Ear, Nose, and Throat Navigation System
DRAFT STATEMENT OF WORK (SOW)**

- I. **Scope:** The Central Arkansas Veterans Health Care System (CAVHS) requires the procurement, delivery, installation, integration, validation, and training for a new ENT(Ear, Nose and Throat) Surgical Navigation System to support Functional Endoscopic Sinus Surgery (FESS), rhinology procedures, and anterior skull-base surgical interventions.
- II. Medtronic brand name or equal ENT(Ear, Nose and Throat) Surgical Navigation System.

	Description	Product Number	QTY	Unit Issue
1	Emitter Axiem 3 Table Top	9733752	1	EA
2	SW Kit Stealth S8 ENT	9735736	1	EA
3	Bracket Support Arm CC	9735833	1	EA
4	Clamp Emitter Holder	9733469	1	EA
5	Arm Side Mount	97357702	1	EA
6	Case Axiem 3 Flat Emitter	9734890	1	EA

- III. **System Requirements include but not limited to:** The Ear, Nose, and Throat Navigation System shall meet the following minimum requirements:
1. **Navigation Technology**
 - Electromagnetic (EM) and/or optical tracking capability.
 - Ability to track multiple surgical instruments simultaneously.
 - Accuracy suitable for ENT/skull-base procedures (≤ 2 mm typical).
 2. **Imaging Compatibility**
 - Full DICOM compatibility for CT and MRI datasets.
 - Ability to import, display, and register patient imaging.
 - Support for automatic, manual, and landmark-based registration.
 3. **OR Integration**
 - Integration with existing endoscopic video towers.
 - Ability to display navigation overlays on primary OR monitors.
 - Support for sterile field interaction (e.g., pointer-driven controls, workflow steps).
 4. **Software Features**
 - Preoperative planning tools.
 - Anatomical mapping and sinus cavity segmentation.
 - Accuracy check function with real-time updates.
 - User profiles, case saving, and documentation tools.
 5. **Hardware Requirements**
 - Mobile cart or tower system with medical-grade displays.
 - Tracking emitter/sensor mounts suitable for ENT surgical positioning.
 - FDA-cleared Class II medical device.

IV. **Specifications:** The system shall include, at minimum:

- One (1) Navigation System Main Console
- One (1) Tracking System (EM or optical)
- One (1) Navigation-ready medical-grade display (24–32 inches)
- Instrument tracking coils or reflectors as applicable
- Navigation-enabled ENT surgical instruments or adapters
- Sterile drapes for navigation sensors and instruments
- All required cabling, power cords, and integration accessories
- Prevalidated software version with all ENT modules licensed
- Compliance with FDA 21 CFR 820 Quality System Regulation
- Compliance with VA cybersecurity and interconnection requirements (MOU-ISA, if applicable)

V. **Maintenance:** The price of the product shall include installation, complete on-site product testing and run through. All site modifications and preparations are the Buyer's responsibility and are to be completed to OEM specifications, prior to the date of the installation.

VI. **Training:** Contractor shall provide all necessary training for the staff on the ENT Navigation System. All surgeons and other hospital staff who will be utilizing the system should be trained by an authorized representative. Upon installation of the ENT Navigation System, contractor representative will schedule training for Trainees at customer's site for no charge on a mutually agreeable date and time.

VII. **Delivery & Installation:** Delivery will be completed within **30 days ARO**

The Contractor shall:

1. Coordinate delivery with **Biomedical Engineering and Surgery/OR**.
2. Provide **inside delivery** to the designated OR or storage area.
3. Complete full **installation, configuration, network connectivity**, and PACS integration (if applicable).
4. Perform **functional testing, calibration, and validation** to ensure compliance with system specifications.
5. Provide a **post-installation report** documenting system readiness, testing results, and training completion.
6. Ensure removal and disposal of all packaging materials.

All contractor personnel entering VA property must comply with facility access, badging, cybersecurity, and infection control procedures.

1. Deliveries will be made to the loading dock at,
Address: 2200 Fort Roots Drive

North Little Rock, AR 72114

Building 182

All items will be turned over to the Receiving/Warehouse staff

Delivery between the hours of 8am -4pm EST Monday-Friday

No deliveries on any Federal Holidays

A lift gate will also be required for all deliveries

Deliver equipment 30 days after purchase

2.Installation to be coordinated by POC

Tyler Jacobs, Surgery Business Manager

501-257-6816

Megan Doyle Administrative Assistant

501-257-6807

Barry Williams Inventory Management Specialist

501-257-4531

Betsi Mergenschroer Inventory Management Specialist

501-257-4538