

N4215826QE043 Enclosure (1)

Vendor shall certify that item quoted meets the capabilities, features and specifications listed below and that vendor can assume all contractor responsibilities stated below by initialing at the bottom. In order to be considered for award, this must be completed and submitted with the vendor quote.

TECHNICAL CAPABILITES STATEMENT		
TOC ANALYZER FOR 6050-ZN01		
Shall be provided with the following capabilities, features and specifications:	Yes	No
CLIN 0001- TOC ANALYZER TOC (TOTAL ORGANIC CARBON) ANALYZER TOC (TOTAL ORGANIC CARBON) ANALYZER, 100/110V, 50/60HZ, MEASUREMENT TECHNIQUE: NON-DISPERSIVE INFRARED (NDIR) DETECTOR, MEASUREMENT RANGE: 10PPB TO 30,000PPM, YSI / XYLEM MODEL 1030W YSI / XYLEM P/N 322110 CERTIFICATE OF COMPLIANCE REQUIRED		

TECHNICAL CAPABILITY/CONTRACTOR RESPONSIBILITY CERTIFICATION

Vendor hereby certifies that vendor's item meets or exceeds all specifications and contractor responsibilities stated in this solicitation, in their entirety, with no exceptions. (initials)

DATE: _____