

PERFORMANCE WORK STATEMENT

Annual Laboratory Instrument Maintenance & Calibration Service

1. BACKGROUND: NAVSTA Rota Fuel Department Laboratory utilizes several petroleum analysis instruments in its testing of jet and diesel fuel and lubricating oil to determine whether fuel meets the required specifications under MIL-DTL-5624, MIL-DTL-83133 and MIL-DTL-16884. To ensure results obtained using these instruments are reliable, calibration is required annually. This contract is to provide the following services to ensure dependable and reliable equipment operation:

- * Visual inspection and verification of instruments and spare parts.
- * Laboratory equipment and instruments cleaning.
- * Laboratory equipment and instruments electronic and mechanic parameters verification and calibration.
- * Minor repairs as identified during calibration.
- * Provide certification of Laboratory equipment maintenance.

2. SCOPE:

Contractor is to provide annual maintenance, calibration, service, and validation of eight petroleum laboratory instruments in accordance with the Original Equipment Manufacturer's (OEM) established requirements. A list of equipment is provided under Section 4.0, herein. The contractor shall be responsible for furnishing all labor and parts necessary to perform an inspection, verification and calibration on the equipment to ensure continued operation at their intended efficiency and capacity.

3. REQUIREMENTS:

3.1 Annual preventive maintenance to ensure precision and reliability - to include calibration and certification documentation in accordance with the OEM standards. Preventive maintenance includes temperature verification, software verification, parameters calibration, pressure verification, and lubrication if necessary, per OEM standards

3.2 Minor repairs may be required upon inspection of the equipment. The contractor shall make minor repairs as determined by inspection tests or disassembly to ensure a functional system that will efficiently serve its intended purpose. Sizable repairs requiring part replacement beyond what is part of calibration are not authorized.

3.3 Contractor personnel shall be qualified to perform the maintenance, calibration, service and validation by the OEM only. Upon request, the Contractor shall provide training certificates (or notarized copies) to the NAVSTA Rota Contracting Officer and Fuels Department for verification.

3.4 Contractor shall ensure that all test equipment and/or calibration standards used in the performance of the contract are traceable and certified. Certification must be made available to the NAVSTA Rota Contracting Officer and/or Fuels Department upon request.

4. GOVERNMENT-FURNISHED PROPERTY AND INFORMATION: The Government is to make available the following instruments for service covered under this contract.

Total	Manufacturer	Model	Instrument Description	Serial Number
1	Precision Analyzer Co. (PAC)	OptiFZP	Freeze Point	7030
1	PAC	FZP 5G2s	Freeze Point	2251
1	Seta Analytics	34300-2	Open Cup Flashpoint	1047789
1	PCS	HFRR	Lubricity	D1948
1	Seta Analytics	SA5000-2	FAME	1040643
1	TBD*	SA200-0	FAME	1097806
1	PAC	OPTICPP	Cloud/Pour Point	1193
1	Cannon Instrument Co.	CAV 4.2	Viscometer	1710-A 825

All GFE in the table above is calibrated in accordance with the OEM and are fully operational.

5. BASE ACCESS: Contractor is to coordinate their site visit with NAVSTA Rota Laboratory personnel, who will assist the Contractor with Security access to the work site. Three weeks' notice in advance is required for obtaining the necessary base access documents.

GOVERNMENT TECHNICAL POINT OF CONTACT: [REDACTED]

6. PLACE OF PERFORMANCE: Naval Station Rota, Fuels Department, Building 600, Rota, ES 11530

7. PERIOD OF PERFORMANCE: Contractors shall provide service to each piece of equipment one time annually. Services will be for a one year base period with (4) additional one year option periods.