

STATEMENT OF WORK (SOW)

Wall-Mounted Intraoral X-Ray Units and Sensors Wichita VA Medical Center Dental Clinic

1. Background

The Robert J. Dole Veterans Health Administration in Wichita, KS seeks the purchase of 8 wall-mounted X-Ray units with sensors for activation of their new Dental Clinic.

2. Scope of Work

- Contractor will provide wall-mounted x-ray units and sensors.
- All equipment will be delivered to the Robert J. Dole VAMC.
- Installation and cleanup will be done by contractor upon delivery.
- On-site training will be provided by the contractor after installation is completed. Training will be scheduled through the dental department.
- Equipment will be under warranty for a minimum of one year after installation.

3. Technical Specifications

- X-ray Tube Voltage: 50–70 kV, with selectable exposure values for different diagnostic needs.
- X-ray Tube Current: 2–8 mA, allowing for adjustable exposure settings.
- Focal Spot Size: 0.4 mm, ensuring optimal resolution and clear images.
- Duty Cycle: 13.5 seconds, with a minimum exposure time of 6 seconds.
- Electrical Classification: Class I Type B, ensuring safety standards are met.
- One size fits all sensors (fits both men and women) with round corners and smooth edges.
- Compatible with MiPACS Enterprise Dental Imaging Software. Must be included in the current, approved VA software inventory for the Enterprise Risk Analysis (ERA).

4. Special Work Requirements

Hours: The regular work hours for this facility is Monday through Friday, from 8:00 am to 4:30 pm (CST), excluding Federal Holidays. All maintenance shall be provided during the regular business hours of this facility. If work is required after hours prior approved will be made by department.

5. **Brand Name or Equal requirement**. The requirement for this acquisition is based on the following items:

- 6150-416 - Planmeca ProX Dual Stud 80" w/remote exposure
- 101810382 - Dexis Ti2 Sensors 2D

Products considered "equal" must meet the salient physical, functional, or performance characteristics specified in this statement of requirements. The contractor is responsible for submitting sufficient documentation for the Government to evaluate "equal" items.