

THIS IS A SOURCES SOUGHT ANNOUNCEMENT, A MARKET SURVEY FOR WRITTEN INFORMATION ONLY. THIS IS NOT A SOLICITATION ANNOUNCEMENT FOR PROPOSALS AND NO CONTRACT WILL BE AWARDED FROM THIS ANNOUNCEMENT. NO REIMBURSEMENT WILL BE MADE FOR ANY COSTS ASSOCIATED WITH PROVIDING INFORMATION IN RESPONSE TO THIS ANNOUNCEMENT OR ANY FOLLOW-UP INFORMATION REQUESTS. NO TELEPHONE CALLS WILL BE ACCEPTED REQUESTING AN RFQ PACKAGE OR SOLICITATION. THERE IS NO RFQ PACKAGE OR SOLICITATION. IN ORDER TO PROTECT THE PROCUREMENT INTEGRITY OF ANY FUTURE PROCUREMENT, IF ANY, THAT MAY ARISE FROM THIS ANNOUNCEMENT, INFORMATION REGARDING THE TECHNICAL POINT OF CONTACT WILL NOT BE GIVEN AND NO APPOINTMENTS FOR PRESENTATIONS WILL BE MADE.

The U.S. Food and Drug Administration (FDA), Office of Acquisitions and Grants Services (OAGS) issues this Sources Sought Notice (SSN) on behalf of the Office of the Chief Scientist (OCS), San Juan Medical Products Laboratory (SJNLMP) in order to determine if there are existing small business sources capable of providing a Mettler Toledo HX204 Halogen Moisture Analyzer (or equivalent) instrument, which meets or exceeds the requirements listed below.

If your firm is considered a small business source under North American Industry Classification System (NAICS) 334516 – Analytical Laboratory Instrument Manufacturing; with a Small Business Size Standard of 1,000 employees, and you believe that your firm would be able to provide the FDA with the product described below, please submit an email to suzanne.martella@fda.hhs.gov.

Other than small business concerns, *especially* vendors who may be able to offer domestic products, are also encouraged to **submit a capability statement** that provides company information, as well as information demonstrating that it can meet all of the minimum specifications included below—**to include the country of manufacture of products offered.**

1. The instrument shall comply with the following quality requirements:
 - a. ISO/IEC 17025-2017 <https://www.iso.org/standard/39883.html>,
<https://www.iso.org/standard/66912.html>
 - b. ISO 9001:2008 <https://www.iso.org/standard/46486.html>
 - c. 21 CFR Part 11 Compliance Features Available
<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/cfrsearch.cfm>
 - d. USP, cGMP/cGLP
<https://login.usp.org/cas/login?service=http%3A%2F%2Fonline.uspnf.com%2Fcas%2Flogin>

Minimum Requirements/Specifications:

The moisture analyzer system shall meet the following salient characteristics:

1. Shall be capable of High-Performance Moisture Analysis
2. Readability: at least 0.001 %MC (0.1mg)
3. Max capacity: at least 200 g
4. Programmable User Roles (Administrator/User)
5. Real-time drying curve
6. Storage for 300 methods
7. Storage for 3000 results
8. Control charts
9. Four drying programs
10. Motorized lid
11. Automated export and print functionality
12. The instrument shall include the current Instrument Controller Software:
 - a. Instrument License
 - b. Easy monitoring for instrument and sample status
 - c. Data collection in SQL database, graphical result evaluation, central instrument management with routine testing scheduler
 - d. Reporting and data storage
 - e. Repository for backups
 - f. Current Instrument Firmware

Delivery, Installation, Training and Additional System Requirements:

1. The system must be a newly manufactured unit, not used and refurbished or previously used for demonstration.
2. Offered systems shall be a turn-key solution i.e., the Contractor shall be responsible for providing all hardware, components, instruments, computers, software, and that otherwise required to meet these specifications and the FDA's stated need.
3. The Contractor shall provide all labor, travel, and tools to install the equipment at the address provided below, to include shipping. The system shall be delivered with all necessary supplies, accessories, and standard kits required for installation, start-up, and qualification tests.
4. Installation shall be conducted by factory certified service personnel, who shall demonstrate that all modules start up properly, are correctly installed and connected.
5. The systems shall include onsite Installation Qualification (IQ), Operational Qualification (OQ) and Performance Verification for the software and hardware. Qualifications tests shall be reviewed and approved by an FDA representative before/after installation of the equipment and execution of protocols. Qualification reports shall include the standard operating procedures and the allowed manufacturer limits for the instrument. Also, qualification tests shall meet Chapter 21 CFR part 11 requirements. All documentation shall be provided in electronic format.
6. The system and associated accessories shall include operations and maintenance manuals covering

proper operation, routine maintenance, and troubleshooting for the system and controlling software. All manuals and documentation shall be provided in hard copy and/or electronic format.

7. Upon acceptance of the system, the Contractor shall provide on-site operator training/familiarization for 3-5 users. Such familiarization shall include system operations, calibration, optimization, troubleshooting and basic operational maintenance procedures.
8. Systems shall be warranted for a minimum of one (1) years from FDA acceptance of the system(s) to include on-site training. Warranty shall be included with the equipment and not separately priced. Warranty services shall include troubleshooting capabilities based on complete knowledge of the entire system, immediate access to replacement parts, immediate access to system improvements and updates, labor and travel costs. Phone and email technical support shall be included for a minimum of one (1) year.

Service Records and Reports: The Contractor shall, commensurate with the completion of each service call (inclusive of warranty service call), provide the end-user of the equipment with a copy of a field service report/ticket identifying the equipment name, manufacturer, model number, and serial number of the equipment being serviced, a detailed description of the work performed, the test instruments or other equipment used to perform the service, the name(s) and contact information of the technician who performed the service, and for information purposes, the on-site hours expended, and parts/components replaced.

Delivery Terms: Delivery, installation, qualification tests, and onsite training shall be completed within 90 days of contract award at the below address:

Food and Drug Administration
San Juan Medical Products Laboratory
466 Fernandez Juncos Ave,
San Juan, PR 00901-3223

POC: will be appointed and contact information will be given upon contract award.

Responses to this Sources Sought shall unequivocally demonstrate that the respondent is regularly engaged in the sale of same or substantially similar services. Though the target audience is small business vendors or small businesses capable of supplying a U.S. product/service of a small business vendor or producer, all interested parties may respond. At a minimum, responses shall include the following:

- Business name and bio, Unique Entity ID (SAM) number, business address, business website, business size status (i.e., SB, VOSB, SDVOSB, HUBZone SB, SDB, WOSB, LB), point of contact name, mailing address (if different from business address), phone number and email address. Provide this same information again if responding to provide a service offered by another firm.
- Past Performance (within the last 3 years) information for the sale of same or substantially

similar product to include date of sale, description of services provided, contract dollar value, client name, client address, client contact name, client point of contact mailing address (if different from that provided for client), client point of contact phone number, client point of contact email address, and name of the manufacturer (to include Unique Entity ID (SAM) number and size status) if not the respondent.

- Descriptive literature, brochures, marketing material, etc. detailing the nature of what the responding firm is regularly engaged in manufacturing or selling.
- The offeror shall furnish sufficient technical information necessary for the Government to conclusively determine that the offered services meet the technical requirements identified above.
- If applicable, identification of the firm's GSA Schedule contract(s) by Schedule number and SINs that are applicable to this potential requirement are also requested.
- If applicable, identification of Best in Class (BIC) contract information or other Government-wide or HHS-wide contracts that the equipment/service is available on.
- If a large business, provide whether subcontracting opportunities exist for small business concerns.
- Standard commercial warranty and payment terms.
- **Provide place of manufacture or any other applicable information to enable review and analysis pertaining to the Buy American statute and requirements relating to Made in America in the event a nonavailability waiver request through the MIAO Digital Waiver Portal is needed.**
- Though this is not a request for quote, informational pricing for the products would be helpful for market research purposes.

The Government is not responsible for locating or securing any information not identified in the response.

Interested Contractors must respond with capability statements which are to be submitted by email only to the point of contact listed below on or before **2:00 pm** (Central Time in Jefferson AR) on **March 13, 2026** to Suzanne Martella, email suzanne.martella@fda.hhs.gov. **Reference: 75F40126SSN132588.**

Disclaimer and Important Notes

This notice does not obligate the Government to award a contract or otherwise pay for the information provided in response. The Government reserves the right to use information provided by respondents for any purpose deemed necessary and legally appropriate. Any organization responding to this notice should ensure that its response is complete and sufficiently detailed to allow the Government to determine the organization's qualifications to perform the work.

Responses with weblinks will not be opened or viewed.

Responses must demonstrate capability, not merely affirm the respondent's capability (i.e. the response must go beyond the statement that, "XYZ company can provide the instrument, technical support etc.").

Respondents are advised that the Government is under no obligation to acknowledge receipt of the information received or provide feedback to respondents with respect to any information submitted. After a review of the responses received, a pre-solicitation synopsis or combined synopsis/solicitation may be published in SAM.gov – Contracting Opportunities. However, *responses to this notice will not be considered adequate responses to a solicitation.*

Confidentiality and Proprietary Information

No proprietary, classified, confidential, or sensitive information should be included in your response. The Government reserves the right to use any non-proprietary technical information in any resultant solicitation(s).

Additional Notes:

If the stated requirements appear restrictive, please submit comments detailing the concern. Failure to provide information indicating that the listed specifications are too restrictive will be interpreted as confirmation that the specifications are adequate for a competitive environment.