

PURCHASE DESCRIPTION
FOR
TORQUE CALIBRATION SYSTEM UPGRADE 250-2000 lbf-ft

1. SCOPE:

The Torque Calibration System is a transducer-based system used by the Air Force Precision Measurement Equipment Laboratories (PMELs) as an Air Force Measurement Standard for calibrating torque wrenches, torque screwdrivers, and torque power tools. The Torque Calibration System Upgrade Package (TCSP) shall consist of torque transducers from 250 to 2000 lbf-ft and a digital indicator to display the torque reading. The uncertainty of the displayed torque shall be within $\pm 0.5\%$ of reading over the range of 250-2000 lbf-ft.

2. PERFORMANCE REQUIREMENTS:

2.1. The contractor shall deliver one each of the following items to the specified PMELs in the contract:

2.1.1. Snap-on TTC610

2.1.2. Snap-on TTC12

2.1.3. Snap-on TTC13

2.1.4. Snap-on TTC14

2.2. Compatibility:

2.2.1. The items in the TCSP shall be compatible with the CDI TFC2000-0-03 Torque Calibration System.

2.2.2. The transducers in the TCSP shall have a four hole bolt pattern.

2.3. Accuracy: The Torque Calibration System shall have an accuracy of $\pm 0.5\%$ of Reading, or better.

3. MECHANICAL & PHYSICAL REQUIREMENTS:

3.1. Ancillary Equipment: Each transducer in the TCSP shall come with a cable for connecting the transducer to the indicator.

3.2. Model and Serial Number Identification: The TCSP shall be identifiable by manufacturer, model or part number, and serial number for identification. Model/Part Number for entire kit shall be an Air Force unique number (i.e. XXXXX-AF) to distinguish between the Air Force and Commercial model. Unique item identification shall follow the item Identification and Valuation clause as stated in the contract.

3.3. Workmanship: The workmanship shall be first class, of neat general appearance in compliance with standard commercial practices. The standards of workmanship shall be such that the system complies with the requirements set forth in this purchase description.

4. WARRANTY: Each transducer and indicator provided shall include, at a minimum, a standard 2 year warranty.

5. TESTING: This item will be inspected and accepted at destination as stated in the contract. The item will be checked with test equipment with parameters equal to or better than those required to ensure that specified performance parameters are met. Test equipment used shall have first echelon certification directly traceable to the National Institute of Standards Technology (NIST) or other AFMETCAL-approved equivalent international laboratory. Tests will be performed to ensure that the

item meets the requirements of this purchase description. Such testing at destination does not relieve the contractor of performing all inspections and quality checks at the point of manufacture as necessary to assure performance as specified herein.