

Statement of Work

Scope

The contractor shall supply FlowMaxx Critical Flow Venturis (CFVs) and provide NIST-traceable air calibrations.

Critical Flow Venturis

The contractor shall deliver FlowMaxx CFVs:

- SNP02-SGR14-100
- SNP02-SGR14-160
- SNP02-SGR14-265
- SNP02-SGR14-430
- SNP04-SGR31-710
- SNP04-SGR31-1160
- SNP04-SGR31-1500

All CFVs must match geometry, throat contour, materials, and interface specifications of the above models

Calibration Requirements

The vender shall provide NIST traceable calibrations for the Reynolds number ranges stated in quotation Q26-115-A.

Mechanical Interfaces

The vendor shall provide adapter flanges and Grayloc hardware (2" and 4") matching SSC's existing mechanical interfaces with no substitutions.

Deliverables

- Serialized CFVs

- Calibration certificates and traceability documents
- Adapter flanges and Grayloc hardware
- Venturi Drawings with MAWP listed

Acceptance Criteria

SSC will accept deliverables when CFVs match specifications listed in this document, calibration data is complete and traceable, documentation is accurate, and components show no damage or deviations.