

Minimum Specifications

Title: Level Sensors

1.0 Scope

NSWCCD Code 633 performs research and development in the areas of non-oily wastewater treatment. Code 633 also provides engineering and technical support to NAVSEA technical warrant holders, various program offices, and other Department of Defense (DoD) organizations. The SCC Facilities funding available for this procurement will support various tasks within the NAVSEA 05P5/F Environmental Quality/Shipboard Waste program. The Endress + Hauser brand of level and pressure sensors in this procurement will allow Code 633 to replace faltering sensors on 6 tanks of the laboratory Liquid Waste Feed System (LWFS) with new and updated models that will ensure the current functionality is maintained within the LWFS.

2.0 Minimum Specifications

2.1 The contractor shall provide six Endress + Hauser Cerabar S PMC71 level sensors with the following specifications:

- Approval: FM IS Cl. I, II, III Div.1 Gr.A-G, NI Cl. I Div.2 Gr.A-D, AEx ia, zone 0,1,2,20,21,22
- Output: 4-20 mA HART
- Housing: T14 Alu IP66/67 NEMA6P
- Cover Sealing: EPDM
- Cable Entry: NPT1/2 thread, T14 = side cover
- Sensor Range: 400 mbar/40 kPa/6 psi gauge
- Sensor Overload Limit: 8 bar/800 kPa/116 psi
- Calibration Unit: Sensor range inH₂O/ftH₂O
- Process Connection: Thread ANSI MNPT1-1/2, 316L, flush mount
- Seal: FKM

2.2 The contractor shall provide six Endress + Hauser Cerabar M PMC51 pressure transmitters with the following specifications:

- Approval: FM NI Cl.I Div.2 Gr.A-D, zone 2
- Output: 4-20 mA HART
- Display and operation: LCD, push button on display/electronics
- Housing: F31 Alu, glass window
- Electrical Connection: Thread NPT1/2, IP66/68 NEMA 4X/6P
- Sensor Range: 1 bar/100 kPa/15 psi gauge, 10 mH₂O/33 ftH₂O/400 inH₂O overload: 10bar/1 MPA/150 psi
- Reference Accuracy: Standard
- Calibration unit: Sensor range in psi
- Process Connection: Thread ANSI MNPT1/2 FNPT1/4, 316L
- Seal: FKM

3.0 Performance and Delivery

The POP for the contract is one year from award of contract. Delivery of the level sensor devices shall be within 15 working days of contract award and receipt of order to the FOB Destination, below. Warranty shall be provided for 12 months after delivery.

Naval Surface Warfare Center, Carderock Division
ATTN: ***** (Bldg 60, Rm 175)
9500 MacArthur Blvd.
West Bethesda, MD 20817

4.0 Government Technical Point of Contact (TPOC)

To be provided upon Award