

Performance Work Statement
(Includes associated Performance Requirements Summary (PRS))

PERFORMANCE WORK STATEMENT (PWS)

Canon USA CT Scheduled Services and Maintenance Repair Contract.
JAN 2026

Part 1
General Information

1. General: This is a non-personal service(s) contract under which the personnel rendering the service(s) are not subject, either by the contract's terms or by the manner of its administration, to the supervision and control usually prevailing in relationships between the Government and its employees.

1.1 Description of Services/Introduction: This is a non-personal service(s) contract to support preventive maintenance and repair of two Computer Tomography (CT) imaging equipment used in LRMC Radiology department.

1.2 Background: This non-personal services contract is in support of the LRMC Radiology Department. The device under contract are used to for diagnostic imaging used to diagnose and determine patient care.

1.3 Scope: The Contractor shall provide non-personal services. The contractors shall provide all personnel, equipment, tools, materials, parts, supervision, transportation and other items and services necessary to properly maintain all equipment identified in the Equipment List in TECHNICAL EXHIBIT C on the PWS. Services to be performed include all unscheduled routine service calls, all emergency service visits, and all scheduled service visits as required for all Line Items. All parts provided by the contractor shall be covered by the terms and conditions of this contract. All work shall be performed in a professional manner by an authorized service representative. If any deficiencies are found due to negligence of the service representative, the contractor will be required to correct the deficiency to a fully operational status in accordance with manufacturer specifications at no additional cost to the U.S. Government.

1.4 Objectives: The objective of this contract is to ensure that all equipment on the equipment list in TECHNICAL EXHIBIT C are functional and calibration is accurate and repairs are performed in a timely manner.

1.5 General Information:

1.5.1 Quality Control Plan (QCP): The Contractor shall develop and maintain a QCP to ensure services are performed in accordance with (IAW) this PWS. The Contractor shall develop and implement procedures to identify, prevent, and ensure non-recurrence of defective services. The Contractor's QCP is the means by which it assures that the work provided complies with the requirements of the contract.

The Contractor's Proposed QCP shall be submitted to the Contracting Officer (KO) through the Contracting Officer's Representative (COR) for review 10 days after date of contract award (Deliverable A.001).

The Government will review and either notify the Contractor in writing of acceptance of the plan or return their comments to the Contractor within 10 days. If the Government has provided comments, the Contractor shall then have 5 days to submit a Final QCP. After receipt of the Final QCP, the Contractor may receive the Contracting Officer's acceptance in writing. Any proposed changes to the accepted QCP are required to be resubmitted for acceptance by the Contracting Officer no later than 30 days prior to the anticipated change and before implementation by the Contractor. The timeline noted above will apply for review and acceptance for proposed changes. At a minimum, the QCP must include and answer the following to be acceptable:

(a) RESERVED

(b) How the Contractor will monitor work to ensure performance complies with all deliverables (etc. timelines, deadlines, and goals);

(c) How the Contractor will monitor work to ensure performance complies with all specifications and requirements of the contract, including the contract's clauses;

(d) How the Contractor will monitor and ensure staff qualifications remain current and valid including Department of Defense (DoD) Contractor Personnel Office (DOCPER) processes/approvals throughout contract performance;

(e) RESERVED

(f) RESERVED

(g) How the Contractor will identify, investigate, and correct any non-conforming performance and prevent similar deficiencies in the future; and

(h) How the Contractor will file and save all Quality Control related documents for the life of the contract plus 5 years.

1.5.1.1 Reserved

1.5.1.2 The Contractor shall inform the COR of issues or potential issues that might affect the performance within 24 hours of identifying those issues. Verbal reports shall be followed up by written reports within 3 days. (A.002).

1.5.2 Quality Assurance: The Government will evaluate the Contractor's performance under this contract in accordance with the Quality Assurance Surveillance Plan (QASP). This plan is primarily focused on what the Government must do to ensure that the Contractor has performed in accordance with the performance standards. It defines how the performance standards will be applied, the frequency of surveillance, and the acceptable quality levels (performance thresholds).

1.5.3 Recognized Holidays: The following provides information on recognized holidays for the purpose of the PWS. If submittal of any documentation (e.g. deliverables, submittals, etc.) deadlines fall on a holiday, the closest work day prior to the holiday will apply as the deadline for submittal.

1.5.3.1 U.S. Holidays: Work shall not be performed on U.S. federally recognized holidays occurring during the normal workweek unless otherwise directed by the Contracting Officer. When a U.S. holiday occurs on a Saturday or a Sunday, the holiday is observed on the preceding Friday or following Monday, respectively.

New Year's Day - January 1st
Presidents Day - 3rd Monday in February
Juneteenth - June 19th
Labor Day - 1st Monday in September
Veteran's Day - November 11th
Christmas Day - December 25th

M. L. King Memorial Day - 3rd Monday in Jan
Memorial Day - last Monday in May
Independence Day - July 4th
Columbus Day - 2nd Monday in October
Thanksgiving Day - 4th Thursday in November

1.5.3.2 Host Nation Holidays: Work shall not be performed on local Host Nation holidays occurring during the normal workweek unless otherwise directed by the Contracting Officer.

1.5.4 Operating Hours: The Government facility office hours, facility operating hours, and the Contractor support hour requirements often coincide, however, they may differ. Please refer to the following.

1.5.4.1 Government Facility Office Hours: The Government facility office hours are 0800-1700 except U.S. Holidays identified in paragraph 1.5.3.1 above or when the Government facility is closed due to local or national emergencies, administrative closings, or similar Government directed facility closings.

1.5.4.2 Government Facility Operating Hours: The Government facility operating hours occur between the hours of 0001 Sunday through 2359 Saturday. The Government facility is open during local or national emergencies, administrative closings, or similar Government directed facility closings. For example, "The Government facility is open for operations during the recognized U.S. Holidays identified in paragraph 1.5.3.1 above. The Government facility may close due to local or national emergencies, administrative closings, or similar Government directed facility closings and will be communicated by the Contracting Officer to the Contractor."

1.5.4.3 Contractor Support Hours: The Contractor shall provide support between the hours of 0800-1700 except U.S. Holidays identified in paragraph 1.5.3.1 above or when the Government facility is closed due to local or national emergencies, administrative closings, or similar Government directed facility closings. The Contractor shall maintain an adequate workforce at all times, for the uninterrupted performance of all tasks defined within this PWS when the Government facility is not closed for the above reasons. When hiring personnel, the Contractor shall keep in mind that the stability and continuity of the workforce is essential.

1.5.4.3.1 Contractor Work Plan: Reserved

1.5.5 Training Holidays: USAREUR has historically granted military personnel training holidays as compensation for time that they spent in field training. These are typically combined with regular holiday leave schedules (for example, Christmas half days or day on day off schedules). Training holidays may affect access to military operated facilities. If scheduled work or access to facilities is required, coordination with occupants will be necessary or the work in affected facilities shall be rescheduled after the training holiday. However, announced training holidays or exodus of military personnel may provide the Contractor with an opportunity for unrestricted access to facilities and areas. The Contractor shall take full advantage of such opportunities by planning maintenance or actions that are better accomplished when such areas are vacant.

1.5.6 Place of Performance: The work to be performed under this contract will be performed at Landstuhl Regional Medical Center and for locations listed in PWS TECHNICAL EXHIBIT C – Equipment list.

1.6 Security Requirements: The following information is provided on security related matters.

1.6.1 **Security Program:** Reserved

1.6.2 **Personnel Security Clearance Requirements:** Reserved

1.6.3 Installation Access: Access to U.S. installations, buildings and controlled areas is limited to personnel who meet security criteria and are authorized by Host Nation law to work in that country. Failure to submit required information/data and obtain required documentation or clearances will be grounds for denying access to U.S. installations, buildings and controlled areas. The Contractor shall ensure that any subcontractors used in performance of this contract complies with these requirements and that all employees, of both the Contractor and any subcontractor utilized by the Contractor, are made aware of and comply with these requirements.

The Contractor shall be aware of and comply with the requirements associated with Installation Access Control. The Government is not liable for any costs associated with performance delays due solely to a firm's failure to comply with Installation Access Control System (IACS) processing requirements.

The Contractor shall return installation passes to the issuing IACS office when the contract is completed or when a Contractor employee no longer requires access.

1.6.3.1 **Installation Pass Log:** Reserved

1.6.3.2 **Individual Termination or Expiration of Employment:** Reserved

1.6.3.3 Contract Termination or Cancellation: Upon termination or cancellation of this contract, the Contractor shall collect all outstanding installation access passes and return them to the issuing office within 5 business days (A.003).

1.6.3.4 **Common Access Cards (CAC) and CAC Log:** Reserved

1.6.4 **Background Checks:** Reserved

1.6.4.1 U.S. Citizen: As needed for base access. In order to receive an installation access pass or work on a US government installation, a background check may be necessary dependent upon base access requirements. The Requiring Activity Manager will determine the extent of background checks required. The Contractor employee may begin performance of the contract after receiving a favorable local background check. The Contractor agrees to replace the employee should results of the background check, if required, be unfavorable as determined by the Requiring Activity Manager. The Contractor's replacement of the employee with the unfavorable background check shall be within a timeframe provided by the Requiring Activity Manager. The Contractor shall also replace employees whose background check, during their tenure of employment, renders them with an unfavorable local or stateside background check, as determined by the Requiring Activity Manager. The Contractor shall forward a copy to the Contracting Officer via the COR and the Requiring Activity Manager of each favorable background check within 10 days of receipt by the Contractor, identifying such with the appropriate contract number (A.006). The Contractor shall forward a copy to the Contracting Officer via the COR and the Requiring Activity Manager of each unfavorable background check within 3 of receipt by the Contractor, identifying such with the appropriate contract number (A.006). When the Government awards a non-personal services contract directly to an individual, the failure of the individual to provide a favorable background check would be grounds for termination.

1.6.4.2 Local Nationals and Third Country Nationals: As needed for base access. In order to receive an installation access pass or work on a US government installation, a background check may be necessary dependent upon base access requirements. As a minimum prior to beginning performance, Contractors employing local national and third country nationals shall furnish the Contracting Officer via the COR and the Requiring Activity Manager current documents required for working on a US installation, no later than 10 days after award (Deliverable A.006) Forwarding correspondence from the Contractors shall identify the associated contract number. The Requiring Activity Manager will determine the favorability of the Polizeiliche Fuehrungszeugnis and notify the Contractor. The Contractor shall forward a copy of unfavorable Polizeiliche Fuehrungszeugnis to the Contracting Officer via the COR and the Requiring Activity Manager within 3 days of receipt, identifying such with the appropriate contract number. Local nationals and third country nationals agree to the same employee termination provisions as stated for U.S. Citizens (see paragraph 1.6.4.1 above, in particular, same requirements to replace the employee (and within timeframe noted) when results of background check are unfavorable.

1.6.5 Physical Security: The Contractor shall safeguard all Government equipment, information, and property provided for Contractor use. At the close of each work period,

Government facilities, equipment, and materials shall be secured IAW the Army Physical Security Program (AR 190-13.”).

1.6.6 Operations Security (OPSEC) Requirements: Contractor personnel shall adhere to facility security policies and restrictions. The Contractor shall immediately report suspicious activities to security personnel. Government issued access badges shall not be worn outside designated facilities where visible to the general public.

1.7 Key Control: Reserved

1.8 Post Award Conference/Periodic Progress Meetings: The Contractor shall attend any post award conference convened by the contracting activity or contract administration office in accordance with FAR Subpart 42.5. The Contracting Officer, COR, and other Government personnel, as appropriate, may meet periodically with the Contractor to review the Contractor's performance. At these meetings the Contracting Officer will apprise the Contractor of how the Government views the Contractor's performance and the Contractor shall apprise the Government of problems, if any, being experienced. Appropriate action shall be taken to resolve outstanding issues. These meetings shall be at no additional cost to the Government.

1.8.1 The Contractor shall attend, participate in, and furnish input to scheduled and unscheduled meetings, conferences, and briefings that relate to the functions and services herein as required by the Government to provide effective communication and impart necessary information. The Contract Manager or designated representative shall attend meetings as requested by the Government. Meeting attendees shall at times include Contractor managerial, supervisory, and other personnel knowledgeable of the subject matter. Meetings may start or end outside of regular duty hours.

1.9 Contracting Officer's Representative (COR): Refer to Part 2 of this PWS for the definition of a COR. As determined by the Contracting Officer, a COR will be appointed and identified by letter of designation, a copy of which will be provided to the Contractor by the Contracting Officer. The designation letter states the responsibilities and limitations of the COR, especially with regard to changes in cost or price, estimates, or changes in delivery dates. The COR is not authorized to change any of the terms and conditions of the resulting order. The COR monitors all technical aspects of the contract and assists in contract administration. The COR is not authorized to obligate the Government. If the work is not written in the contract, the COR is not authorized to request new work. The COR is authorized to perform the following functions: assure that the Contractor performs the technical requirements of the contract, perform inspections necessary in connection with contract performance, maintain written and oral communications with the Contractor concerning technical aspects of the contract, issue written interpretations of technical requirements,

including Government drawings, designs, specifications, monitor Contractor's performance and notifies both the Contracting Officer and Contractor of any deficiencies, coordinate availability of Government property, and coordinate site entry of Contractor personnel.

1.10 Key Personnel: The following personnel are considered key personnel by the Government: Contract Manager, Alternate Contract Manager, and Systems Administrator. The Contractor shall provide a Contract Manager who shall be responsible for the performance of the work. The name of this person and an Contract Manager, who shall act for the Contractor when the Contract Manager is absent. The Contract Manager and Alternate Contract Manager shall have full authority to act for the Contractor on all contract matters relating to daily operation of this contract. The Contract Manager or Alternate Contract Manager shall be available between 0800-1700 normal business hours as discussed in paragraph 1.5.4. Qualifications for all key personnel are listed below:

- a. **Contract Manager and Alternate Contract Manager:** The Contract Manager and Alternate Contract Manager must demonstrate a minimum of 3 years' experience within the past 10, working as a Contract Manager.
- b. Quality Control Manager: Reserved
- c. Systems Administrator: Reserved
- d. Special Qualifications: Reserved

1.11 Identification of Contractor Employees: All Contractor personnel attending meetings, answering Government telephones, and working in other situations where their Contractor status is not obvious to third parties must identify themselves, to include proper marking of signature blocks in correspondence, to avoid creating an impression in the minds of members of the public that they are Government officials. The Contractor shall ensure that all documents or reports, produced by Contractors are suitably marked as Contractor products or that Contractor participation is appropriately disclosed.

1.12.1 Contractor Workspace: Reserved

1.12.2 Contractor Identification Badges: Contractor personnel shall be easily identifiable through the display of badges IAW Contractor Identification (AE Reg. 27-715). (For services at Child, Youth, and School Services where children are present, recommend personnel wear identification when children are present.) The Contractor (to include subcontractors) shall provide each of their employees an easily readable identification badge. The badge must include the employee's name, the contractor's company name, the functional area of

assignment, and a recent color photograph of the employee. Companies with U.S. Government contracts will require their employees to wear these badges visibly when performing work at Government workplaces unless this would be impractical because of mission or safety considerations.

1.13 **Contractor Travel:** Reserved

1.14 **Other Direct Costs (ODCs):** Reserved

1.15 **Data Rights:** Reserved

1.16 Non-Disclosure Requirements: Performance under this contract may require the Contractor to access data and information proprietary to a Government agency, another Government Contractor, or of such nature that its dissemination or use other than as specified in this work statement would be adverse to the interests of the Government or others. The Contractor and Contractor personnel, shall not divulge or release data or information developed, or obtained under performance of this PWS, except to authorized Government personnel or upon written approval of the Contracting Officer. The Contractor shall not use, disclose, or reproduce proprietary data, which bears a restrictive legend, other than as specified in this PWS. All documentation showing individual names or other personal information shall be controlled and protected under the provisions of the Privacy Act of 1974, Public Law 93-579, 5 United States Code (U.S.C.) Section 552a.

1.17 Protection of Government and Contract Information: Per Public Use Notice of Limitations stated by Defense imagery Management operations Center and contained at www.dimoc.mil/resources/limitations/, the Contractor shall not cite any information (e.g., contract information, pictures, locations, etc.) obtained through this contract on any hard copy or digital marketing tools to include its company website.

1.18 **Non-Disclosure Statements:** Reserved

1.19 Organizational Conflict of Interest (OCI): Contractor and subcontractor personnel performing work under this contract may receive, have access to or participate in the development of proprietary or source selection information (e.g., cost or pricing information, budget information or analyses, specifications or work statements, etc.) or perform evaluation services which may create a current or subsequent OCI as defined in FAR Subpart 9.5. The Contractor shall notify the Contracting Officer immediately whenever it becomes aware that such access or participation may result in any actual or potential OCI and shall promptly submit a plan to the Contracting Officer to avoid or mitigate any such OCI (Deliverable A.004). The Contractor's OCI Mitigation Plan will be determined to be acceptable solely at the

discretion of the Contracting Officer. In the event the Contracting Officer unilaterally determines that any such OCI cannot be satisfactorily avoided or mitigated, the Contracting Officer may affect other remedies as he or she deems necessary, including prohibiting the Contractor from participation in subsequent contracted requirements which may be affected by the OCI.

1.20 **Phase In /Phase Out Period:** Reserved

1.21 **Safety:** The following provides information pertaining to Safety and Accident Prevention:

1.21.1 **Safety Laws and Regulations:** The Contractor shall comply with FAR 52.236-13, Accident Prevention (November 1991), the references contained therein (e.g. U.S. Army Corps of Engineers Safety and Health Requirements Manual (EM 385-1-1)), and applicable Host Nation Safety Laws and Regulations.

1.21.1.1 In addition to other requirements for accident prevention specified in this contract, the Contractor shall comply with all Federal and State requirements governing safety and health protection on construction sites and other applicable occupational safety codes. In particular, within the requirements of the German Safety and Health Protection Regulation for Construction Sites, Baustellenverordnung (BaustellV) (current version in effect), the Contractor shall coordinate all work at the site, including development of plans and providing advance notices.

1.21.1.2 The Contractor shall complete and provide safety notifications required by host nation authorities and will provide documented proof to the Contracting Officer via the COR, no later than 5 days (A.005) that such notifications were made.

1.21.1.3 **OCONUS-Italy language:** Reserved

1.22 **Environmental Compliance:** The following provides information on Environmental Compliance.

1.22.1 **Environmental Compliance Plan and Program:** The Contractor shall ensure that all of its operations and activities conform to the USAG Kaiserslautern's environmental policy requirements. The Contractor shall ensure that all contract personnel have received appropriate levels of training in the handling, use, storage, and disposal of hazardous material as it pertains to this contract. In addition, the Contractor shall ensure that all Contractor controlled activities on USAREUR installations are conducted in accordance with the USAG Kaiserslautern Spill Prevention and Response Plan, Solid Waste Policy, Final Governing Standards-Germany, and applicable German environmental laws/standards.

If contract requires the use of hazardous materials/substances, the Contractor shall, prior to mobilization, inform the Contracting Officer (KO)/Contracting Officer Representative (COR) of the types and amount of hazardous materials needed to be brought onto the Garrison in order to fulfill their contractual requirements. Contractor shall also provide the Safety Data Sheets, storage requirements and proposed storage locations for the hazardous material. The KO/COR shall provide a copy of this information to the Garrison Fire Department and the Environmental Division.

Prior to contract completion, the Contractor shall remove all hazardous materials and dispose of all hazardous waste (HW) generated in accordance with Final Governing Standards Germany and applicable German environmental laws/standards. A copy of the HW disposal documents must be forwarded to the HW program manager at the Department of Public Works (DPW) Environmental Division.

The Contractor shall clean up at its own expense any contamination, including soil and ground water contamination caused by hazardous materials/substances brought onto the garrison, or by his activities. The Contractor must immediately report any hazardous substance spills to the Garrison Fire Department at CIV 0611-705-117.

Current Environmental Stewardship Policy (Command policy Number 11) is available at the Garrison website:

https://home.army.mil/rheinland-pfalz/application/files/4716/0501/6392/CP_11_Environmental_Stewardship.pdf

Current Hazardous Waste Handling Policy (Garrison policy Number 10) is available at the Garrison website:

https://home.army.mil/rheinland-pfalz/application/files/6216/0501/8981/SOP_Hazardous_Waste.pdf

Current Refuse Collection and Recycling Policy is available at the Garrison Website

https://home.army.mil/rheinland-pfalz/application/files/5616/0501/9038/SOP_Refuse_and_Recycling.pdf

1.23 Required Training: The following provides information on training requirements.

1.23.1 **Anti-Terrorism (AT) Level I Training:** Reserved

1.23.2 **OPSEC Training:** Reserved

1.23.3 **iWATCH Training**: Reserved

1.23.4 **Information Assurance (IA) Training** Reserved

1.23.5 **Personally Identifiable Information (PII)**: Reserved

1.24 **Authorization to Perform Services in Germany**

1.24.1 Contractors performing services in the Federal Republic of Germany (FRG) shall comply with German law. The Contractor shall determine whether performance requires registration with German authorities or authorization to do business in Germany and, if so, shall comply with all requirements. Whether or not registration or authorization to do business is required, the Contractor shall determine what documents or authorization its employees and any subcontractor employees must possess to work in Germany. The Contractor shall ensure affirmatively that its employees and subcontractor employees possess such documents or authorizations.

1.24.2 Contractor employees who: (a) are not nationals of Germany or other European Union countries; (b) are not members of the force, the civilian component or their dependents; or c) do not have assimilated status under Articles 71, 72, or 73 of the Supplementary Agreement to the North Atlantic Treaty Organisation Status of Forces Agreement shall possess work and residence permits.

1.24.3 By acceptance of and performance under this contract, the Contractor affirms that it has complied with the requirements above. Compliance with this requirement and German law is a material contract requirement. Noncompliance by the Contractor or subcontractor at any tier shall be grounds for issuing a negative past performance evaluation and terminating this contract for default.

1.25 **Notice to Proceed**: Reserved

1.26 The Contractor certifies that all electrical and electronic equipment and appliances furnished are in compliance with the Electromagnetic Compatibility Directive 2014/30/EU. For radio transmitting and telecommunication devices furnished in the FRG, the Contractor certifies that all such items have been approved by the Bundesamt fuer Zulassungen. Any penalties for noncompliance imposed by the Government of the European country where the equipment or appliances are delivered or product liability claims caused by interference generated by furnished appliances or equipment are solely the responsibility of the Contractor.

1.27 **Government Property and Services**: Reserved

1.28 **Conventional Forces Europe (CFE) Treaty Verification Inspection:** Reserved

PART 2

DEFINITIONS & ACRONYMS

2. Definitions and Acronyms

2.1 Definitions: Although not inclusive of every term used within this PWS, the following provides a list of definitions used throughout this PWS and commonly used in the acquisition field.

Contracting Officer – means a person with the authority to enter into, administer, and/or terminate contracts and make related determinations and findings on behalf of the Government. Note: The only individual who can legally bind the Government.

Contracting Officer's Representative (COR) – As defined in DFARS 202.101, means an individual designated and authorized in writing by the Contracting Officer to perform specific technical or administrative functions. DoD Instruction (DoDI) 5000.72, Part II Definitions states the following when defining a COR: “Defined in subpart 202.101 of Reference (f). Any individual delegated responsibilities pursuant to subpart 1.602-2 of Reference (e), regardless of local terminology, must be certified in accordance with this instruction. For example, local terminology can be COR, Contracting Officer’s technical representative, technical point of contact, technical representative, alternate COR, administrative COR, assistant COR, line item manager, task order manager, quality assurance personnel, quality assurance evaluator, or COR management.” In addition, Army Regulation 70-13, Chapter 2, paragraph 2-2g, states, in part, the following when providing other surveillance support personnel to assist the COR when needed, “...These other surveillance support personnel may serve as on-site representatives of the COR in performance of actual contract surveillance if they meet all COR requirements and have been appointed by the Contracting Officer as alternate CORs.”

Contractor – means a supplier or vendor awarded a contract to provide specific supplies or service to the Government. The term used in this contract refers to the prime.

Contractor-acquired Property - means property acquired, fabricated, or otherwise provided by the Contractor for performing a contract and to which the Government has title.

Day – means, unless otherwise specified, a calendar day.

Defective Service – means a service output that does not meet the standard of performance associated with the Performance Work Statement.

Deliverable – means anything that can be physically delivered, but may include non-manufactured things such as meeting minutes or reports.

Government-furnished Property – As reflected in FAR 52.245-1, Government-furnished Property “means property in the possession of, or directly acquired by, the Government and subsequently furnished to the Contractor for performance of a contract. Government-furnished property includes, but is not limited to, spares and property furnished for repair, maintenance, overhaul, or modification. Government-furnished property also includes contractor-acquired property if the contractor-acquired property is a deliverable under a cost contract when accepted by the Government for continued use under the contract

Government Property - means all property owned or leased by the Government. Government property includes both Government-furnished and Contractor-acquired property. Government property includes material, equipment, special tooling, special test equipment, and real property. Government property does not include intellectual property and software.

Property Administrator - means an authorized representative of the Contracting Officer appointed in accordance with agency procedures, responsible for administering the contract requirements and obligations relating to Government property in the possession of a Contractor.

High Level Objective (HLO) – means a key overarching result-based objective for a project necessary to achieve the project’s vision. HLOs are similar to Level 2 in a Work Breakdown Structure. Each HLO may contain several statements to flesh out the areas necessary to meet the objective.

Physical Security – means that part of security concerned with physical measures designed to safeguard personnel; to prevent unauthorized access to equipment, installations, material, and documents; and to safeguard against espionage, sabotage, damage, and theft.

Quality Assurance – (or Government contract quality assurance) means the various functions, including, inspection, performed by the Government to determine whether a Contractor has fulfilled the contract obligations pertaining to quality and quantity.

Quality Assurance Surveillance Plan (QASP) – means the key Government-developed surveillance process document, and is applied to Performance-Based Service Contracting (PBSC). The QASP is used for managing Contractor performance assessment by ensuring that systematic quality assurance methods validate that Contractor quality control efforts are timely, effective, and are delivering the results specified in the contract or task order. The QASP directly corresponds to the performance objectives and standards (i.e., quality, quantity,

timeliness) specified in the Performance Work Statement (PWS). It provides specific details on how the Government will survey, observe, test, sample, evaluate, and document Contractor performance results to determine if the Contractor has met the required standards for each objective in the PWS. The QASP, with very few if any exceptions, is an internal to Government document.

Quality Control – means all necessary measures taken by the Contractor to assure that the quality of an end product or service shall meet contract requirements.

Statement – means the specific results-based activities required to satisfy HLOs. A statement contains a result, the context of the statement, and the required action(s). Statements focus on “what” is to be accomplished; however they are not prescriptive in describing “how” the outcome is to be achieved. Each HLO may have several statements to flesh out the areas necessary to meet the objective. Statements are similar to Level 3 in a Work Breakdown Structure.

Subcontractor – means one that enters into a contract with a prime Contractor. The Government does not have privity of contract with the subcontractor.

Work Day - The number of hours per day the Contractor provides services in accordance with the contract.

Work Week - Monday through Friday, unless otherwise specified.

2.2 Acronyms: Although not inclusive of every term used within this PWS, or that may be included in an acquisition, the following provides a list of acronyms commonly used in the acquisition field.

ACOR	Alternate Contracting Officer's Representative
AE	Army in Europe
AFARS	Army Federal Acquisition Regulation Supplement
AHA	Activity Hazard Analysis
AOR	Area of Responsibility
AR	Army Regulation
AT	Anti-terrorism
ATCTS	Army Training Certification Tracking System
CAC	Common Access Card
CCE	Contracting Center of Excellence
CFE	Conventional Forces Europe
CFR	Code of Federal Regulations

CONUS	Continental United States (excludes Alaska and Hawaii)
COR	Contracting Officer's Representative
COTS	Commercial-Off-the-Shelf
DA	Department of the Army
DD250	Department of Defense Form 250 (Receiving Report)
DD254	Department of Defense Contract Security Requirement List
DFARS	Defense Federal Acquisition Regulation Supplement
DMDC	Defense Manpower Data Center
DoD	Department of Defense
FAR	Federal Acquisition Regulation
FRG	Federal Republic of Germany
GFP	Government Furnished Property
HLO	High Level Objective
HIPAA	Health Insurance Portability and Accountability Act of 1996
IA	Information Awareness
IAC	Installation Access Control
IACO	Installation Access Control Office
ID	Identification
IGCE	Independent Government Cost Estimate
IT	Information Technology
JTR	Joint Travel Regulation
JPAS	Joint Personnel Adjudication System
KO	Contracting Officer
OCI	Organizational Conflict of Interest
OCONUS	Outside Continental United States (includes Alaska and Hawaii)
ODC	Other Direct Costs
OPSEC	Operations Security
PA	Property Administrator
PII	Personally Identifiable Information
PIPO	Phase In/Phase Out
POC	Point of Contact
PRS	Performance Requirements Summary
PWS	Performance Work Statement
QA	Quality Assurance
QAP	Quality Assurance Program
QASP	Quality Assurance Surveillance Plan
QC	Quality Control
QCP	Quality Control Program
RA	Requiring Activity
RCO	Regional Contracting Office

SIGE	Safety and Health Protection Plan
TE	Technical Exhibit
USAG	United States Army Garrison
USAREUR	United States Army Europe
UOM	Unit of Measure

PART 3
GOVERNMENT PROPERTY (GP) AND SERVICES

3. Government Property and Services

3.1 Services: Reserved

3.2 Facilities: Reserved

3.3 Utilities: Reserved

3.4 Equipment: Reserved

3.5 Special Tooling and Test Equipmen: Reserved

3.6 Materials: Reserved

3.7 Training: Reserved

3.8 Common Access Cards (CAC): Reserved

PART 4

CONTRACTOR FURNISHED ITEMS AND SERVICES

4. Contractor Furnished Property and Services

4.1 General: The Contractor shall furnish all personnel, equipment, supplies, facilities, transportation, tools, materials, supervision, and other items necessary to perform services as described in the PWS and associated contract documents, except for those items specified in Part 3 Government Property and Services.

4.2 Secret Facility Clearance: Reserved

4.3 Materials: The Contractor shall provide personal Protective Equipment, PPE, Office supplies and materials such as, paper, printer cartridges, and all other office supplies and materials needed in the performance of this contract; automated data processing equipment (ADPE) such as computers, copy, printer, and fax machines. All ADPE connecting to the Government network shall be reviewed and approved by the Government prior to purchase by the Contractor; office furniture; cellular and desktop telephones, etc. The Contractor may request a letter of authorization from the Contracting Officer to utilize Government supply sources such as Government Services Administration (GSA), UNICOR, etc.

4.4 Vehicles and Equipment: The Contractor shall provide all necessary vehicles (no GSA vehicles provided to Contractor for use on this contract) and equipment to perform the work described in this PWS. Contractor's vehicles and equipment used in the performance of this contract shall meet all local and Host Nation traffic, safety, and environmental requirements. Vehicles and equipment found to be unsafe shall be removed from the installation and replaced. The Contractor shall not use any Government-owned tools, materials, or parts to maintain Contractor vehicles and equipment without prior written approval of the Contracting Officer. The COR may inspect the Contractor's vehicles and equipment at any time.

4.5 Background Checks: The Contractor shall provide Background Checks in accordance with this PWS and as determined by the Government.

PART 5 SPECIFIC TASKS

5.1 Basic Services: Non-personnel service to include, but not limited to, providing all personnel, equipment, tools, supervision and other items necessary to maintain equipment as listed on the equipment list. Services must be in compliance with

5.2 Tasks: Tasks consist of the following:

5.2.1 Contractor Resources Management

5.2.1.1 Work Order Reception: Contractor shall establish a Work Reception System for tracking all requests for recurring and preventative maintenance work or services and maintaining active and historical workload records.

- (a) STD: The Work Reception Desk and telephone shall be attended continuously during normal working hours including lunchtime. The contractor shall provide voice messaging service management and telephonic support in English Language is required. The contractor shall return calls for repair service left on voice messaging services within 1 business day.

AQL: 95% Compliant

- (b) STD: The Work Reception Desk shall inform the Contracting Officer through the COR on expected timelines for phone support, field service engineer on-site time, or expected delivery date of parts.

AQL: 95% Compliant

- (c) STD: The contractor shall only accept calls for service/repairs from the COR or alternate COR on this contract.

AQL: Pass or Fail

- (d) STD: The contractor shall notify the POC (Point of Contact) of the exact date and time within ten (10) days prior to scheduled preventive maintenance services.

AQL: 95% Compliant

- (e) STD: The contractor's service technician shall provide remote support / troubleshooting, if the repair condition allows, over the phone when available to limit equipment downtime.

AQL: 95% Compliant

- (f) All contractors' representatives will be sufficiently fluent in the English language to effectively communicate with LRMC's representatives.

AQL: 95% Compliant

5.2.1.2 ROUTINE/EMERGENCY SERVICE:

- (a) STD: Prior to performing work, a contractor's representative will report to the Clinical Engineering Branch (LRMC, Bldg. 3760) and sign for a Visitor's Badge. The badge must be returned to the Clinical Engineering Branch prior to departing LRMC.

SQL: Pass or Fail

- (b) STD: The contractor shall maintain the equipment listed in proper operating condition as described in the manufacturers' literature.

AQL: Pass or Fail

- (c) STD: Unscheduled emergency / contingency visits to sites in order to address repairs shall be accepted only from the COR, Alternate COR, or KO.

AQL: Pass or Fail

- (d) STD: Completion of Unscheduled Repairs. For On Site routine services, Contractor's technician must be on site within Five (5) working days and the repair must be completed within two (2) working days after contractor arrives on-site.

AQL: 90% Compliant

- (e) STD: Contractor must furnish all software updates issued by the equipment manufacturer (if applicable).

AQL: Pass or Fail

- (f) STD: Within 5 days after completion of service services (scheduled and/or unscheduled) are completed, the contractor's service technician shall prepare and initial a legible, preferably typed service report in English and present it to the COR for signature. This report shall include the following information:

- i. Services performed
- ii. What systems or components failed (reason for call)
- iii. Start date/time
- iv. Completion date/time
- v. Name of service technician
- vi. Part numbers, nomenclatures, quantities and prices of replaced parts
- vii. Travel time

- viii. Repair time
 - ix. Travel, repair and parts cost
- SQL: 95% Compliant

(g) STD: The contractor “must” obtain authorization from the COR, or alternate COR, prior to starting “any” repair that is attributable to instrument abuse and/or instrument misuse, theft, neglect, fire and anything else not covered in this contract. If these types of repairs are performed without the prior approval of the COR, or alternate COR, the contractor shall not be paid by the U.S. government. If approved, a separate purchase order shall be issued prior to performance.

SQL: Pass or Fail

(h) STD: Repairs that cannot be performed at the Government site may be performed at the contractor’s site provided the POC approves removal of the equipment. The contractor’s representative shall sign a receipt for the equipment. The repaired equipment shall be returned to the original premises within (2) two working days or the contractor will provide a temporary replacement at no additional cost to the government.

SQL: 95% Compliant

(i) STD: If the equipment is out of service for longer than 48 hours (from the time of service technician’s arrival on site), the contractor shall notify the COR or alternate COR in writing as to the reason(s) (i.e. parts nonavailability, etc.) for non-compliance with paragraph 5.2.1.2. The contractor shall then provide, at no additional cost to the U.S. government, a substitute piece of monitoring equipment to be utilized by the government until the original equipment is fully repaired and accepted for service.

AQL: 90% Compliant

(j) STD: All correspondence, service reports and invoicing shall be in English and clearly legible.

SQL: Pass or Fail

(k) STD: The U.S. government will make the equipment available to the contractor for servicing at such time and duration necessary to perform needed repairs.

AQL: Pass or Fail

5.2.1.3 Reserved

5.2.1.4 PREVENTATIVE MAINTENANCE INSPECTIONS:

(a) STD: Preventive maintenance inspections shall be performed in accordance with the manufacturer's standards and procedures. The preventive maintenance inspection shall include, but is not limited to safety, calibration, completed operational testing, adjustments and cleaning of equipment to which the operator does not have access. It also includes installation of all parts required to ensure proper operation.

SQL: 95% Compliant

(b) STD: Safety shall be performed in accordance with NFPA 99, Chapter 7. This regulation can be reviewed in full text at the Equipment Management Branch. When required, the contractor's service representative will comply with the Occupational Safety and Health Administration lockout/tagout standards while performing maintenance on equipment. The government and the contractor's service representative will exchange hazard communication information before the commencement of any repair.

SQL: Pass or Fail

(c) STD: Preventive maintenance shall include cleaning, replacement of missing or defective parts, modules, covers, a complete inspection, an inspection or repair of cables, plugs and wiring, a thorough function test of the input and safety circuits, checking the topicality of the operator manuals and information and training of new personnel. The safety inspection shall include a mechanical check (examination of the casing, the readings, user interfaces, connections, cables, labeling, soiling that would affect safety, etc.) and function test, as services include written documentation of servicing, cleaning, adjustments, operational tests and calibration/verification in accordance with the manufacturer's specifications and all applicable U.S. government and Army regulations.

AQL: 95% Compliant

(d) STD: The contractor's qualified service technician shall provide preventative maintenance services to each site per the manufacturers recommended schedule.

SQL: 95% Compliant

(e) STD: All preventive maintenance inspections/verifications/calibrations shall be performed in accordance with the manufacturer's service literature, which will be located with the equipment or provided by the contractor.

SQL: 95% Compliant

(f) Contractor's service representatives will be company trained and have a minimum of 2 years' experience working on contracted equipment.

SQL: 95% Compliant

- (g) Contractor must furnish all software updates issued by the equipment manufacturer.

SQL: Pass or Fail

- (h) Contractor shall have access to all necessary diagnostic software (if applicable).

SQL: Pass or Fail

5.2.1.5 Replacement Parts

- (a) All replacement parts shall be certified (*), new and have the same material as the original part. The original design of the equipment shall not be changed, modified or the required functional capacities altered. (*) Certified parts are parts that comply with standards set forth in 21 CFR 1020.30 (applicable to radiation producing equipment only), for installation in medical equipment. This regulation can be reviewed in full context at the Equipment Management Branch.

SQL: Pass or Fail

- (b) In place of new parts, factory refurbished exchange parts per the manufacturer's spare parts procedures and policies may be used. The contractor's representative shall inform the POC and request approval prior to the use of a refurbished part.

SQL: Pass or Fail

- (c) The contractor shall furnish a detailed listing of any replacement parts provided under the terms of this contract. This listing shall contain the part number, nomenclature, normal life expectancy and price and shall indicate whether the item is an exchange part.

SQL: 95% Compliant

5.2.1.6 Receipt of Services

- (a) Receipt of services will be acknowledged after receipt of all completed service reports for any scheduled calibration, preventive maintenance or unscheduled routine and emergency service performed during the respective quarter. Receiving reports will not be prepared and submitted until all such completed service reports containing all required information are in the possession of the Chief, Supply Management Branch or his/her designated representative.

AQL: Pass or Fail

(b) The Contractor shall complete DD Form 2164 (X-Ray Verification/Certification Worksheet) in accordance with instructions provided in TB 38-750-2 if the equipment requires such documentation. A continuation sheet shall be attached to the DD Form 2164 indicating the manufacturer, model, serial number and date of the calibration expiration of all items of test, measurement and diagnostic equipment used to perform the calibration.

AQL: Pass or Fail

(c) If the Contractor's calibration equipment produces a printed summary of the calibration procedure used, attach the printed summary to the DD Form 2164. Ensure the heading of the DD Form 2164 is filled out and the form is properly signed.

AQL: Pass or Fail

5.2.2 Reserved

5.2.2.1 Reserved

5.3 Service Contract Reporting

5.3.1 System for Award Management (SAM) Service Contract Report (SCR): The Contractor shall report ALL Contractor labor hours (including subcontractor labor hours) required for performance of services provided under this contract for the Defense Health agency via a secure data collection site. The Contractor is required to completely fill in all required data fields through the following web address: www.sam.gov. Reporting inputs will be for the labor executed during the period of performance during each Government FY, which runs from October 1 through September 30. While inputs may be reported any time during the FY, all data shall be reported no later than October 31 of each calendar year. Contractors may direct questions to the help desk by clicking "View Assistance for SAM.gov" which is located at the top of the SAM.gov website. From there, you can the select "Contact Our Service Desk" which will allow you to contact SAM.gov directly. If contract period of performance ends prior to September 30, the Contractor has 30 calendar days from end date of the contract to complete the SAM SCR requirement.

Steps for Submitting a Service Contract Report (SCR)

1. Go to www.sam.gov and log in.
2. Select Entity Registrations and then select Service Contract Reporting.
3. SAM displays your entities which have service contracts and meet the reporting criteria. Select View by entity to see the service contracts for each entity.

4. Next, select Add for the service contract against which you want to create a Service Contract Report. Each service contract which meets the FAR Subpart 4.1703 reporting thresholds is displayed.
5. You will be taken to the Complete Service Contract Report page. SAM displays the contract details and allows you to report. You are required to enter the following information:
 - **Total Amount Invoiced:** Total dollar amount invoiced for services performed during the previous Government fiscal year under the contract (this amount should include the prime and any subcontract amount).
 - **Prime Contractor Hours Expended:** Prime contractor direct labor hours expended on the services performed during the previous Government fiscal year. The amount you enter is automatically divided by 2,080 hours to calculate a Full Time Employee (FTE) equivalent, displayed under the Prime Contractor Hours Expended as Prime Contractor FTEs.
6. Report any required Tier 1 subcontractor information by selecting the Add Tier 1 Subcontract Information button.
7. When you are ready to submit the report, select Submit. This saves your report and returns you to the Select Service Contract page where you can create other SCRs or edit an existing SCR.

Steps for Editing a Submitted Report

1. Follow steps 1-4 above. Any previously submitted SCR will have a View/Edit button instead of an Add button.
2. Select View/Edit. You will be able to view the current SCR, edit available fields, and resubmit the record. You can also delete previously entered information altogether.

STD: All information provided by the Contractor shall be accurate, complete, and not exceed suspense dates noted in corresponding paragraph above.

AQL: 100% Compliant

PART 6

APPLICABLE PUBLICATIONS

6. Applicable Publications (Current Editions): The following publications, manuals, regulations, etc. are mentioned in this PWS and are listed below.

- 6.1.1 Department Of Defense Contract Security Classification Specification (DD Form 254)
- 6.1.2 Department of Defense Security Agreement (DD Form 441)
- 6.1.3 National Industrial Security Program Operating Manual (DoD 5220.22-M)
- 6.1.4 Installation Access Control (AE Reg 190-16)
- 6.1.5 The Army Physical Security Program (AR 190-13)

- 6.1.6 Contractor Identification (AE Reg 27-715)
- 6.1.7 U.S. Army Corps of Engineers Safety and Health Requirements Manual (EM 385-1-1)
- 6.1.8 Information Assurance Workforce Improvement Program (DoD 8570-M)
- 6.1.9 Information Assurance Training Certification and Workforce Management Directive (DoDD 8570.01)
- 6.1.10 Information Assurance (AR 25-2)
- 6.1.11 Electromagnetic Compatibility Directive (2004/108/EC)
- 6.1.12 Policies and Procedures for Property Accountability (AR 735-5)

PART 7

TECHNICAL EXHIBIT LISTING

7 Technical Exhibit List

7.1 Technical Exhibit A – Performance Requirements Summary

7.2 Technical Exhibit B – Deliverables Schedule

7.3 Technical Exhibit C - Equipment list

TECHNICAL EXHIBIT A
Performance Requirements Summary (PRS)

PERFORMANCE OBJECTIVE	STANDARD	PERFORMANCE THRESHOLD ACCEPTABLE QUALITY LEVELS (AQLs)	METHOD OF SURVEILLANCE
PRS#1 - PWS para (5.2.1.2) Maintain the equipment listed in proper operating condition as described in the manufacturers' literature.	The contractor shall maintain the equipment listed in proper operating condition as described in the manufacturers' literature.	Pass or Fail	Periodic Inspection
PRS#2 - PWS para (5.2.1.2) The contractor shall Complete unscheduled repairs. For On Site routine services.	Contractor's technician must be on site within Five (5) working days and the repair must be completed within two (2) working days after contractor arrives on-site.	90% Compliant	100% Inspection
PRS#3 - PWS para (5.2.1.4) The contractor shall perform preventive maintenance in accordance with the manufacturer's standards and procedures.	The preventive maintenance inspection shall include, but is not limited to safety, calibration, completed operational testing, adjustments and cleaning of	95% Compliant	Periodic Inspection

	equipment to which the operator does not have access. It also includes installation of all parts required to ensure proper operation.		
PRS#4 – PWS para (5.2.1.4) The contractor's qualified service technician shall provide preventative maintenance services to each per the manufacturer's recommended schedule.	Scheduled services for preventive maintenance, safety and calibration will be performed as required by the manufacturer's specifications and this statement of work. The equipment shall be repaired and operational after initial response by the contractor.	95% Compliant	100% Inspection
PRS #5 - Service Contract Reporting (SCR) PWS 5.3 and CLIN 0001 and CLIN 0002	The Contractor shall completely fill in all the information in the format using the following web address www.sam.gov The contractor shall provide the estimated total cost (if any) incurred to comply with this reporting requirement. Reporting period	Zero Deviation from Standard	100% Inspection

	shall be the period of performance not to exceed 12 months ending September 30 of each government fiscal year and must be reported by 31 October of each calendar year. If contract period of performance ends prior to September 30, the contractor has 30 days from end date of the contract to complete the SCR requirement.		
--	--	--	--

TECHNICAL EXHIBIT B

DELIVERABLES SCHEDULE

DELIVERABLE	FREQUENCY	# OF COPIES	MEDIUM/FORMAT	SUBMIT TO
A.001 Quality Control Plan (paragraph 1.5.1)	Once within 10 days after contract award and no later than 30 days prior to any anticipated changes.	One each to the Contracting Officer	Electronic delivery via email using Microsoft Office Word or Power Point.	To the Contracting Officer through the Contracting Officer's Representative (COR)
A.002 Quality Control Plan issues (paragraph 1.5.1.2)	Within 24 hours of identifying issues or potential issues that might affect the performance of the contract.	One each to the Contracting Officer	Electronic delivery via email using Microsoft Office Word or Power Point.	To the Contracting Officer through the Contracting Officer's Representative (COR)
A.003 Contract Termination or Cancellation badge turn-in (para 1.6.3.3)	Within 4 business days of contract termination or cancellation, turn in all badges to COR or installation command.	One each to the Contracting Officer	ID's and Badges	COR or Installation command.
A.004 Organizational Conflict of Interest (OCI) (para 1.19)	The contractor shall notify the Contracting Officer immediately whenever it becomes aware that such access or participation may result in any actual or potential OCI and shall	One each to the Contracting Officer	Electronic delivery via email using Microsoft Office Word or Power Point.	To the Contracting Officer through the Contracting Officer's Representative (COR)

DELIVERABLE	FREQUENCY	# OF COPIES	MEDIUM/FORMAT	SUBMIT TO
	promptly submit a plan to the Contracting Officer to avoid or mitigate any such OCI.			
A.005 Safety notifications. (para 1.21.1.2)	The Contractor shall complete and provide safety notifications required by host nation authorities and will provide documented proof to the Contracting Officer via the COR, no later than 5 days that such notifications were made	One each to the Contracting Officer	Electronic delivery via email using Microsoft Office Word or Power Point.	To the Contracting Officer through the Contracting Officer's Representative (COR)
A.006 Background Checks Para 1.6.4	As needed for base access. A background check will be requested by the contracting officer through the COR for a base access request if necessary. If requested, the contractor must supply the requested background check or Polizeiliche Fuehrungszeugnis within 10 days of receipt. For each unfavorable background check, the report will be provided within 3 days.	One each to the Contracting Officer	Electronic delivery via email using Microsoft Office Word or Power Point.	To the Contracting Officer through the Contracting Officer's Representative (COR)

TECHNICAL EXHIBIT C

Equipment List

CLIN	ECN	Nomenclature	Model	Serial	Location
0001	035024	SCANNING SYSTEM, COMPUTED TOMOGRAPHY	AQUILION 64	JDA1072183	CT Department
0002	103208	SCANNING SYSTEM, COMPUTED TOMOGRAPHY	AQUILION ONE	6LB1962032	CT Department