

Performance Work Statement

16 Mar 2026

General Description

1. Contractor shall provide and install all equipment necessary to perform Hematology testing according to this performance work statement.
2. All equipment must be FDA approved.
3. Must be capable of host-query bidirectional interface with the existing Laboratory Information System (LIS), Composite Healthcare System (CHCS).
4. The system must fit within an 18" by 26" countertop space.
5. Provide one Uninterruptible Power Supply (UPS) unit with the instrument.
6. Instrument software must be Windows based.
7. Provide service for the instrument during normal business hours, excluding holidays and weekends for the duration of the contract. Technical support via telephone must be available for any reagent, software, and instrument problems between the hours of 8:00 AM and 4:00 PM CST. On-site service must be available within 24 hours of notification. Defective parts and software shall be replaced at no charge.
8. Provide semiannual and as needed preventive maintenance visits to include instrument calibration which will be included in the price of the contract.
9. Provide instrument software/hardware upgrades at no cost during the term of the contract.
10. Provide off-site training for two key operators during the first contract year to include transportation, food, and lodging.
11. Additional on-site training shall be provided as necessary and at no extra charge.
12. Perform correlation, precision, linearity, and all other studies required during installation of system. Vendor will provide all materials necessary for above studies at vendor's expense.
13. The instrument must have automatic reagent expiration date tracking.
14. The instrument must have on-board user manuals for troubleshooting.
15. Instrument must use only one probe for both automated and manual specimen sampling to minimize QC requirements.
16. Reagents and calibrators must be ready-to-use: reagents must not require preparation prior to loading on analyzer. Reagent kits must have 2D bar-coded labels.

17. The analyzer must have a minimum throughput of 40 hematology tests per hour.
18. The analyzer must be able to achieve a lower linearity down to 0 for WBC, RBC, and Platelet parameters to allow for clinical diagnosis of critical conditions.
19. The analyzer must be able to perform full hematology test with as little as 25 µL of sample.
20. The on-board software must be capable of storing 2 years of quality control data with Levey Jennings charts.
21. The software must include patient moving averages (Xbar).
22. The bid must include cost per reportable pricing which includes the full cost of the equipment, service, reagents, quality control materials, training, and consumables.
23. Analyzer must be able to support test menu specified with no changes, and must be upgradeable to meet future testing needs.
24. Test Menu: The analyzer must be capable of reporting the same parameters that we currently report with our Complete Blood Counts:

Complete Blood Count – 4500/year

- WBC
- RBC Count
- HGB
- HCT
- MCV
- MCH
- MCHC
- PLATELET COUNT
- RDW
- MPV
- NEUT%
- LYMPH%
- MONO%
- EOS%
- BASO%
- NEUT#
- LYMPH#
- MONO#
- EOS#
- BASO#