

**PURCHASE DESCRIPTION
FOR
*AFPSL – Load Cells Exchange Kits***

1.0 Scope This procurement is for additional CS Series Load Cells, 9150-2-300 digital indicators, adapters, and travel cases that were originally procured under FA2263-11-C-0002. The current exchange pool is depleted and additional units are required.

2.0 Functional Requirements

2.1. The contractor shall deliver the following load cells and indicators to the Air Force Primary Standards Lab:

- 2.1.1. Tovey CS10-2.5K-B300
- 2.1.2. Tovey CS10-10K-B300
- 2.1.3. Tovey CS20-30K-B300
- 2.1.4. Tovey CS20-60K-B300
- 2.1.5. Tovey 9150-2-300
- 2.1.6. Tovey 2133-0007-02
- 2.1.7. Tovey 2133-0009-02
- 2.1.8. Tovey 2133-0008
- 2.1.9. Tovey 2133-0010
- 2.1.10. Tovey 210506
- 2.1.11. Tovey 2133-0012
- 2.1.12. Tovey TC-10-300
- 2.1.13. Tovey TC-20-30K-300
- 2.1.14. Tovey TC-20-60K-300
- 2.1.15. Tovey TC-Inst-300

2.2. Compatibility: The load cells, digital indicators, adapters, and travel cases shall be compatible with each other along with the Tovey Engineering, Inc. P/N C3-100 Vertical Force Press.

2.3. Uncertainty: The uncertainty of the C3-100 Vertical Force Press shall not exceed 0.025% of reading when utilizing the load cells and digital indicators being provided.

3.0 Physical/Mechanical Requirements

3.1. Storage Box(s): A simple box shall be provided that will safely store each load cell and digital indicator while on a shelf or in a cabinet.

3.2. Load Cell Cable: Each load cell shall have an integrated 6-wire, 10 foot cable.

4.0 Warranty

4.1. Each load cell and indicator provided shall include a standard 1 year warranty.

5.0 Testing

5.1. The load cells shall be inspected and accepted at a destination specified elsewhere in this contract. The item will be checked with test equipment and parameters equal to or better than those required assuring specified performance requirements are met. Test equipment has first echelon certification directly traceable to the National Institute of Standards and Technology (NIST) or other AFMETCAL approved sources where required. Tests will be performed to assure the item meets or exceeds all of the requirements of this purchase description. Such

testing does not relieve the contractor of performing all inspections and quality control checks at the point of fabrication as necessary to assure performance as specified.