

**PERFORMANCE WORK STATEMENT
FOR
EQUIPMENT CALIBRATION AND
SERVICING**

**EGLIN AFB
13 April 2026**

**PERFORMANCE WORK STATEMENT (PWS)
FOR
EQUIPMENT CALIBRATION AND SERVICING**

1. **DESCRIPTION OF SERVICES.** The contractor shall provide all parts, tools, supplies, equipment, and labor necessary to conduct an annual Preventative Maintenance (PM) visit. The PM visit will repair, calibrate and service multiple presses, load frame, and Triaxial system at AFRL/RWTE, Eglin AFB. The repairs, calibrations and servicing will meet ASTM accreditation and ensure continuous, safe operation. Following is a list of equipment to be repaired, calibrated and serviced. A second maintenance visit will also be provided each year to repair and recalibrate any press, load frame, or Triax equipment failing between annual PM visits.

Equipment	Service or Product
Landmark 100kip Model 370.50 press	Calibration to ASTM accredited standards and routine maintenance
Criterion Model C43-504 presses	Calibration to ASTM accredited standards and routine maintenance
MTS Exceed 42 load frame,	Calibration and routine maintenance
Landmark 500kip Model 311.41 press	Calibration to ASTM accredited standards and routine maintenance of the 311.41 press system and HPU
Triaxial system	Calibration and routine maintenance of the Triaxial system

- 1.1 CALIBRATION/MAINTENANCE SERVICES.** The contractor shall perform repairs, calibrations and routine maintenance on multiple presses, load frame, and Triax system within the AFRL/RWTE complex. All repairs, calibrations and servicing will take place at the AFRL/RWTE complex at Eglin AFB, Florida. The contractor shall ensure the repairs and calibrations are to required standards. The contractor will ensure all repairs, calibrations and maintenance are conducted by qualified personnel. A summary of the calibrations and maintenance services are provided in the following table.

Equipment	Service Description
MTS Exceed 42	Axial Force Calibration, 0 to 30 kN, ASTM Accredited
MTS Exceed 42	Speed & Displacement, Criterion, Single Range, ASTM Accredited
MTS Exceed 42	Routine Maintenance: MTS Exceed EM 2,000 hr
Landmark Model 370.50/ Criterion Model C43-504	Axial Force Calibration, S/H, 0-100 kN (0-22 kip), ASTM Accredited
Landmark Model 370.50/ Criterion Model C43-504	Axial Force Calibration, S/H 0-500 kN (0-110 kip), ASTM Accredited
Landmark Model 370.50/ Criterion Model C43-504	LVDT/TEMPO Displacement Transducer Calibration
Landmark Model 370.50/ Criterion Model C43-504	COD Gage Calibration, ASTM Accredited
Landmark Model 370.50/ Criterion Model C43-504	Axial Force Calibration, 0 to 100 kN, ASTM Accredited
Landmark Model 370.50/ Criterion Model C43-504	Speed & Displacement, Criterion, Single Range, ASTM Accredited
311.41 press system	Axial Force Calibration, S/H, 0-5000 kN (0-1000 kip), ASTM Accredited

i311.41 press system	LVDT/TEMPO Displacement Transducer Calibration
311.41 press system	Return Filter and Seal Kit; 515, 3 and 6 Bay
311.41 press system	Fluid Sample Bottle; North American FSE Use
311.41 press system	Main Filter Kit; Model 293.22, 74 Micron
311.41 press system	Pilot Pressure Filter Kit; Model 293.xx, 3 Micron
311.41 press system	Routine Maintenance; HPU 1000 hr
311.41 press system	Routine Maintenance; S/H Frame 1000 hr
Triaxial System	2 Axial Force Load cell Calibrations, S/H, 0-100 kN (0-22 kip), ASTM Accredited
Triaxial System	4 LVDT/TEMPO Displacement Transducer Calibrations
Triaxial System	String Potentiometer Calibration
Triaxial System	Pressure transducer calibration
Triaxial System	Routine Maintenance; HPU 500 hr

- 1.1.1. The contractor will provide all parts and equipment needed for repairs, calibrations and routine maintenance on the listed press systems, load frames, and Triax system.
- 1.1.2. The contractor will ensure all repairs, calibrations and maintenance is conducted by qualified personnel.
- 1.1.3. During calibrations and routine maintenance, parts/equipment needing replacement would be purchased from a \$2500 parts fund. Any parts above this amount will be identified to AFRL/RWTE personnel and purchased by the government.
- 1.1.4. Repairs, calibrations and maintenance will be conducted on site unless an exemption is agreed to by the government.

1.2 ADDITIONAL REPAIR AND CALIBRATION VISIT. An additional service visit will be available between annual Preventative Maintenance visits to repair and recalibrate any presses or load frames that fail.

- 1.2.1. The contractor will provide all equipment needed for repairs and recalibrations on listed press systems and load frames.
- 1.2.2. The government will purchase the parts needed to repair the listed press systems and load frames.
- 1.2.3. The contractor will ensure all repairs and maintenance is conducted by qualified personnel.
- 1.2.4. Repairs and recalibrations will be conducted on site unless an exemption is agreed to by the government.

2. **SERVICE DELIVERY SUMMARY.**

Performance Objective	SOW Para	Performance Threshold
Calibrations Calibrations for listed presses and load frames must meet ASTM accredited standards	1.1, 1.1.1, 1.1.3, 1.2, 1.2.1, 1.2.3	Final calibrations meet ASTM standards.
Repairs Repairs for listed presses and load frames must allow for safe operations and ensure the equipment works within specifications	1.1, 1.1.1, 1.1.3, 1.2, 1.2.1, 1.2.3	Final repairs meet specifications and minimize equipment downtime.
Routine Maintenance Routine maintenance will be conducted on listed presses and load frames during calibration visit. Parts/equipment needing replacement or repair will be identified	1.1 1.1.2 1.1.4	Maintenance will allow continued operation. Parts/equipment needed replacement or repair will be identified 95% of the time

3. **GOVERNMENT FURNISHED PROPERTY AND SERVICES.** N/A

4. **QUALITY CONTROL**

4.1. QUALITY CONTROL. As a minimum, the contractor shall develop quality control procedures that address the areas identified in paragraph 2, Service Delivery Summary. The government evaluator must have a specific quality control inspector to notify in case of customer complaints.

4.2. QUALITY ASSURANCE. The government will periodically evaluate the contractor's performance in accordance with the quality control procedures established in paragraph 4.1.

4.3. GOVERNMENT REMEDIES. The contracting officer shall follow the requirements of FAR 52.212-4, Contract Terms and Conditions for Commercial Items (May 1997), for contractor's failure to correct nonconforming services.

5. **APPENDICES.**

A. Estimated Workload Data

ESTIMATED WORKLOAD DATA

ITEM	NAME	ESTIMATED QUANTITY	
1	Exceed Model E42-503 calibration and maintenance	8 (est)	hours
2	Landmark 100kip Model 370.50	8 (est.)	hours
3	Criterion Model C43-504 calibration and maintenance	8 (est)	hours
4	Landmark 500kip Model 311.41 press system calibration and maintenance	8 (est.)	hours
5	Triaxial System calibration and maintenance	8 (est.)	hours
6	Additional repair visit	12 (est.)	hours