

Flash Point Tester

STATEMENT OF NEED (SON)

05 Mar 2026

1. Requirement: T.O. 42B-1-1 Dictates the testing of various fuels flash point. The flash point of fuel is a critical factor in determining if fuel is safe and in specification to be used by Department of War assets. This equipment is essential in accomplishing this standard operation procedure.

2. Characteristics: A flash point tester capable of performing to the standard of the following ASTM's must be met: D56/D93/D3828/D7236/IP 160/IP 523/ IP 534. The model PMA 500 accomplishes this requirement as it can be programed from the manufacturer.

CLIN 0001: *Pensky-Martins Flash Point Tester PMA 500/Qty: 01 – High precision flash point testing of petroleum products. — Anton Paar, PMA 500.*

The PMA 500 is capable of being programed from the manufacturer, it is capable of being calibrated IAW USAF Precision Measurement Equipment Laboratory directives. The PMA 500 meets the American Standards Testing Method standards as required by T.O. 42B-1-1.

3. Inherent Services: The equipment will need to be programmed at manufacturer prior to delivery at MacDill AFB. We will need technical data to provide PMEL with calibration specifications.

4. Delivery: The item will need to be delivered to MacDill AFB by however means is appropriate.

2914 Egret Ave, Tampa FL, 33621.

5. Terms and Conditions: The item will come with a limited manufacturer warranty for a period of time.

3-year warranty starts from the date of delivery and applies solely on purchased instrument(s). Consumable items and any spare parts purchased separately are excluded from this warranty. (1) This instrument requires regular scheduled service. The 3 year warranty will only be applicable if the service is carried out in accordance with the instrument maintenance plan.