

DEPARTMENT OF VETERANS AFFAIRS

**Justification for Single Source Awards IAW [FAR 13.106-1](#)
For
Over Micro-Purchase Threshold but Not Exceeding the Simplified Acquisition
Threshold**

Acquisition Plan Action ID: 36C244-26-AP-1987

1. **Contracting Activity:** Department of Veterans Affairs, VISN 4, Station 646 VA Pittsburgh Healthcare System, University Drive C, Pittsburgh, PA 15240 | 646-26-2-233-0300
2. **Brief Description of Supplies/ Services required and the intended use/Estimated Amount:** The VA Pittsburgh Healthcare System (VAPHS) Sterile Processing Services requires the brand name procurement for Nutrace R+ Scanners. The compatibility of these scanners with the Censitrac system will allow the VA Pittsburgh Healthcare System the ability to remain on the path of 100% instrument traceability and accountability in support the VA mission of providing quality care to our Veterans.
The estimated dollar value for this requirement is [REDACTED]
3. **Unique characteristics that limit availability to only one source, with the reason no other supplies or services can be used:** The equipment cleaned and reprocessed by Sterile Processing Services (SPS) must be effectively traced via the Censitrac equipment tracking software system already implemented by SPS. This system allows for the precise tracking and utilization of equipment and reusable medical devices (RMDs) in patient care, inventory management, and repairs. A key component of this process is the unique device identification (UDI) laser mark specifically, a 2D data matrix code used to commission RMDs within the Censitrac system.

The Nutrace R+ scanner is uniquely designed to read the most challenging 2D data matrix codes using its built-in magnifier, thereby significantly improving scan capture reliability. This reduces scan failures and minimizes the need for repeat scans, thereby increasing productivity and efficiency for sterile processing technicians.

Additionally, these scanners have a compact design, taking up minimal space at workstations and eliminating the need for stands, which are required by our current scanners. Made with a stainless-steel rugged design, they are easily disinfected, ensuring high hygiene standards. Importantly, unlike the current scanner models that emit potentially eye-damaging lights, the Nutrace R+ scanners do not produce any hazardous flashing lights, thereby enhancing user safety.

Given these unique features and the compatibility with the Censitrac system, the Nutrace R+ scanners are indispensable for our operations. Martin AIJS is the authorized vendor and sales channel for NuTrace products and services.

- 4. Description of market research conducted and results or statement why it was not conducted:** Market research was performed by searching GSA Advantage, where we found no other vendors or contracts available. Specifically, there were no vendors or contracts that could meet the stringent performance standards set by Directive 1116.

A sources sought notice was posted seeking alternate sources resulting in only one authorized SDVOSB vendor capable of supplying the required item.

According to the VHA Directive 1116, which mandates 100% accountability in Sterile Processing Services (SPS), precise and high-speed tracking capabilities are essential. The Nutrace R+ scanners were deemed necessary because they meet these specifications. For instance, "the R+ healthcare scanner is capable of reading a 2D Data Matrix in 0.5 seconds per instrument, regardless of contrast, material, or position"—a capability that other vendors could not match (https://www.patientcare.va.gov/Office_Sterile_Processing/VHA_Directive_1116.asp).

Martin AIJS is the authorized vendor and sales channel for NuTrace products and services.

- 5. Contracting Officer's Certification:** *Purchase is approved in accordance with FAR 13.106-1(b). I certify that the foregoing justification is accurate and complete to the best of my knowledge and belief.*

BENJAMIN VILLASENOR

Name
Title

 Digitally signed by BENJAMIN VILLASENOR
Date: 2026.03.16 09:07:43 -04'00'

Date