
Background

The Gallup Indian Medical Center (GIMC) Housekeeping-EVS Department is seeking service support for four ultraviolet-c machines.

Objective

Gallup Indian Medical Center objective is seeking to obtain services including maintenance, repair, and replacement of key components such as:

UVC light bulbs: 254nm UVC-lamps

Batteries: Eco Flow River, Eco Flow Delta Mini, Eco Flow River 2 Pro

Tablets: Android

Phones: Android

OhmniClean UVC base autonomous drive system, UV light module, depth camera

Period of Performance:

The base year begins January 1, 2026, to December 31, 2026, with two option years if exercised

- i. Option Year 1: January 1, 2027, to December 31, 2027
- ii. Option Year 2: January 1, 2028, to December 31, 2028

Scope

The service provider shall furnish all labor, materials, tools, and expertise necessary to support the functionality of the UVC machines during the warranty period.

Services include:

Preventative Maintenance

Routine inspection of each UVC machine annually on site at Gallup Indian Medical Center.

Inspection will include verification of device performance, calibrations, and safety features. An assessment of operating hours, lamp output, and battery condition.

Corrective Maintenance

Onsite or offsite repair of any device malfunction
Diagnostics and troubleshooting of hardware or software issues affecting device functionality.

Parts Replacement

The following components shall be covered under the service warranty and replaced at no additional cost when failure or degradation occurs under normal use.

Batteries: Eco Flow River, Eco Flow Delta Mini, Eco Flow River 2 Pro
Tablet: Android
Phones: Android
UVC Light Bulbs: 254nm UVC lamps

The following items are not included

Damage caused by improper use, accidents, unauthorized repairs, physical damage to tablets and phones due to drops.

Task

The service provider shall perform the following tasks:

Conduct an initial inspection of all four UVC machines. Document the operating condition, lamp hours, battery status, tablet, and phone functionality

Perform scheduled maintenance annually

Respond to reported issues within 3 to 5 business days

Diagnose any malfunction or abnormal operation

Repair or replace faulty components covered

Replace end of life or malfunctioning UVC lamps

Replace defective or degraded batteries

Replace malfunctioning tablets or phones

Verify successful installation and proper operation after replacement

Security

Contractors must follow GIMC security and identification protocols

Place of Performance

Gallup Indian Medical Center, 516 East Nizhoni Blvd, Gallup NM 87301.

Inspection

The government reserves the right to inspect the contractor's place of performance, daily operations, equipment, licenses and certifications, and quality of work at any given time.

Point of Contact

[REDACTED]

Contracting Officer Representative:

GIMC will identify a contracting officer representative (COR) to be determined at the time award