COHORT SHEET HLB-fem-profsurgery-troubleshooting

32 mouse recordings from 4 nights. Every mouse is checked by the same six questions, in order. A question is only asked if the one before it passed.

READ EACH ROW LEFT TO RIGHT. The first red box is what went wrong for that mouse. Read a column downwards to see whether the whole cohort failed the same way.

		0 FILE	1 IDENTITY	2 CONNECTION	3 SCOREABILITY	4 ARCHITECTURE	5 EXPORT	usable	longest block	scoreable	persistence
2026-07-31 1 uid0731-18-06-37	A1 HLB-A1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	A2 HLB-A2-NO-MOUSE original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B1 HLB-B1a original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B2 HLB-B2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C1 HLB-C1 original	✓	✓	✓	✓	◐	?	30.3%	46 min	8.73 h	+0.761
	C2 HLB-C2 original	✓	✓	✓	✓	◐	?	6.6%	28 min	2.34 h	+0.708
	D1 HLB-D1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D2 HLB-D2 original	✓	✓	✗	–	–	–	8.7%	13 min	–	–
2026-08-02 2 uid0802-01-46-57	A1 HLB-A1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	A2 HLB-A2-NO-MOUSE original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B1 HLB-B1a original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B2 HLB-B2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C1 HLB-C1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C2 HLB-C2 original	✓	✓	✗	–	–	–	0.2%	0 min	–	–
	D1 HLB-D1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D2 HLB-D2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
2026-08-06 0 uid0806-12-45-02	A1 HLB-A1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	A2 HLB-A2-NO-MOUSE original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B1 HLB-B1b short soldered	✓	✓	✓	✓	◐	✓	99.7%	75 min	3.18 h	+0.481
	B2 HLB-B2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C1 HLB-C1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C2 HLB-C2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D1 HLB-D1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D2 HLB-D2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
2026-08-06 1 uid0806-16-57-38	A1 HLB-A1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	A2 HLB-A2-NO-MOUSE original	✓	✓	✗	–	–	–	0.0%	–	–	–
	B1 HLB-B1b short soldered	✓	✓	✓	✓	◐	✓	99.6%	216 min	23.66 h	+0.603
	B2 HLB-B2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C1 HLB-C1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	C2 HLB-C2 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D1 HLB-D1 original	✓	✓	✗	–	–	–	0.0%	–	–	–
	D2 HLB-D2 no slip ring	✓	✓	✓	✓	◐	✓	86.5%	87 min	23.28 h	+0.499

✓ PASS = asked and answered yes ◐ PROVISIONAL = answered, but gated behind work not yet done ? INSUFFICIENT = the question could not be asked ✗ FAIL = asked and answered no – BLOCKED = a lower rung failed, so this was never evaluated

Tethers tested in this cohort: original = original tether short soldered = short 19.5 cm, soldered joints, resin encapsulated no slip ring = long, soldered joints, direct to headstage, no slip ring

"Usable" means the mouse channel is measurably different from amplifier inputs with no wire attached, recorded at the same moment in the same file.
"Persistence" is how strongly the sleep or wake state at one minute predicts the state a minute later; shuffled data scores about zero.
The sleep column stays PROVISIONAL for every mouse until a person has blind scored a set of clips and we can measure how closely the automatic scoring matches a human.