`allego_1__uid0731-18-06-37`

post-surgery day not recorded · day 1 post-tether · body weight not recorded · duration unknown · ZT span unknown · criterion HUM · QC epoch 10 s

LIFE TIMELINE



VERDICT LADDER

✓ 0 FILE — PASS
Is the recording file physically sound?
100,454,784 samples, 27.90 h
vs data samples == timestamp samples, remainder 0
100,454,784 data samples = 100,454,784 timestamp samples, remainder 0 and 0

✓ 1 IDENTITY — PASS
Which native channels belong to this animal?
EEG B/33 EMG B/34
vs layout derived from the file, never assumed
2 channels resolved from sensors_by_port; 112 wire-free inputs available as an in-file floor

✗ 2 CONNECTION — FAIL
Is the electrode connected, judged against wire-free inputs in the same file?
0.0% usable, longest unbroken block 0.0 min
vs 60 Hz < 98.4 uV (p1 of 112 wire-free inputs in this file), rms 10-150 uV, zero railed samples
no epoch passes on every role; weakest is EEG at 0.0%


– 3 SCOREABILITY — BLOCKED
Can the sleep question even be asked?
rung 2 CONNECTION did not pass

– 4 ARCHITECTURE — BLOCKED
What did the animal do?
rung 3 SCOREABILITY did not pass

– 5 EXPORT — BLOCKED
Can a human check us?
rung 3 SCOREABILITY did not pass

A rung is not evaluated if the rung below it failed: it reads BLOCKED. ? INSUFFICIENT means the question could not be asked, which is NOT the same as ✗ FAIL.

RUNG 2 evidence — usable epochs against the in-file wire-free floor



falsification: the same criterion scores the 112 wire-free inputs in this file at 0.88% usable, and must be near zero · floor 60 Hz p1 98 uV, 1-40 Hz slope p1 -3.48

– RUNG 4 ARCHITECTURE: BLOCKED

rung 3 SCOREABILITY did not pass

NO PANEL IS DRAWN FOR A RUNG THAT DID NOT PASS.

Shared-path fault. Its own two channels correlate at 0.50 where working animals sit at 0.07, and 100 percent of its EEG-railed epochs also rail on EMG, which localises the fault to a path shared by both channels rather than to two coincidental electrode failures.