
HLB-B1b · F · age not recorded · position B1 · tether arm B

neurosurveillance-eeg · HLB-fem-profsurgery-troubleshooting

allego_1__uid0806-16-57-38

day 16 post-surgery · day 0 post-tether · body weight not recorded · 23.66 h · ZT 8.0 to 7.6 · criterion HUM · QC epoch 10 s

LIFE TIMELINE



VERDICT LADDER

✓ 0 FILE — PASS

Is the recording file physically sound?

85,172,608 samples, 23.66 h

vs data samples == timestamp samples, remainder 0

85,172,608 data samples = 85,172,608 timestamp samples, remainder 0 and 0

✓ 1 IDENTITY — PASS

Which native channels belong to this animal?

EEG B/1 EMG B/2

vs layout derived from the file, never assumed

2 channels resolved from sensors_by_port; 112 wire-free inputs available as an in-file floor

✓ 2 CONNECTION — PASS

Is the electrode connected, judged against wire-free inputs in the same file?

99.6% usable, longest unbroken block 215.8 min

vs 60 Hz < 415.0 uV (p1 of 112 wire-free inputs in this file), rms 10-150 uV, zero railed samples

8,485 of 8,517 epochs usable on every role, with an unbroken block of 215.8 min

✓ 3 SCOREABILITY — PASS

Can the sleep question even be asked?

23.66 h in 1 block(s), r@1min 0.603

vs >= 15 min inside blocks of >= 124 s, within-block r@1min > 0.30, wake 15-90%

longest block 1419.5 min; persistence 0.603 against a shuffled 0.010

◐ 4 ARCHITECTURE — PROVISIONAL

What did the animal do?

wake 50.2%, NREM 46.5%, asleep in light 81.2%

vs no pass mark: this rung reports, it does not judge

the light-phase figure swings 36.0 points across a +/-25% threshold sweep; no human has scored any epoch, so accuracy is unmeasured; ...

✓ 5 EXPORT — PASS

Can a human check us?


4/4 pipeline artifacts, plain text written

vs EDF for a scorer, plain text for the professor, both present

EDF and plain-text export both present

A rung is not evaluated if the rung below it failed: it reads BLOCKED. ? INSUFFICIENT means the question could not be asked, which is NOT the same as ✗ FAIL.


RUNG 2 evidence — usable epochs against the in-file wire-free floor



falsification: the same criterion scores the 112 wire-free inputs in this file at 0.00% usable, and must be near zero · floor 60 Hz p1 415 uV, 1-40 Hz slope p1 -1.35

RUNG 4 evidence — how much the wake threshold moves the answer

the light-phase sleep figure moves 36.0 points across a +/-25% sweep of that one threshold



READ WITH THE PANEL ON THE LEFT

- PROVISIONAL until a blind-scoring Cohen kappa exists. No human has scored any epoch from this project, so these labels have no known accuracy.
- The light-phase sleep figure moves 36.0 points across a +/-25% sweep of the wake threshold, so it is NOT threshold-independent and agreement with the literature is not independent confirmation.
- NO REM PERCENTAGE IS REPORTED: the percentile rule and an independent Otsu cut disagree by roughly a factor of five on this montage.

A DIFFERENT ANIMAL from HLB-B1a in the same rig position. Implanted 21 July with the rest of the cohort, then left in its home cage untethered for 16 days, which makes it the first animal this project ever tethered in line with its own 14 day protocol. Produced 23.66 h of usable EEG in ONE unbroken block, the best recording the project has.