
# HLB-C1 · F · age not recorded · position C1 · tether arm A

neurosurveillance-eeg · HLB-fem-profsurgery-troubleshooting

allego_1__uid0806-16-57-38

post-surgery day not recorded · day 7 post-tether · body weight not recorded · 23.66 h · ZT 8.0 to 7.6 · criterion HUM · QC epoch 10 s

## LIFE TIMELINE




## VERDICT LADDER

**✓ 0 FILE** — PASS

*Is the recording file physically sound?*

**85,172,608 samples, 23.66 h**

vs data samples == timestamp samples, remainder 0

85,172,608 data samples = 85,172,608 timestamp samples, remainder 0 and 0

**✓ 1 IDENTITY** — PASS

*Which native channels belong to this animal?*

**EEG C/1 EMG C/2**

vs layout derived from the file, never assumed

2 channels resolved from sensors_by_port; 112 wire-free inputs available as an in-file floor

**✗ 2 CONNECTION** — FAIL

*Is the electrode connected, judged against wire-free inputs in the same file?*

**0.0% usable, longest unbroken block 0.0 min**

vs 60 Hz < 415.0 uV (p1 of 112 wire-free inputs in this file), rms 10-150 uV, zero railed samples

no epoch passes on every role; weakest is EEG at 0.0%

**– 3 SCOREABILITY** — BLOCKED

*Can the sleep question even be asked?*

rung 2 CONNECTION did not pass

**– 4 ARCHITECTURE** — BLOCKED

*What did the animal do?*

rung 3 SCOREABILITY did not pass


**– 5 EXPORT** — BLOCKED

*Can a human check us?*

rung 3 SCOREABILITY did not pass

A rung is not evaluated if the rung below it failed: it reads BLOCKED. ? INSUFFICIENT means the question could not be asked, which is NOT the same as ✗ FAIL.

## RUNG 2 evidence — usable epochs against the in-file wire-free floor




falsification: the same criterion scores the 112 wire-free inputs in this file at 0.00% usable, and must be near zero · floor 60 Hz p1 415 uV, 1-40 Hz slope p1 -1.35

## – RUNG 4 ARCHITECTURE: BLOCKED

rung 3 SCOREABILITY did not pass

*NO PANEL IS DRAWN FOR A RUNG THAT DID NOT PASS.*

The single most informative animal in the cohort. Best EEG in the fleet on 31 July at 0.465 MOhm, survived deliberate joint-by-joint provocation of its cable, and produced 8.73 h of sleep in one 8.1 h unbroken block with r@1min +0.744, still the highest this project has recorded. Two days later its EEG was electrically open, and it never recorded again even though the impedance returned to a good value on 4 August. A channel can read connected and still record nothing.

HLB-fem-profsurgery-troubleshooting · card schema v2 · generated 2026-08-09T11:11:53+09:00 · git not a git repo · thresholds c635d5245a5b (qc_thresholds.py) · criterion HUM at 10 s epochs. SLOPE and HUM usable fractions are NOT comparable, and neither is comparable to a fraction computed at the 4 s staging epoch.