



Patient Monitoring Solutions





PAVO Vital Sign Monitor

Features

- 8" color TFT LCD Screen (touch screen optional)
- Portable, lightweight and sturdy design
- Flexible parameters configuration for different clinical environments
- Rechargeable Li-ion Battery (up to 12 hours uninterrupted work)
- Big font and font color display setting
- Spot-check and continuous monitoring mode
- Selectable for Adult, Pediatric and Neonatal patients
- Wired/Wireless CMS, support HL7 protocol to HIS
- Barcode scanner support
- Thermal recorder support
- Graphical & tabular trend review
- 48 hours holographic wave review for each patient (stored in SD card)



For Out-Patient Department, Spot-check, Transport, Ward and other Basic Monitoring.

Configuration

Optional

SpO2 + NIBP, Li-ion battery	Masimo/Nellcor SpO2, Quick Temp, Barcode scanner
SpO2+NIBP+ECG+TEMP, Li-ion battery	Masimo/Nellcor SpO2, EtCO2, Quick Temp, Barcode scanner, Thermal Recorder

Technical Specifications

Display

8" color TFT LCD Screen, resolution: 800 x 600

ECG

Lead type

3-lead: I, II, III

5-lead: I, II, III, aVR, aVL, aVF, V

Display sensitivity:

2.5 mm/mV (×0.25), 5 mm/mV (×0.5),

10 mm/mV (×1.0), 20 mm/mV (×2.0)

Wave sweep speed: 6.25 mm/s, 12.5 mm/s,

25 mm/s, 50 mm/s

Bandwidth

Diagnostic mode: 0.05Hz~100Hz

Monitor mode: 0.5Hz~40Hz

Surgery mode: 1Hz~20Hz

Strong filter mode: 5Hz~20Hz CMRR

>100dB

Notch: 50/60 Hz notch filter can be set to on or off

Differential input impedance >5 MΩ

Electrode polarization voltage range: ±400 mV

Baseline recovery time <3 s after defibrillation (in monitor and surgery mode)

Calibration signal: 1 mV (peak - peak), accuracy ±3%

RESP

Measurement method: Thoracic electrical bioimpedance

Rate: 0 – 150 bpm

Measuring lead: Lead I, II

Wave gain: ×0.25, ×0.5, ×1, ×2

Respiratory impedance range: 0.5-5 Ω

Baseline impedance: 500-4000 Ω

Gain: 10 grades

Scan speed: 6.25 mm/s, 12.5 mm/s, 25 mm/s

TEMP

Measurement method: Thermistor

Measuring range: 5~50 °C (41~122 °F)

Resolution: 0.1 °C

Measurement accuracy: ±0.1 °C

Recorder (optional)

Built-in, Thermal dot array

Horizontal resolution: 16 dots/mm (25 mm/s paper speed)

Vertical resolution: 8 dots/mm

Paper speed: 25 mm/s, 50 mm/s

Number of waveform channels: 3

PAVO Vital Sign Monitor

Technical Specifications

NIBP

Measurement method: Automatic oscillometric method

Operating mode: Manual, automatic, continuous

Measurement unit: mmHg/kPa selectable

Typical measurement time: 20~40 s

Measurement type: Systolic, Diastolic, Mean

Measurement range (mmHg)

Range of Systolic pressure:	Adult	40-270
	Pediatric	40-230
	Neonatal	40-135

Range of Diastolic pressure:	Adult	10-210
	Pediatric	10-150
	Neonatal	10-100

Range of Mean pressure:	Adult	20-230
	Pediatric	20-165
	Neonatal	20-110

Measurement accuracy

Maximum average error: ± 5 mmHg

Maximum standard deviation: 8 mmHg

Resolution: 1 mmHg

Interval: 1, 2, 3, 4, 5, 10, 15, 30, 60, 90, 120, 180, 240, 480 minutes

Overpressure protection: Software and hardware, double safety protection

Cuff pressure range: 0-280 mmHg

Standard SpO2

Measurement range: 0-100%

Resolution: 1%

Accuracy: $\pm 2\%$ (70-100%, Adult/Pediatric);
 $\pm 3\%$ (70-100%, Neonate);
0-69%, unspecified

Refreshing Rate: 1s

Masimo SpO2 (optional)

Measurement range: 0-100%

Resolution: 1%

Accuracy: $\pm 2\%$ (70-100%, Adult/Pediatric),
non-motion, low
 $\pm 3\%$ (70-100%, Neonate,
non-motion);
 $\pm 3\%$ (70-100%, motion);
0-69%, unspecified

Refreshing Rate: 1s



Portable Design



Touch Screen
(Optional)



Quick Temp
(Infrared Ear Thermometer)



Infrared Ear Thermometer (optional)

Displayed range: 34~42.2 °C (93.2~108 °F)

Operation ambient temperature range:

10~40 °C (50~104 °F)

Accuracy for displayed temperature range:

≥35 °C (95.9 °F) ~ ≤42.2 °C (107.6 °F) range

±0.2 °C (0.4 °F)

<35 °C (95.9 °F) ~ ≥34 °C (93.2 °F) range

±0.3 °C (0.5 °F)



Phasein IRMA™ Sidestream CO2 (optional)

Warm-up time: Full accuracy within 10 seconds

Sampling flow rate: 50 ml/min (+/-10/min)

Accuracy: ± (0.2% +2% of the reading)

Measurement Range: 0 -15%

Rise time: 200 ms, typical at 50 ml/min flow rate

Total response time: within 3 seconds (with 2m Momoline sampling line)

AWRR Range: 0-150 bpm

AWRR Accuracy: ±1 breath

Phasein IRMA™ Mainstream CO2 (optional)

Measurement Range: 0-15%

Warm-up time: Full accuracy within 10 seconds

Accuracy: ± (0.2% +2% of the reading)

AWRR Range: 0-150 bpm

AWRR Accuracy: ±1 breath

Operation Environment

Power: AC 100-250 V, 50/60 Hz

Temperature: 0-40 °C

Humidity: 15-85%

Patient Range: Adult, Pediatric, Neonate

Battery backup: Standard 4-5 hrs (2.600 mAh), optional 8-10 hrs (5.200 mAh) or 12-15 hrs (7.800 mAh)



Patient Monitoring Solutions

For more information, please contact us.

aXcent medical GmbH

Josef-Görres-Platz 2
56068 Koblenz
Germany

Phone: +49 261 30 11 117
Fax: +49 261 30 11 111
Mail: info@axcentmedical.com



www.axcentmedical.com

Alkaram - Medical

- 📍 10th of Ramadan City, Alordonia, Al Jawhara Mall, Office 447.
- 📍 32 Obur Bulding - Salah Salem Street, Cairo, Egypt.

✉ sales@alkaram-medical.com
🌐 www.alkaram-medical.com
☎ 01099442054 - 01206467528
☎ 01272398884 - 01032671638



www.alkaram-medical.com